# Agilent 1290 Infinity Quaternary Pump

User Manual

Agilent Technologies

# Notices

© Agilent Technologies, Inc. 2012-2014, 2015

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

## Manual Part Number

G4204-90001

## Edition

04/2015

Printed in Germany

Agilent Technologies
Hewlett-Packard-Strasse 8
76337 Waldbronn

## Warranty

**The material contained in this document is provided "as is," and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.**

## Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

## Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed as "Commercial computer software" as defined in DFAR 252.227-7014 (June 1995), or as a "commercial item" as defined in FAR 2.101(a) or as "Restricted computer software" as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Agilent Technologies' standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

## Safety Notices

**CAUTION**

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

**WARNING**

**A WARNING notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.**

# In This Guide...

This manual covers the Agilent 1290 Infinity Quaternary Pump (G4204A).

### 1 Introduction

This chapter gives an introduction to the module, instrument overview and internal connectors.

### 2 Site Requirements and Specifications

This chapter provides information on environmental requirements, physical and performance specifications.

### 3 Installing the Module

This chapter gives information about the preferred stack setup for your system and the installation of your Agilent 1290 Infinity Quaternary Pump.

### 4 Using the Pump

This chapter explains the operational parameters of the Agilent 1290 Infinity Quaternary Pump.

### 5 How to Optimize the Performance of Your Module

This chapter gives hints on how to optimize the performance or use additional devices.

### 6 Troubleshooting and Diagnostics

Overview about the troubleshooting and diagnostic features.

### 7 Error Information

This chapter describes the meaning of error messages, and provides information on probable causes and suggested actions how to recover from error conditions.

### 8 Test Functions and Calibrations

This chapter will describe the tests for the module.

### 9 Maintenance

This chapter describes the maintenance of the Agilent 1290 Infinity Quaternary Pump.

### 10 Parts and Materials

This chapter provides information on parts for maintenance.

### 11 Identifying Cables

This chapter provides information on cables used with the Agilent 1200 Infinity Series modules.

### 12 Hardware Information

This chapter describes the pump in more detail on hardware and electronics.

### 13 LAN Configuration

This chapter provides information on connecting the module to the controller software.

### 14 Appendix

This chapter provides additional information on safety, legal and web.

# Contents

**1 Introduction 9**

Features 10
Overview of the Quaternary Pump 11
Operating Principle 12
Positions of the Multi Purpose Valve 14
System Overview 18

**2 Site Requirements and Specifications 21**

Site Requirements 22
Physical Specifications 25
Performance Specifications 26

**3 Installing the Module 29**

Unpacking the Module 30
Optimizing the Stack Configuration 31
Installation Information on Leak and Waste Handling 36
Removing the Transport Foam 39
Installing the Pump 40
Flow Connections to the Pump 42
Installation of Seal Wash Function 45

**4 Using the Pump 47**

Leak and Waste Handling 48
Best Practices 49
Setting up the Pump with the Instrument Control Interface 53
Flushing the Filter 66
Solvent Information 68
Algae Growth in HPLC Systems 75

**5 How to Optimize the Performance of Your Module 77**

Using the Degasser 78
Operational Hints for the Multi Channel Gradient Valve (MCGV) 79
Delay Volume and Extra-Column Volume 80
How to Configure the Optimum Delay Volume 81
How to Achieve Higher Resolution 83
Using Solvent Calibration Tables 86

**6 Troubleshooting and Diagnostics 87**

Overview of the Module's Indicators and Test Functions 88
Status indicators 89
Available Tests vs User Interfaces 91
Agilent Lab Advisor Software 92

**7 Error Information 93**

What Are Error Messages 95
General Error Messages 96
Pump Error Messages 102

**8 Test Functions and Calibrations 121**

**9 Maintenance 123**

Introduction to Maintenance 125
Warnings and Cautions 126
Overview of Maintenance 128
Cleaning the Module 129
Installing Fittings and Capillaries 130
Replacing the Pressure Sensor 131
Replacing the Inlet Weaver 134
Replacing the Inlet Valve 136
Replacing the Outlet Valve 138
Removing the Jet Weaver 141
Installing the Jet Weaver 144
Replacing the Seal Wash Pump 146
Replacing the Multi-Channel Gradient Valve (MCGV) 148
Releasing a Stuck Inlet Valve 152
Replacing the Pump Head 155

Disassembling the Pump Head 163
Disassembling the Primary Pump Head 165
Disassembling the Secondary Pump Head 170
Replacing the Heat Exchanger 174
Replacing Wash Seal and Gasket 177
Assembling the Pump Head 179
Replacing the Multi Purpose Valve 191
Replacing Parts of the Multi Purpose Valve 194
Replacing the Outlet Filter 196
Installing the Inline Filter 198
Removing the Inline Filter 200
Replacing Parts of the Inline Filter 202
Installing the Valve Rail Kit 205
Replacing Module Firmware 206
Preparing the Pump Module for Transport 207

**10 Parts and Materials 211**

Overview of Main Assemblies 212
Flow Connections 214
Seal Wash Function 215
Pump Head Assembly Parts 216
Primary Pump Head Parts 218
Secondary Pump Head Parts 222
Multi Purpose Valve 225
Solvent Cabinet 226
Cover Parts 228
Leak Parts 229
Accessory Kit 230
Others 231

**11 Identifying Cables 233**

Cable Overview 234
Analog cables 236
Remote Cables 238
BCD Cables 241
CAN/LAN Cable 243
RS-232 Cable Kit 244
Agilent 1200 Module to Printer 245

**12 Hardware Information 247**

Firmware Description 248
Electrical Connections 251
Interfaces 253
Setting the 8-bit Configuration Switch 260
Early Maintenance Feedback 263
Instrument Layout 264

**13 LAN Configuration 265**

What You Have To Do First 266
TCP/IP Parameter Configuration 267
Configuration Switch 268
Initialization Mode Selection 269
Dynamic Host Configuration Protocol (DHCP) 273
Link Configuration Selection 276
Automatic configuration with Bootp 277
Manual Configuration 287
PC and User Interface Software Setup Setup 293

**14 Appendix 295**

General Safety Information 296
The Waste Electrical and Electronic Equipment (WEEE) Directive (2002-96-EC) 299
Radio Interference 300
Sound Emission 301
Agilent Technologies on Internet 302

# 1 Introduction

Features 10
Overview of the Quaternary Pump 11
Operating Principle 12
Positions of the Multi Purpose Valve 14
System Overview 18
Leak and Waste Handling 18

This chapter gives an introduction to the module, instrument overview and internal connectors.

# Features

The G4204A Quaternary Pump is designed for highest performance, GLP compliance and easy maintenance. It includes the following features:

- Seal wash function for continued high lifetime of pump seals for buffer applications.
- Optional Jet Weaver for optimum mixing performance with a minimum of delay volume.
- Automatic purge function for ease of use and unattended preparation of the system.
- Auto tuning of the delivery cycle for compensation of elasticity and solvent volume change effects (compressibility, thermal expansion).
- Solvent selection for optimum flow accuracy.
- Fast defill function for improved intake and delivery performance.
- Two pistons in series design for increased reliability.
- High resolution piston movement control for smooth and reliable motion.

For specifications, see "Performance Specifications" on page 26.

# Overview of the Quaternary Pump

The Agilent 1290 Infinity Quaternary Pump has a built-in 4-channel vacuum degasser for best flow stability, especially at low flow rates and maximum detector sensitivity. It uses a multi-channel gradient valve (MCGV) for formation of quaternary gradients at low pressure. The low-pressure Inlet Weaver based on patented Agilent microfluidic technology ensures highest mixing performance and lowest mixing noise. The pump head offers a high power range with a maximum pressure of 1200 bar and a maximum flow rate of 5 mL/min. The Multi Purpose Valve can be used for automatic purging, using an optional Jet Weaver high-performance mixer, automatic back-flushing of the optional inline filter or for diagnostic.

The Agilent 1290 Infinity Quaternary Pump is suitable for a wide range of columns and HPLC und UHPLC applications starting from typical 250 x 4.6 mm HPLC columns going down to high resolution 50 x 2.1 mm UHPLC columns and can be used in a flow range between 0.05 – 5 mL/min. The active seal wash function can be used with concentrated buffer solutions.

# Operating Principle

The pump head comprises two pump chambers in series with independent high-resolution motion control. A pressure sensor in the flow path monitors the pressure. The pump control uses this signal for minimizing the pressure ripple in order to achieve highest flow precision. A stable flow can be delivered even in case of eventual small internal leaks, which can be compensated automatically. A heat exchanger between two pump chambers strongly reduces thermal effects due to solvent compression under very high pressures.

As solvents are compressed by the pump head and expand further down the flow path, for example in the column, the volumetric flow is changed depending on the compressibility of the liquid. Agilent control software allows specifying pure solvents, pre-mixed solvents and solvent gradients. Associated Agilent solvent libraries are used by the pump control for enhanced flow accuracy, which is required for cross-instrument or cross-system reproducibility and method compatibility.

A high resolution encoder unit is attached to the pump drives, which divides a single turn into 65000 steps. Each step corresponds to a volume of about 300 pL, which allows an extremely precise control.




**Figure 1** The hydraulic path

# Positions of the Multi Purpose Valve

The Multi Purpose Valve allows easy software controlled switching between different modes of operation.

## Normal Operating Mode Without Mixer

In normal operating mode, the flow comes from the pump head, passes the pressure sensor and arrives at the central port of the Multi Purpose Valve. The flow leaves the valve through port 4 to the system (autosampler etc.).




**Figure 2** Valve position in normal operating mode without mixer

## Purge Mode

In purge mode, the flow is diverted to the waste container.




**Figure 3** Valve position in purge mode

## Normal Operating Mode With Jet Weaver and Optional Inline Filter

In this mode, the flow passes an optional Jet Weaver and the optional inline filter. This configuration is recommended for special applications which require an increased mixing efficiency.




**Figure 4** Valve position in normal operating mode with Jet Weaver

## Filter Flush Mode

This mode is used for cleaning the inline filter by back-flushing it. The flow goes to port 5, passes the inline filter in opposite direction and leaves to the waste through port 7.




**Figure 5** Valve position in filter flush mode

**CAUTION**

Damage to the valve

➔ Use the filter flush mode only if the optional inline filter is installed.

# System Overview

## Leak and Waste Handling

The 1200 Infinity Series has been designed for safe leak and waste handling. It is important that all security concepts are understood and instructions are carefully followed.




**Figure 6** Leak and waste handling concept (overview - typical stack configuration as an example)

The solvent cabinet (1) is designed to store a maximum volume of 6 L solvent. The maximum volume for an individual bottle stored in the solvent cabinet should not exceed 2.5 L. For details, see the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets (a printed copy of the guideline has been shipped with the solvent cabinet, electronic copies are available on the Internet).

The leak pan (2) (individually designed in each module) guides solvents to the front of the module. The concept covers also leakages on internal parts (e.g. the detector's flow cell). The leak sensor in the leak pan stops the running system as soon as the leak detection level is reached.

The leak pan's outlet port (3, A) guides excessive overfill from one module to the next, as the solvent flows into the next module's leak funnel (3, B) and the connected corrugated waste tube (3, C). The corrugated waste tube guides the solvent to the next lower positioned module's leak tray and sensor.

The waste tube of the sampler's needle wash port (4) guides solvents to waste.

The condense drain outlet of the autosampler cooler (5) guides condensate to waste.

The waste tube of the purge valve (6) guides solvents to waste.

The waste tube connected to the leak pan outlet on each of the bottom instruments (7) guides the solvent to a suitable waste container.

# 2
# Site Requirements and Specifications

Site Requirements 22
Physical Specifications 25
Performance Specifications 26

This chapter provides information on environmental requirements, physical and performance specifications.

# Site Requirements

A suitable environment is important to ensure optimal performance of the instrument.

## Power Considerations

The module power supply has wide ranging capability. It accepts any line voltage in the range described in Table 1 on page 25. Consequently there is no voltage selector in the rear of the module. There are also no externally accessible fuses, because automatic electronic fuses are implemented in the power supply.

**WARNING**

**Hazard of electrical shock or damage of your instrumentation**

**can result, if the devices are connected to a line voltage higher than specified.**

➔ Connect your instrument to the specified line voltage only.

**WARNING**

**The module is partially energized when switched off, as long as the power cord is plugged in.**

**Repair work at the module can lead to personal injuries, e.g. electrical shock, when the cover is opened and the module is connected to power.**

➔ Always unplug the power cable before opening the cover.

➔ Do not connect the power cable to the instrument while the covers are removed.

**CAUTION**

Inaccessible power plug.

In case of emergency it must be possible to disconnect the instrument from the power line at any time.

➔ Make sure the power connector of the instrument can be easily reached and unplugged.

➔ Provide sufficient space behind the power socket of the instrument to unplug the cable.

## Power Cords

Different power cords are offered as options with the module. The female end of all power cords is identical. It plugs into the power-input socket at the rear. The male end of each power cord is different and designed to match the wall socket of a particular country or region.

**WARNING**

**Absence of ground connection or use of unspecified power cord**

**The absence of ground connection or the use of unspecified power cord can lead to electric shock or short circuit.**

➔ Never operate your instrumentation from a power outlet that has no ground connection.

➔ Never use a power cord other than the Agilent Technologies power cord designed for your region.

---

**WARNING**

**Use of unsupplied cables**

**Using cables not supplied by Agilent Technologies can lead to damage of the electronic components or personal injury.**

➔ Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.

---

**WARNING**

**Unintended use of supplied power cords**

**Using power cords for unintended purposes can lead to personal injury or damage of electronic equipment.**

➔ Never use the power cords that Agilent Technologies supplies with this instrument for any other equipment.

---

## Bench Space

The module dimensions and weight (see Table 1 on page 25) allow you to place the module on almost any desk or laboratory bench. It needs an additional 2.5 cm (1.0 inches) of space on either side and approximately 8 cm (3.1 inches) in the rear for air circulation and electric connections.

If the bench shall carry a complete HPLC system, make sure that the bench is designed to bear the weight of all modules.

The module should be operated in a horizontal position.

## Condensation

**CAUTION**

Condensation within the module

Condensation can damage the system electronics.

➔ Do not store, ship or use your module under conditions where temperature fluctuations could cause condensation within the module.

➔ If your module was shipped in cold weather, leave it in its box and allow it to warm slowly to room temperature to avoid condensation.

# Physical Specifications

**Table 1** Physical Specifications

| Type | Specification | Comments |
|---|---|---|
| Weight | 15.2 kg (33.4 lbs) | |
| Dimensions (height × width × depth) | 200 x 345 x 435 mm (8 x 13.5 x 17 inches) | |
| Line voltage | 100 – 240 V~, ± 10 % | Wide-ranging capability |
| Line frequency | 50 or 60 Hz, ± 5 % | |
| Power consumption | 220 VA / 180 W / 615 BTU/h | Maximum |
| Ambient operating temperature | 4–55 °C (39–131 °F) | |
| Ambient non-operating temperature | -40 – 70 °C (-40 – 158 °F) | |
| Humidity | < 95 % r.h. at 40 °C (104 °F) | Non-condensing |
| Operating altitude | Up to 2000 m (6562 ft) | |
| Non-operating altitude | Up to 4600 m (15092 ft) | For storing the module |
| Safety standards: IEC, CSA, UL | Installation category II, Pollution degree 2 | For indoor use only. |

# Performance Specifications

**Table 2** Performance specifications

| Type | Specification | Comments |
|---|---|---|
| Hydraulic system | Dual pistons in series pump with proprietary servo-controlled variable stroke design, power transmission by ball screws, smooth motion control of pistons for active damping. | |
| Settable flow range | 0.001—5 mL/min, in 0.001 mL/min increments. | Executed in 300 pL/st ep increments |
| Flow precision | ≤0.07 % RSD or 0.01 min SD, whatever is greater (0.2—5.0 mL/min). | Based on retention time at constant room temperature. |
| Flow accuracy | ±1 % or ± 10 µL/min, whatever is greater. | Pumping degassed $H_2O$ at 10 MPa (100 bar) |
| Maximum operating pressure | Operating range up to 120 MPa (1200 bar), up to 2 mL/min, ramping down to 80 MPa (800 bar) up to 5 mL/min. | |
| Pressure pulsation | <1 % amplitude or < 0.5 MPa (5 bar), whatever is greater. | At 1 mL/min water |
| Compressibility compensation | Automatic, pre-defined, based on mobile phase selection. | |
| Gradient formation | Low pressure quaternary mixing | |
| Delay volume | Standard configuration: <350 µL<br>With optional V380 Jet Weaver: <500 µL | |
| Composition range | Settable range: 0 – 100 % | Recommended range: 1 – 99 % or 5 µL/min per channel, whatever is greater. |

**Table 2** Performance specifications

| Type | Specification | Comments |
|---|---|---|
| Composition precision | <0.15 % RSD, or 0.02 min SD, whatever is greater (1 mL/min). | Based on retention time at constant room temperature |
| Composition accuracy | ±0.40 % absolute (1 – 99 % B, 0.5 – 2.0 mL/min with water/caffeine tracer, 400 bar) | |
| Integrated degassing unit | Number of channels: 4<br>Internal volume per channel: 1.5 mL | |
| Control | Agilent ChemStation for LC (C.01.04 or above)<br>OpenLAB (A.04.04)<br>Masshunter (B.05.01 or above) | |
| Local control | Agilent Instant Pilot (G4208A) (B.02.08 or above) | |
| Communications | Controller-area network (CAN), RS232C, APG remote: ready, start, stop and shutdown signals, LAN | |
| Safety and maintenance | Extensive diagnostics, error detection and display through Agilent LabAdvisor, leak detection, safe leak handling, leak output signal for shutdown of the pumping system. Low voltage in major maintenance areas. | |
| GLP features | Early maintenance feedback (EMF) for continuous tracking of instrument usage in terms of seal wear and volume of pumped mobile phase with pre-defined and user settable limits and feedback messages. Electronic records of maintenance and errors. | |
| Housing | All materials recyclable. | |

# 3
# Installing the Module

Unpacking the Module 30
Optimizing the Stack Configuration 31
One Stack Configuration 31
Two Stack Configuration 34
Installation Information on Leak and Waste Handling 36
Removing the Transport Foam 39
Installing the Pump 40
Flow Connections to the Pump 42
Installation of Seal Wash Function 45

This chapter gives information about the preferred stack setup for your system and the installation of your Agilent 1290 Infinity Quaternary Pump.

# Unpacking the Module

## Damaged Packaging

If the delivery packaging shows signs of external damage, please call your Agilent Technologies sales and service office immediately. Inform your service representative that the instrument may have been damaged during shipment.

**CAUTION**

"Defective on arrival" problems

If there are signs of damage, please do not attempt to install the module. Inspection by Agilent is required to evaluate if the instrument is in good condition or damaged.

➔ Notify your Agilent sales and service office about the damage.

➔ An Agilent service representative will inspect the instrument at your site and initiate appropriate actions.

## Delivery Checklist

Ensure all parts and materials have been delivered with your module. The delivery checklist is included to your shipment. For parts identification please check the illustrated parts breakdown in "Parts and Materials" on page 211. Please report any missing or damaged parts to your local Agilent Technologies sales and service office.

# Optimizing the Stack Configuration

If your module is part of a complete Agilent 1290 Infinity Liquid Chromatograph, you can ensure optimum performance by installing the following configurations. These configurations optimize the system flow path, ensuring minimum delay volume.

For other possible configurations, please refer to the Agilent 1290 Infinity System Manual.

## One Stack Configuration

Ensure optimum performance by installing the modules of the Agilent 1290 Infinity Quaternary LC System in the following configuration (see Figure 7 on page 32 and Figure 8 on page 33). This configuration optimizes the flow path for minimum delay volume and minimizes the bench space required.

The Agilent 1290 Infinity Quaternary Pump should always be installed at the bottom of the stack.




**Figure 7** Recommended stack configuration for 1290 Infinity with quaternary pump (front view)




**Figure 8** Recommended stack configuration for 1290 Infinity with quaternary pump (rear view)

## Two Stack Configuration

In case the autosampler thermostat is added to the system, a two-stack configuration is recommended, which places both heavy modules (1290 Infinity pump and thermostat) at the bottom of each stack and avoids high stacks. Some users prefer the lower height of this arrangement even without the autosampler thermostat. A slightly longer capillary is required between the pump and autosampler. (See Figure 9 on page 34 and Figure 10 on page 35).




**Figure 9** Recommended two stack configuration for 1290 Infinity with quaternary pump (front view)




**Figure 10** Recommended two stack configuration for 1290 Infinity with quaternary pump (rear view)

# Installation Information on Leak and Waste Handling

The Agilent 1200 Infinity Series has been designed for safe leak and waste handling. It is important that all security concepts are understood and instructions are carefully followed.

**WARNING**

**Toxic, flammable and hazardous solvents, samples and reagents**

**The handling of solvents, samples and reagents can hold health and safety risks.**

- ➔ When working with these substances observe appropriate safety procedures (for example by wearing goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet supplied by the vendor, and follow good laboratory practice.
- ➔ The volume of substances should be reduced to the minimum required for the analysis.
- ➔ Never exceed the maximal permissible volume of solvents (6 L) in the solvent cabinet.
- ➔ Do not use bottles that exceed the maximum permissible volume as specified in the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets.
- ➔ Arrange the bottles as specified in the usage guideline for the solvent cabinet.
- ➔ A printed copy of the guideline has been shipped with the solvent cabinet, electronic copies are available on the Internet.

**NOTE**

**Recommendations for Solvent Cabinet**

For details, see the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets.




**Figure 11** Leak and waste handling (overview - typical stack configuration as an example)

| | |
|---|---|
| 1 | Solvent cabinet |
| 2 | Leak pan |
| 3 | Leak pan's outlet port (A), leak funnel (B) and corrugated waste tube (C) |
| 4 | Waste tube of the sampler's needle wash |
| 5 | Condense drain outlet of the autosampler cooler |
| 6 | Waste tube of the purge valve |
| 7 | Waste tube |

1 Stack the modules according to the adequate stack configuration.

   The leak pan outlet of the upper module must be vertically positioned above the leak tray of the lower module, see Figure 11 on page 37.

2 Connect data and power cables to the modules, see section *Installing the Module* below.

3 Connect capillaries and tubes to the modules, see section *Flow Connections to the module* below or the relevant system manual.

**WARNING**

**Toxic, flammable and hazardous solvents, samples and reagents**

- ➔ Keep solvent path free from blockages.
- ➔ Keep the flow path closed (in case the pump in the system is equipped with a passive inlet valve, solvent may leak out due to hydrostatic pressure, even if your instrument is off).
- ➔ Avoid loops.
- ➔ Tubes must not sag.
- ➔ Do not bend tubes.
- ➔ Do not immerse tube end in waste liquid.
- ➔ Do not intubate tubes in other tubes.
- ➔ For correct tubing follow instructions on label attached to the module.

Warning
Attention
Warnung
Attenzione
Advertencia
Disconnect power cord before moving.
Débranchez le cordon secteur avant tout déplacement.
Entfernen Sie das Netzkabel bevor Sie das Gerät bewegen.
Scollegare il cavo di alimentazione prima di spostare lo strumento.
Desconectar el aparato de la red antes de moverlo.

**Figure 12** Warning label (illustration for correct waste tubing)

# Removing the Transport Foam

**1** Open the front cover of the module.



**2** Carfully remove the protective foam.



**3** Close the front cover.



# Installing the Pump

| | | |
|---|---|---|
| **Parts required** | **#** | **Description** |
| | 1 | Pump |
| | 1 | Power cord |
| | 1 | Agilent Control Software and/or Instant Pilot G4208 |
| **Preparations** | Locate bench space | |
| | Provide power connections | |
| | Unpack the pump | |

1 Place the module on the bench in a horizontal position.

2 Ensure the power switch on the front of the module is OFF (switch stands out).




**Figure 13** Front view of the quaternary pump

3 Connect the power cable to the power connector at the back of the module.

4 Connect the required interface cables to the rear of the pump.




**Figure 14** Rear view of the quaternary pump

**NOTE** In an Agilent 1290 Infinity System, the individual modules are connected by CAN cables. An Agilent 1200 Series Instant Pilot can be connected to the CAN bus of any module. Connection to an Agilent data system is established through the built-in LAN port of the detector. The LAN port of the detector must be used as the detector generates the highest data rate of all modules. For more information about connecting the Instant Pilot or Agilent Data System, please refer to the respective user manual. For setting up the LAN access, see "LAN Configuration" on page 265.

5 Turn on the power by pushing the button at the lower left hand side of the module.

The power button stays pressed in and the status LED should be green.

**NOTE** When the line power button stands out and the green light is off, the module is turned off.

**NOTE** The module was shipped with default configuration settings. For changing these settings, refer to section *Setting the 8-bit configuration switch*.

# Flow Connections to the Pump




The pump is shipped with tubing and capillary connections installed between degassing unit, MCGV, pump heads, pressure sensor, and Multi Purpose Valve. This section describes the installation of additional flow connections.

| **Parts required** | **p/n** | **Description** |
|---|---|---|
| | | Other modules |
| | G4220-68755 | Accessory Kit |
| | 5067-4644 | Solvent Cabinet Kit 1290 Infinity Pump |

**Preparations** Pump is installed in the LC system.

**WARNING**

**Toxic, flammable and hazardous solvents, samples and reagents**

**The handling of solvents, samples and reagents can hold health and safety risks.**

➔ When working with these substances observe appropriate safety procedures (for example by wearing goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet supplied by the vendor, and follow good laboratory practice.

➔ The volume of substances should be reduced to the minimum required for the analysis.

➔ Do not operate the instrument in an explosive atmosphere.

1 Remove the front cover by pressing the snap fasteners on both sides.

2 Place the solvent cabinet on top of the UHPLC stack.

3 Put the bottle-head assemblies into empty solvent reservoirs and place the bottle in the solvent cabinet.

4 Route tubing connections along the left side of the UHPLC stack using tube clips.

5 Connect the inlet tubes of the bottle-head assemblies to the inlet connectors A to D at the left hand side of the vacuum degasser. Fix the tubes in the tubing grommets of the pump.



6 Connect the capillary from the autosampler to port 4 of the Multi Purpose Valve.

7 Connect the waste tubing to port 7 of the Multi Purpose Valve and place it in your waste system.




8 If the pump is not part of an Agilent 1290 Infinity system stack or placed on the bottom of a stack, connect the waste tube to the waste outlet of the pump leak handling system.

9 Fill solvent reservoirs with your mobile phase.

10 Fill solvent lines with a syringe; prime, purge and condition your pump before first use.

# Installation of Seal Wash Function




The 1290 Infinity Quaternary Pump has a built-in seal wash function. The Seal Wash Function is recommended when using buffers or other non-volatile solvents or additives that could deposit on pistons and seals. It is used for regularly cleaning these parts automatically.

1 Place a wash solvent reservoir into the solvent cabinet. A mixture of distilled water and isopropanol (90/10) is a good choice for many applications.

2 Put the solvent inlet tube into the solvent reservoir, close it and connect the tube to the seal wash pump.

3 Route the outlet of the wash tube into a waste container.

# 4
# Using the Pump

Leak and Waste Handling 48
Best Practices 49
- Daily / Weekly tasks 49
- Power up / Shut-down the pump 49
- Prepare the pump 50
- How to deal with solvents 51
- Select channels for Multi-Channel Gradient Valve (MCGV) 51
- Optional Inline Filter 52

Setting up the Pump with the Instrument Control Interface 53
- Overview 53
- Instrument Configuration 53
- The Pump User Interface (Dashboard Panel) 55
- Control Settings 58
- Method Parameter Settings 60

Flushing the Filter 66
Solvent Information 68
Algae Growth in HPLC Systems 75
- How to Prevent and-or Reduce the Algae Problem 75

This chapter explains the operational parameters of the Agilent 1290 Infinity Quaternary Pump.

# Leak and Waste Handling

**WARNING**

**Toxic, flammable and hazardous solvents, samples and reagents**

**The handling of solvents, samples and reagents can hold health and safety risks.**

- When working with these substances observe appropriate safety procedures (for example by wearing goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet supplied by the vendor, and follow good laboratory practice.
- The volume of substances should be reduced to the minimum required for the analysis.
- Do not operate the instrument in an explosive atmosphere.
- Never exceed the maximal permissible volume of solvents (6 L) in the solvent cabinet.
- Do not use bottles that exceed the maximum permissible volume as specified in the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets.
- Arrange the bottles as specified in the usage guideline for the solvent cabinet.
- A printed copy of the guideline has been shipped with the solvent cabinet, electronic copies are available on the Internet.
- Ground the waste container.
- The residual free volume in the appropriate waste container must be large enough to collect the waste liquid.
- Check the filling level of the waste container regularly.
- To achieve maximal safety, check the correct installation regularly.
- Do not use solvents with an auto-ignition temperature below 200 °C (392 °F).

**NOTE**

**Recommendations for Solvent Cabinet**

For details, see the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets.

For details on correct installation, see section *Installation Information on Leak and Waste Handling* in the service manual.

# Best Practices

## Daily / Weekly tasks

### Daily tasks

- Replace mobile phase based on water/buffer.
- Replace organic mobile phase latest every 2nd day.
- Check seal wash solvent.
- Run conditioning with composition of your application.

### Weekly tasks

- Change seal wash solvent (10 % / 90 % isopropanol/water) and bottle.
- Flush all channels with water to remove salt deposits.
- Visually inspect solvent filters. Clean or exchange if necessary.

## Power up / Shut-down the pump

### Power up the pump

- Use new or different mobile phase (as required).
- Purge pump heads with 2.5 – 3 mL/min for 5 min.
- Condition pump heads for 10 – 20 min.

### Long-term shut-down of the pump

- Flush system with water to remove buffer.
- Use recommended solvents to store the system.
- Power off the pump or system.

# Prepare the pump

## Purge

Use the Purge function to:

- fill the pump,
- exchange a solvent,
- remove air bubbles in tubes and pump heads.

## Condition

Use the Conditioning function:

- daily when starting the pump,
- to minimize pressure ripple by dissolving air bubbles in pump heads.

**NOTE** Condition your complete system with solvents and composition of your application (for example 50 %/50 % A/B at 0.5 mL/min.

## Seal wash

Using the seal wash function is recommended when using buffers or other non- volatile solvents or additives that could deposit on pistons and seals. The seal wash function regularly cleans these parts automatically.

**CAUTION** Contaminated seal wash solvent

➔ Do not recycle seal wash solvent to avoid contamination.

➔ Weekly exchange seal wash solvent.

## How to deal with solvents

- Use clean bottles only.
- Exchange water-based solvents daily.
- Select solvent volume to be used up within 1 – 2 days.
- Use only HPLC-grade solvents and water filtered through 0.2 µm filters.
- Label bottles correctly with bottle content, and filling date / expiry date.
- Use solvent inlet filters.
- Reduce risk of algae growth: use brown bottles for aqueous solvents, avoid direct sunlight.

## Select channels for Multi-Channel Gradient Valve (MCGV)

- Use lower channels (A and/or D) for buffer solutions.
- Regularly flush all MCGV channels with water to remove possible salt deposits.
- Check compatibility of buffers and organic solvents to avoid precipitation.

## Optional Inline Filter

The pump can be equipped with an additional inline filter ( In-Line Filter Assembly for 1290 Infinity Quaternary Pump (5067-5407)) with a nominal filter pore size of 0.3 µL.

Advantages of the inline filter:

- Very small internal volume
- Specified for working at high pressures
- Possibility of back-flushing the filter

Using the inline filter is recommended:

- to protect the downstream system from blockages,
- for solvent combinations that can form precipitation after mixing,
- for applications running with buffers.

General hints for effective usage of the inline filter:

- filter solvents before usage,
- follow best practices ,
- back-flush the filter weekly,
- exchange the filter frit on a regular basis.

**CAUTION**

Damage to the valve

➔ Use the filter flush mode only if the optional inline filter is installed.

See Technote G4226-90131 for further reference.

# Setting up the Pump with the Instrument Control Interface

## Overview

Parameters described in following sections is offered by the instrument control interface and can usually be accessed through Agilent instrument control software. For details, please refer to manuals and online help of respective user interfaces.

## Instrument Configuration

Use the **Instrument Configuration** dialog box to examine and, if necessary, modify your instrument configuration. The **Configurable Modules** panel contains a list of all modules available for configuration. The **Selected Modules** panel contains the list of configured modules.

**Auto Configuration**: Under **Communication settings**, select either the **Host Name** option or the **IP address** option and enter the appropriate value for the host computer to enable automatic detection of the hardware configuration. The system configures the instrument automatically with no further manual configuration necessary.

The Quaternary Pump configuration parameters are in two sections:

- **Communication**
- **Options**

**Communication**: The parameters in this dialog box are detected automatically during autoconfiguration.

- **Device name**,
- **Type ID**,
- **Serial number**,
- **Firmware revision**,
- Button **Connection settings**

**Options**:

- **Pressure Unit**:

  select the pressure units from the drop-down list (bar, psi or MPa).

- **Seal wash installed**:

  This check box is marked to indicate that an optional seal wash has been detected during autoconfiguration.

- **Installed mixer**:

  The installed mixer is detected during autoconfiguration. For manual configuration, click the down-arrow and select the installed mixer from the list or choose **No mixer installed**.

**Configure Solvent Type Catalogs**: Displays the **Solvent Type Catalogs** dialog box, which allows you to import and export solvent calibration data. See "Importing Solvent Calibration Tables" on page 86.

Please refer to the online help of your user interface for more detailed information.

# The Pump User Interface (Dashboard Panel)

## Module Graphic

The items in the pump graphic have the following meaning and function:

| | |
|---|---|
| | Indicates that an External Contacts board is installed. |
| | The level of solvent in the bottle is denoted by the green area; when the solvent level falls below the specified volume, the area turns yellow; when the bottle is empty, the area turns red. Clicking on the solvent bottle displays the **Bottle Fillings** dialog box. The tooltip for the bottle shows the solvent name. |
|  | The pressure setpoints. The red line shows the current maximum pressure limit; the green area shows the current pressure (also shown as text). |

The current solvent flow rate (in mL/min) is displayed above the pressure display.

## Instrument Signals

The following pump signals are displayed:

| | |
|---|---|
| **Flow** | The current solvent flow rate (in mL/min). |
| **Pressure** | The current pump ressure (in bar, psi or MPa, see "Instrument Configuration" on page 53). |
| **Tuning** | Indicates the tuning efforts of 1290 Infinity pumps. For pumps operating as expected, the signal should stay in a range of -1 to +1 within the full scale of -2 to +2. |
| **Pressure Limit** | The current maximum pressure limit. |
| **Composition A:B** | The contributions of channels A and B to the current solvent composition. |
| **Composition C:D** | The contributions of channels C and D to the current solvent composition. |
| **Mixer** | The installed mixer type. |
| **Valve position** | The current valve position. |

## Context Menu

The context menu of the dashboard panel contains the following commands:

| | |
|---|---|
| **Control** | Displays the pump's **Control** dialog box. |
| **Method** | Displays the pump's **Method Setup** dialog box. |
| **Set Error Method** | Sets the method that is loaded if an error occurs to the method that is currently available in the hardware. |
| **Identify Device** | Causes the LED on the front of the module to blink for a few seconds. |
| **Switch Pump On/Off** | Toggles the status of the pump, on or off. |
| **Bottle Fillings** | Displays the **Bottle Fillings** dialog box. |
| **Purge On/Off** | Allows you to control the purging of the system. |
| **Prime On/Off** | Allows you to prime the pump heads for initially drawing solvent. |
| **Conditioning On/Off** | Allows you to switch pump conditioning on and off. The conditioning function is useful for removing small air bubbles inside the pump flow path. |
| **Flush Filter On/Off** | Allows you to flush a clogged inline filter, which is connected to the Multi Purpose Valve, see "Filter Flush Mode" on page 17. Use the pump self-test for checking the filter back pressure. Do not use this option if no filter is installed! |

## Control Settings

The Quaternary Pump control parameters are in six sections:

- **Pump**
- **Seal Wash**
- **Automatic Turn On**
- **Purge**
- **Prime**
- **Conditioning**

**Table 3** Pump control parameters

| Parameter | Limits | Description |
|---|---|---|
| **Pump** | | Enables you to switch the pump **On**, **Off** or to a **Standby** condition.<br>In the **Standby** condition, the pump motor is still active, and when the pump is switched on again, does not need to be re-initialized. |
| **Seal Wash** | | The seal wash can be set up to be run once (**Single wash**) or periodically (**Periodic**).<br>• **Off**: no seal wash is used.<br>• **Single wash**: the seal will be purged for a specified time.<br>• **Periodic**: a periodic wash will be applied for a defined period in minutes.<br>The option is available only when the pump has seal wash capability. The seal wash capability is detected by the module itself. If seal wash is installed, it is recommended to use it in order to increase the primary seal lifetime. |
| **Seal Wash Run Mode** | | Allows you to define when to use the seal wash:<br>• **Off**: The seal wash is inactive.<br>• **On when pump is on**: The seal wash is active only when the pump is on.<br>• **On all the time**: The seal wash is active when the pump is on or in standby mode. |
| **Automatic Turn On** | | Module can be turned on at a specified date/time. This feature can only be used if the module power switch is turned on. |

**Table 3** Pump control parameters

| Parameter | Limits | Description |
|---|---|---|
| **Purge** | **Time**: 0 – 100.00 min in steps of 0.01 . **Flow**: 0.000 – 5.000 mL/min for each channel, in steps of 0.001 | Setup and activation of **Purge** parameters. The automatic purge valve can be used for purging the system. The process has been automated for ease of use.<br>• **Off**: Turns off the purge.<br>• **On**: The device is purged.<br>• **Purge Flow**, **Time** and **Composition** during purge have to be defined.<br>As soon as the duration time of the purge ends, the module automatically switches to analytical conditions again. |
| **Prime** | | Select **On** to start priming, **Off** to turn priming off.<br>The **Prime** function is helpful for filling empty solvent lines or if air has entered the pump heads. The module draws solvent, at high speed with both pump drives simultaneously, and dispenses it against the waste position of the Multi Purpose Valve. This is done 20 times, before the process comes to an end. |
| **Conditioning** | at least 200 bar (> 500 bar is better). | Use this function if you see excessive pressure or composition ripple, and you are sure that the solvent type (aqueous/organic or specific solvent/solvent mix) is correctly set, and there is no evidence of leakage in the pump.<br>Conditioning may be necessary if the pump may contain air, for example after running out of solvent, after a long period of standby or after service or repair. |

## Method Parameter Settings

The Quaternary Pump method setup parameters are in nine sections:

- **Flow**
- **Solvents A to D**
- **Stoptime**
- **Posttime**
- **Pressure Limits**
- **Timetable**
- **Advanced**
- **Blend Assist**
- **External Contacts**

**Table 4** Method parameters

| Parameter | Limits | Description |
|---|---|---|
| **Flow** | 0.00 – 5.00 mL/min in steps of 0.001 . Recommended flow range: 0.05 – 5.00 mL/min . | The flow is the rate of movement of eluent along the column. It is important that the flow rate is kept constant to ensure precise retention time, and peak measurements. Variations in flow rate can occur as a result of the partial failure of the pumping system, air in the pumping system, a change in the mobile phase viscosity or a temperature change. |
| **Enable Blend Assist** | | Mark this check box to switch on Blend Assist, which allows you to set up solvent mixtures from stock solutions. When this check box is marked, the **Blend Assist** section of the method setup is available. |

**Table 4** Method parameters

| Parameter | Limits | Description |
|---|---|---|
| **Solvents** | | **Blend Assist Disabled**: When **Blend Assist** is disabled, you can set the percentages of solvents B, C and D to any value from 0 through 100 %. **Solvent A** always delivers the remaining volume: 100 - (%B + %C + %D). The check boxes allows you to turn the solvent channels on (checked) or off (cleared). Click the solvent name down arrow and select the solvent from the list of calibrated solvents and solvent mixtures. For solvent mixtures, specify the percentage of additive. You can enter your own name for the solvent or solvent mixture in the adjacent field.<br>**Blend Assist Enabled**: When **Blend Assist** is enabled, the table shows the solvent blends that have been set up in the **Blend Assist** section of the method setup.<br>• **Solvent**: The solvent or blend of solvents as set up in the Blend Assist section.<br>• **Used**: Mark this check box if you want to use this solvent or blend in the method.<br>• **%**: Enter the percentage of the solvent or blend in this field.<br>• **Name**: Type a name for the solvent or blend in this field. |
| **Stoptime** | 0.01 – 99999 min or **As Injector/No Limit** (an infinite run time). | The stoptime sets a time limit for your analysis. After the stoptime, all gradients are stopped and the pump parameters return to their initial values. The pump can be used as a stoptime master for the complete analytical system. The pump also stops the detectors if they have a **No Limit** stoptime setting. If no limit is given, a method will have to be stopped manually. |
| **Posttime** | 0.01 – 99999 min or **Off** (0.0 min ). | Your instrument remains in a not ready state during the posttime to delay the start of the next analysis. You can use the **Posttime** to allow your column to equilibrate after changes in solvent composition (for example after gradient elution). |

**Table 4** Method parameters

| Parameter | Limits | Description |
|---|---|---|
| **Pressure Limits** | **Max**: 1200 bar (17400 psi ) for flow rates up to 2 mL/min . For flow rates between 2 mL/min and 5 mL/min , the maximum pressure ramps down to 800 bar (11600 psi).<br>**Min**: any value between 0 and the upper pressure limit setting. | Sets the maximum and minimum pressure limits for the pump.<br>• **Max** is the maximum pressure limit at which the pump will switch itself off, protecting the analytical system against over-pressure.<br>• **Min** is the minimum limit at which the pump will switch itself off, for example, if any solvent reservoir is empty, this prevents system damage by pumping air. |
| **Timetable** | | See "Timetable Settings" on page 64 |
| **Advanced** | | See "Advanced Settings" on page 63 |
| **External Contacts** | | The **External Contacts** section enables you to set up the switching of the external contacts.<br>**NOTE**<br>The **External Contacts** section is present only when a BCD/external contacts board is installed. |

## Advanced Settings

The Quaternary Pump advanced method setup parameters are in five sections:

- **Minimum Stroke**
- **Compressibility**
- **Maximum Flow Gradient**
- **Primary Channel**
- **Mixer Selection**

**Table 5** Advanced method parameters

| Parameter | Limits | Description |
|---|---|---|
| **Minimum Stroke** | 20 – 100 µL | The Stroke Volume is used for optimizing between performance of the module and seal life time. For performance a low stroke volume is beneficial, as it divides disturbances into smaller packages, but a larger volume is extending the life time of the pump seals.<br>If **Automatic** is activated, the pump tries to achieve an optimized stroke volume for the Jet Weaver geometry. |
| **Compressibility** | | The compressibility of the mobile phase has an effect on the performance of the pump. For best flow accuracy and mixing performance, you can set the parameter according to the mobile phase being used.<br>**Use solvent types**:<br>• Select this check box (recommended) for using the enhanced and automatic compressibility calibration. Then select the calibrated solvent from the drop-down lists using the combo boxes in the **Solvents** section. Using this checkbox hides compressibility fields for manual settings.<br>• Clear this check box to display the compressibility fields, which allow you to enter manual compressibility values, which are constant over pressure. This setting is available for method backward compatibility e.g. from 1260 Infinity pumps. For best performance, use solvent types. |
| **Maximum Flow Gradient** | 1.000 – 1000.000 mL/min/min in steps of 0.001 mL/min/min<br>Default value: 100.000 mL/min/min | You can set a limit on the rate of change of the solvent flow to protect your analytical column. You can set individual values for **Flow ramp up** and **Flow ramp down**. |

**Table 5** Advanced method parameters

| Parameter | Limits | Description |
|---|---|---|
| **Primary Channel** | | Using **Automatic** is recommended.<br>The primary channel can be specified as A to D for optimizing highly specific methods. It is split up to deliver the first and last solvent package created by the MCGV in order to optimize composition precision. The primary channel does not change during a gradient, as long as the channel is used. Using **Automatic** chooses the channel with the highest percentage at start conditions before a gradient. |
| **Mixer Selection** | | Click the down-arrow and select the mixer to use from the list:<br>• **Use any mixer**: The currently installed mixer is used, irrespective of its type.<br>• **Do not use mixer**: The valve is set to bypass the mixer so that it is not in the flow path.<br>• **<Mixer Name>**: Only the specified mixer may be used; it the mixer is not found, the pump goes into a Not Ready state. |

## Timetable Settings

Use the **Timetable** to program changes in the pump parameters during the analysis by entering a time in the **Time** field and appropriate values in the following fields of the timetable. Changes in flow rate occur linearly from either time zero or the time of the last defined change; other parameters change instantaneously at the time defined in the timetable.

Show **Advanced Timetable** toggles the timetable display between standard mode and advanced mode.

The following parameters can be changed:

- **Change Contacts**
- **Change Flow**
- **Change Max. Pressure Limit**
- **Change Solvent Composition** - You can only use solvents, which have been enabled in the solvents section.
- **Function centric view** - This checkbox allows you displaying parameter changes instead of a time table.

## Blend Assist

The **Blend Assist** table allows you to blend two or more solvents or solvent mixtures from stock solutions. The blends must be of pure solvents or pure solvents with additives. For example, you can blend 100 % water with 10 % isopropanol in water.

- **Channel**: The channel name.
- **Type**: The type of solvent
  - **Solvent <n>**: Pure solvent
  - **Solvent <n> Additive**: Solvent mixture
- **Calibration**: Click the down arrow and select the solvent or solvent mixture from the list.
- **Name**: Enter a name for the solvent or solvent mixture in this field.
- **Stock conc.**: For solvent mixtures, specify the concentration of the additive in the stock solution in this field. Pure solvents are always 100 %.
- **Final conc.**: Enter the concentration of the additive that you want to achieve in this field. The pure solvent and solvent mixture will be blended to achieve the **Final conc.** For the relationship of stock concentration and concentration in the mixture, the composition accuracy needs to be considered (see "Performance Specifications" on page 26).
- **Conc. unit**: The concentration can be defined as mM (mmol/L) or as %.

# Flushing the Filter

For highest performance and robustness, the 1290 Infinity Quaternary Pump uses 3 solvent filters:

1 Solvent inlet filter, 20 µm pore size (5041-2168) as part of Bottle Head Assembly (G4220-60007) have a large pore size of about 20 µm and filter out particles before they reach the pump.




2 An outlet filter (average pore size 5 µm; Outlet filter 1290 Infinity Quaternary Pump (G4204-60004)) between pump head and pressure sensor filters out particles which may be created in the pump by wear of piston or wash seals.

This filter can be replaced as required.

3 An optional inline filter connected to the Multi Purpose Valve with a small pore size of about 0.3 µm ( In-Line Filter Assembly for 1290 Infinity Quaternary Pump (5067-5407)).

This filter can be flushed using the graphical user interface or replaced as required.




In the instrument control panel of Agilent user interfaces, use the context menu and select **Flush Filter On**, see also "Context Menu" on page 57.

# Solvent Information

## Introduction

Observe the following recommendations on the use of solvents.

- Follow recommendations for avoiding the growth of algae, see "Algae Growth in HPLC Systems" on page 75.
- Small particles can permanently block capillaries and valves. Therefore, always filter solvents through 0.4 µm filters.
- Avoid or minimize the use of solvents that may corrode parts in the flow path. Consider specifications for the pH range given for different materials like flow cells, valve materials etc. and recommendations in subsequent sections.

## Materials in Flow Path

Following materials are used in the flow path of this module:

| Part | Materials |
|---|---|
| Degasser chamber | TFE/PDD copolymer, PFA (internal tubings); PEEK (inlets); FEP (tubings); ETFE (fittings) |
| Ultra clean tubings[1] | PFA (tubings), PEEK (fittings) |
| Microfluidic structures[2] | SST |
| MCGV | PEEK, FEP, PFA, $Al_2O_3$-based ceramic, ruby, sapphire, SST |
| Passive inlet valve | SST, gold, ruby, $ZrO_2$-based ceramic, tantalum |
| Outlet valve | SST, gold, ruby, $ZrO_2$-based ceramic, tantalum |
| Pump head | SST |
| Pistons | $ZrO_2$-based ceramic |
| Piston/wash seals | UHMW-PE, SST |
| Pressure sensor | SST |
| Multi Purpose Valve | Polyimide, SST, DLC |

1 Ultra clean tubings are available for the use with high-end MS detectors. They are also compatible to THF.

2 Inlet Weaver, Jet Weaver, Heat Exchanger

# Material Information

Materials in the flow path are carefully selected based on Agilent's experiences in developing highest quality instruments for HPLC analysis over several decades. These materials exhibit excellent robustness under typical HPLC conditions. For any special conditions, please consult the material information section or contact Agilent.

## Disclaimer

Subsequent data were collected from external resources and are meant as a reference. Agilent cannot guarantee the correctness and completeness of such information. Data is based on compatibility libraries, which are not specific for estimating the long-term life time under specific but highly variable conditions of UHPLC systems, solvents, solvent mixtures and samples. Information can also not be generalized due to catalytic effects of impurities like metal ions, complexing agents, oxygen etc. Apart from pure chemical corrosion, other effects like electro corrosion, electrostatic charging (especially for non-conductive organic solvents), swelling of polymer parts etc. need to be considered. Most data available refers to room temperature (typically 20 – 25 °C, 68 – 77 °F). If corrosion is possible, it usually accelerates at higher temperatures. If in doubt, please consult technical literature on chemical compatibility of materials.

## PEEK

PEEK (Polyether-Ether Ketones) combines excellent properties regarding biocompatibility, chemical resistance, mechanical and thermal stability. PEEK is therefore the material of choice for UHPLC and biochemical instrumentation.

It is stable in a pH range between 1 – 12, and inert to many common solvents.

There is still a number of known incompatibilities with chemicals such as chloroform, methylene chloride, THF, DMSO, strong acids (nitric acid > 10 %, sulphuric acid > 10 %, sulfonic acids, trichloroacetic acid), halogenes or aequous halogene solutions, phenol and derivatives (cresols, salicylic acid etc.).

### Polyimide

Agilent uses semi-crystalline polyimide for rotor seals in valves and needle seats in autosamplers. One supplier of polyimide is DuPont, which brands polyimide as Vespel, which is also used by Agilent.

Polyimide is stable in a pH range between 1 and 10 and in most organic solvents. It is incompatible with concentrated mineral acids (e.g. sulphuric acid), glacial acetic acid, DMSO and THF. It is also degraded by nucleophilic substances like ammonia (e.g. ammonium salts in basic conditions) or acetates.

### Polyethylene (PE)

Agilent uses UHMW (ultra-high molecular weight)-PE/PTFE blends for yellow piston and wash seals, which are used in 1290 Infinity pumps and for normal phase applications in 1260 Infinity pumps.

Polyethylene has a good stability for most common inorganic solvents including acids and bases in a pH range of 1 to 12.5. It is compatible to many organic solvents used in chromatographic systems like methanol, acetonitrile and isopropanol. It has limited stability with aliphatic, aromatic and halogenated hydrocarbons, THF, phenol and derivatives, concentrated acids and bases. For normal phase applications, the maximum pressure should be limited to 200 bar.

### Tantalum (Ta)

Tantalum is inert to most common HPLC solvents and almost all acids except fluoric acid and acids with free sulfur trioxide. It can be corroded by strong bases (e.g. hydroxide solutions > 10 %, diethylamine). It is not recommended for the use with fluoric acid and fluorides.

### Stainless Steel (ST)

Stainless steel is inert against many common solvents. It is stable in the presence of acids and bases in a pH range of 1 to 12.5. It can be corroded by acids below pH 2.3. It can also corrode in following solvents:

- Solutions of alkali halides, their respective acids (for example, lithium iodide, potassium chloride, and so on) and aqueous solutions of halogens.
- High concentrations of inorganic acids like nitric acid, sulfuric acid and organic solvents especially at higher temperatures (replace, if your chromatography method allows, by phosphoric acid or phosphate buffer which are less corrosive against stainless steel).
- Halogenated solvents or mixtures which form radicals and/or acids, for example:

  $2\ CHCl_3 + O_2 \rightarrow 2\ COCl_2 + 2\ HCl$

  This reaction, in which stainless steel probably acts as a catalyst, occurs quickly with dried chloroform if the drying process removes the stabilizing alcohol.
- Chromatographic grade ethers, which can contain peroxides (for example, THF, dioxane, di-isopropylether). Such ethers should be filtered through dry aluminium oxide which adsorbs the peroxides.
- Solutions of organic acids (acetic acid, formic acid, and so on) in organic solvents. For example, a 1 % solution of acetic acid in methanol will attack steel.
- Solutions containing strong complexing agents (for example, EDTA, ethylene diamine tetra-acetic acid).
- Mixtures of carbon tetrachloride with 2-propanol or THF.

### Diamond-Like Carbon (DLC)

Diamond-Like Carbon is inert to almost all common acids, bases and solvents. There are no documented incompatibilities for HPLC applications.

### Fused silica and Quartz ($SiO_2$)

Fused silica is used in 1290 Infinity Flow Cells and capillaries. Quartz is used for classical flow cell windows. It is inert against all common solvents and acids except hydrofluoric acid and acidic solvents containing

fluorides. It is corroded by strong bases and should not be used above pH 12 at room temperature. The corrosion of flow cell windows can negatively affect measurement results. For a pH greater than 12, the use of flow cells with sapphire windows is recommended.

### Gold

Gold is inert to all common HPLC solvents, acids and bases within the specified pH range. It can be corroded by complexing cyanides and concentrated acids like aqua regia.

### Zirconium Oxide ($ZrO_2$)

Zirconium Oxide is inert to almost all common acids, bases and solvents. There are no documented incompatibilities for HPLC applications.

### Platinum/Iridium

Platinum/Iridium is inert to almost all common acids, bases and solvents. There are no documented incompatibilities for HPLC applications.

### Fluorinated polymers (PTFE, PFA, FEP, FFKM)

Fluorinated polymers like PTFE (polytetrafluorethylene), PFA (perfluoroalkoxy) and FEP (fluorinated ethylene propylene) are inert to almost all common acids, bases, and solvents. FFKM is perfluorinated rubber, which is also resistant to most chemicals. As an elastomer, it may swell in some organic solvents like halogenated hydrocarbons.

TFE/PDD copolymer tubings, which are used in all Agilent degassers except G1322A, are not compatible with fluorinated solvents like Freon, Fluorinert, or Vertrel. They have limited life time in the presence of Hexafluoroisopropanol (HFIP). To ensure the longest possible life with HFIP, it is best to dedicate a particular chamber to this solvent, not to switch solvents, and not to let dry out the chamber. For optimizing the life of the pressure sensor, do not leave HFIP in the chamber when the unit is off.

### Sapphire, Ruby and $Al_2O_3$-based ceramics

Sapphire, ruby and ceramics based on aluminum oxide $Al_2O_3$ are inert to almost all common acids, bases and solvents. There are no documented incompatibilities for HPLC applications.

# Solvent Handling

## Handling of Buffers

The following recommendations should be observed when using buffer solutions:

- Buffers and aqueous solutions are possible sources of algae contamination, for avoiding related problems, please read "Algae Growth in HPLC Systems" on page 75.
- For buffer concentrations of 0.1 M or higher using the seal wash function periodically with a runtime of 0.3 min every 3 min is strongly recommended.
- Filter buffer solutions to avoid increased wear or blockages that are caused by undissolved crystals. Always use solvent inlet filters.
- Avoid conditions where mixing of buffers and organic solvents may cause precipitation, as this impairs the reproducibility of chromatographic experiments and may also reduce the system life time. For example in reversed phase chromatography, avoid buffers (especially phosphate buffers) with a concentration higher than 20 mmol/L. For phosphate buffers, avoid compositions containing more than 65 % acetonitrile or other organic solvents.
- Use a minimum flow rate of 5 µL/min or 1 % composition per solvent channel (whatever is greater) to avoid cross-flow. Cross-flow can be caused by micro leaks in pump heads and can result in buffer precipitation in pump heads, channel blocking, or reduced pump head life time through wear of seals and pistons.
- When installing tubing connections to the MCGV, use lower channels (A/D) for aqueous solvents and upper channels for organic solvents. This will re-dissolve precipitates more easily.
- Never leave buffers in a system without flow. Before shutting down a system, flush it extensively with warm water to avoid clogging of valves, capillaries, or flow cells or reducing the life time of your column. If the system is not used for some time, for example more than a day depending on lab temperature, fill all solvent lines with organic solvent or water with at least 10 % isopropanol.
- Regularly maintain the LC system.

## Handling of Acetonitrile

Acetonitrile is a solvent that is frequently used in reversed-phase chromatography. Despite of its common use, it can be a source of issues if not handled correctly. Acetonitrile degrades through polymerization and such polymers can stick to surfaces in LC systems and e.g. cause issues with valve performance and therefore affect retention time precision. Polymers can also show up as background noise in MS detectors.

When using acetonitrile:

- Use high-quality solvents from renowned suppliers.
- Use fresh solvents and filter them.
- Minimize exposure to light and air/oxygen.
- Choose a bottle size which fits to your application and usage.
- Acids accelerate polymerization. If possible avoid such additives or refresh solvents more frequently.
- Pure acetonitrile polymerizes faster. If your application allows, add about 5 % water and adjust gradient compositions.
- Do not leave acetonitrile in unused systems to avoid aging. If not in use, flush all solvent lines with a mixture of water and 10 % isopropanol.
- In case of blocked valves, flush the system with hot water. Knock at valves, flush them (see "Releasing a Stuck Inlet Valve" on page 152) or ultrasonicate them, e.g. in methanol.

## Handling of Acids

Acids can corrode stainless steel and other materials in the flow path of LC systems. For stainless steel, the minimum pH is 2.3 for corrosive acids and pH 1 for non-corrosive acids.

Please note that for non-volatile acids like phosphoric acid or perchloric acid concentrations increase after evaporation of water. This means that originally diluted acids can damage parts over time, e.g. because of liquid, which has left the solvent path through micro leaks. Such systems should be flushed regularly with pure water and may require shorter maintenance cycles. Using the seal wash function should be considered for protecting pump heads.

# Algae Growth in HPLC Systems

The presence of algae in HPLC systems can cause a variety of problems that may be incorrectly diagnosed as instrument or application problems. Algae grow in aqueous media, preferably in a pH range of 4-8. Their growth is accelerated by buffers, for example phosphate or acetate. Since algae grow through photosynthesis, light will also stimulate their growth. Even in distilled water small-sized algae grow after some time.

### Instrumental Problems Associated With Algae

Algae deposit and grow everywhere within the HPLC system causing:

- Blocked solvent filters or deposits on inlet or outlet valves resulting in unstable flow, composition or gradient problems or a complete failure of the pump.
- Small pore high pressure solvent filters, usually placed before the injector to plug resulting in high system pressure.
- PTFE frits blockage leading to increased system pressure.
- Column filters to plug giving high system pressure.
- Flow cell windows of detectors to become dirty resulting in higher noise levels (since the detector is the last module in the flow path, this problem is less common).

## How to Prevent and-or Reduce the Algae Problem

- Always use freshly prepared solvents, especially use demineralized water which was filtered through about 0.2 µm filters.
- Never leave mobile phase in the instrument for several days without flow.
- Always discard old mobile phase.
- Use the amber solvent bottle (Solvent bottle, amber (9301-1450)) supplied with the instrument for your aqueous mobile phase.
- If possible add a few mg/l sodium azide or a few percent organic solvent to the aqueous mobile phase.

# 5 How to Optimize the Performance of Your Module

Using the Degasser 78
Operational Hints for the Multi Channel Gradient Valve (MCGV) 79
Delay Volume and Extra-Column Volume 80
Delay Volume 80
How to Configure the Optimum Delay Volume 81
How to Achieve Higher Resolution 83
Using Solvent Calibration Tables 86

This chapter gives hints on how to optimize the performance or use additional devices.

# Using the Degasser

The quaternary pump has a built-in degasser, which should always be included to the flow path.

# Operational Hints for the Multi Channel Gradient Valve (MCGV)

In a mixture of salt solutions and organic solvent the salt solution might be well dissolved in the organic solvent without showing precipitations. However in the mixing point of the gradient valve, at the boundary between the two solvents, micro precipitation is possible. Gravity forces the salt particles to fall down. Normally the A channel of the valve is used for the aqueous/salt solution and the B channel of the pump is used for the organic solvent. If used in this configuration the salt will fall back into the aqueous solution and will be dissolved. When using the pump in a different configuration (e.g., D - salt solution, A - organic solvent) the salt can fall into the port of the organic solvent and may lead to performance problems.

**NOTE** When using salt solutions and organic solvents it is recommended to connect the salt solution to one of the bottom ports of the MCGV and the organic solvent to one of the upper gradient valve ports. It is best to have the organic channel directly above the salt solution channel. Regular flushing with water of all MCGV channels is recommended to remove all possible salt deposits in the valve ports.

**NOTE** Precipitations formed during the mixing of buffers and organic solvents which do not dissolve salts may cause a loss of pump performance (flow/retention time stability), a blockage or internal leak of the pump. Avoid the use of such solvent combinations, as they can cause irreproducible chromatographic results.

# Delay Volume and Extra-Column Volume

The *delay volume* is defined as the system volume between the point of mixing in the pump and the top of the column.

The *extra-column volume* is defined as the volume between the injection point and the detection point, excluding the volume in the column.

## Delay Volume

In gradient separations, this volume causes a delay between the mixture changing in the pump and that change reaching the column. The delay depends on the flow rate and the delay volume of the system. In effect, this means that in every HPLC system there is an additional isocratic segment in the gradient profile at the start of every run. Usually the gradient profile is reported in terms of the mixture settings at the pump and the delay volume is not quoted even though this will have an effect on the chromatography. This effect becomes more significant at low flow rates and small column volumes and can have a large impact on the transferability of gradient methods. It is important, therefore, for fast gradient separations to have small delay volumes, especially with narrow bore columns (e.g., 2.1 mm i.d.) as often used with mass spectrometric detection.

The delay volume in a system includes the volume in the pump from the point of mixing, connections between pump and autosampler, volume of the flow path through the autosampler and connections between autosampler and column.

For the 1290 Infinity Quaternary Pump, all pump parts downstream the MCGV contribute to the delay volume, i.e. inlet weaver, pump heads, capillary connections, filters and the optional Jet Weaver.

# How to Configure the Optimum Delay Volume

The design of the 1290 Infinity Quaternary Pump offers a strongly decreased delay volume compared to standard 600 bar pressure pumps. For the 1290 Infinity Quaternary Pump, mixing is done in the multi-channel gradient valve at ambient pressure. As all pump parts in the flow path after mixing contribute to the delay volume, this includes also pump heads of the quaternary pump, flow connections, filters, mixers etc. Therefore the delay volume of a quaternary pump is by design larger than that of a binary pump.

All listed components including the inlet weaver and pump heads ensure a good mixing performance resulting in excellent composition precision and accuracy, highly reproducible retention times and low baseline noise. This ensures best results for most applications.

Per default, the 1290 Infinity Quaternary Pump does not require and include a Jet Weaver, as solvents are mixed in the MCGV and mixing is further improved in the inlet weaver, pump heads and subsequent parts in the flow path. Therefore, no Jet Weaver is required for most applications.

The V380 Jet Weaver high performance mixer is optionally available for demanding applications, which use solvents in different channels (for example A versus B), that differ strongly in their UV/Vis absorption, for example by using trifluoroacetic acid (TFA) as a modifier, which has a high absorbance. Solvent packages created by the pump may persist until the solvent reaches the detector flow cell. Absorption fluctuations can then show up as baseline noise, also referred to as mixing noise. Applications like impurity quantitation or lowest level compound detection require minimizing this noise. The V380 Jet Weaver strongly improves mixing and therefore reduces baseline noise and improves sensitivity in detection. Patented Agilent microfluidic technology offers high mixing performance at a low internal volume of 380 µL, which is the physical volume of all channels. It contributes with 150 µL to the pump delay volume (< 350 µL without Jet Weaver), which is the partial mixer volume that creates a composition change corresponding to the delay volume.



**Figure 15** The Jet Weaver mixer

The installation procedure is illustrated in “Installing the Jet Weaver” on page 144.

# How to Achieve Higher Resolution

Increased resolution in a separation will improve the qualitative and quantitative data analysis, allow more peaks to be separated or offer further scope for speeding up the separation. This section explains how resolution can be increased by examining the following points:

- Optimize selectivity
- Smaller particle-size packing
- Longer Columns
- Shallower gradients, faster flow

Resolution between two peaks is described by the resolution equation:

$$Rs = \frac{1}{4}\sqrt{N}\,\frac{(\alpha - 1)}{\alpha}\,\frac{(k_2 + 1)}{k_2}$$

where

- $R_S$=resolution,
- N=plate count (measure of column efficiency),
- $\alpha$=selectivity (between two peaks),
- $k_2$=retention factor of second peak (formerly called capacity factor).

The term that has the most significant effect on resolution is the selectivity, $\alpha$, and practically varying this term involves changing the type of stationary phase (C18, C8, phenyl, nitrile etc.), the mobile phase and temperature to maximize the selectivity differences between the solutes to be separated. This is a substantial piece of work which is best done with an automated method development system which allows a wide range of conditions on different columns and mobile phases to be assessed in an ordered scouting protocol. This section considers how to get higher resolution with any chosen stationary and mobile phases. If an automated method development system was used in the decision on phases it is likely that short columns were used for fast analysis in each step of the scouting.

The resolution equation shows that the next most significant term is the plate count or efficiency, N, and this can be optimized in a number of ways. N is inversely proportional to the particle size and directly proportional to the length of a column and so smaller particle size and a longer column will give a higher plate number. The pressure rises with the inverse square of the particle size and proportionally with the length of the column. This is the reason that the 1290 Infinity LC system was designed to go to 1200 bar so that it can run sub-two-micron particles and column length can be increased to 100 mm or 150 mm. There are even examples of 100 mm and 150 mm columns linked to give 250 mm length. Resolution increases with the square root of N so doubling the length of the column will increase resolution by a factor of 1.4. What is achievable depends on the viscosity of the mobile phase as this relates directly to the pressure. Methanol mixtures will generate more back pressure than acetonitrile mixtures. Acetonitrile is often preferred because peak shapes are better and narrower in addition to the lower viscosity but methanol generally yields better selectivity (certainly for small molecules less than about 500 Da). The viscosity can be reduced by increasing the temperature but it should be remembered that this can change the selectivity of the separation. Experiment will show if this leads to increase or decrease in selectivity. As flow and pressure are increased it should be remembered that frictional heating inside the column will increase and that can lead to slightly increased dispersion and possibly a small selectivity change both of which could be seen as a reduction in resolution. The latter case might be offset by reducing the temperature of the thermostat by a few degrees and again experiment will reveal the answer.

The van Deemter curve shows that the optimum flow rate through an STM column is higher than for larger particles and is fairly flat as the flow rate increases. Typical, close to optimum, flow rates for STM columns are: 2 ml/min for 4.6 mm i.d.; and 0.4 ml/min for 2.1 mm i.d. columns.

In isocratic separations, increasing the retention factor, k, results in better resolution because the solute is retained longer. In gradient separations the retention is described by $k^*$ in the following equation:

$$k^* = \frac{t_G}{\Delta\%B} \cdot \frac{F}{V_m} \cdot \frac{100}{S}$$

where:

- $k^*$ = mean k value,
- $t_G$ = time length of gradient (or segment of gradient) (min),
- F = flow (ml/min),
- $V_m$ = column delay volume,
- Δ%B = change in fraction of solvent B during the gradient,
- S = constant (ca. 4-5 for small molecules).

This shows that k and hence resolution can be increased by having a shallower gradient (2 to 5 %/min change is a guideline), higher flow rate and a smaller volume column. This equation also shows how to speed up an existing gradient – if the flow is doubled but the gradient time is halved, $k^*$ remains constant and the separation looks the same but happens in half the time. Recently published research has shown how a shorter STM column (at temperatures above 40 °C) can generate higher peak capacity than a longer STM column by virtue of running it faster. (Refer to *Petersson et al., J.Sep.Sci, 31, 2346-2357, 2008, Maximizing peak capacity and separation speed in liquid chromatography*).

# Using Solvent Calibration Tables

## Importing Solvent Calibration Tables

RC.NET based Agilent graphical user interfaces (ChemStation, EZChrom Elite, OpenLab etc.) include data for most commonly used solvents in HPLC. This data contains solvent properties and is used for optimum pump control in order to ensure best flow and composition accuracy.

If your solvent is not included to the software, please check the Agilent web site http://www.chem.agilent.com/_layouts/agilent/downloadFirmware.aspx?whid=69761 for additional libraries (registration required), which also provides updates and optimized data.

If your solvent is neither available in the user interface nor in the library, please use generic solvents. "Generic aqueous" gives good results for most solvent mixtures with at least 50 % water, which have similar properties as pure water. For other solvents with high organic percentage, "Generic organic" gives a good approximation.

## Importing Solvent Calibration in ChemStation

1 Go to menu **Instrument > Instrument configuration**.

2 In the **Instrument Configuration** screen choose your module and click **Configure**.

3 Click **Configure Solvent Type Catalogs**.

4 In **Solvent Type Catalogs** click **Import**.

5 Navigate to the location of the solvent calibration table and click **Open**.

6 The new solvent will now appear in the **Solvent Type Catalogs**.

7 The imported solvent is now available for selection as a solvent type, see Table 4 on page 60.

# 6 Troubleshooting and Diagnostics

Overview of the Module's Indicators and Test Functions 88
Status indicators 89
Power Supply Indicator 89
Module Status Indicator 90
Available Tests vs User Interfaces 91
Agilent Lab Advisor Software 92

Overview about the troubleshooting and diagnostic features.

# Overview of the Module's Indicators and Test Functions

## Status Indicators

The module is provided with two status indicators which indicate the operational state of the module. The status indicators provide a quick visual check of the operation of the module.

## Error Messages

In the event of an electronic, mechanical or hydraulic failure, the module generates an error message in the user interface. For each message, a short description of the failure, a list of probable causes of the problem, and a list of suggested actions to fix the problem are provided (see chapter Error Information).

## Test Functions

A series of test functions are available for troubleshooting and operational verification after exchanging internal components (see Tests and Calibrations).

## Diagnostic Signals

The module has several signals (internal temperatures, voltages and currents of lamps) that can be used for diagnosing baseline problems. These can be added like normal signals in the Agilent ChemStation software.

# Status indicators

Two status indicators are located on the front of the module. The lower left indicates the power supply status, the upper right indicates the module status.

**Figure 16** Location of status indicators

## Power Supply Indicator

The power supply indicator is integrated into the main power switch. When the indicator is illuminated (*green*) the power is *ON*.

## Module Status Indicator

The module status indicator indicates one of six possible module conditions:

- When the status indicator is *OFF* (and power switch light is on), the module is in a *prerun* condition, and is ready to begin an analysis.
- A *green* status indicator, indicates the module is performing an analysis (*run* mode).
- A *yellow* indicator indicates a *not-ready* condition. The module is in a not-ready state when it is waiting for a specific condition to be reached or completed (for example, immediately after changing a set point), or while a self-test procedure is running.
- An *error* condition is indicated when the status indicator is *red*. An error condition indicates the module has detected an internal problem which affects correct operation of the module. Usually, an error condition requires attention (e.g. leak, defective internal components). An error condition always interrupts the analysis.

  If the error occurs during analysis, it is propagated within the LC system, i.e. a red LED may indicate a problem of a different module. Use the status display of your user interface for finding the root cause/module of the error.
- A *blinking* indicator indicates that the module is in resident mode (e.g. during update of main firmware).
- A *fast blinking* indicator indicates that the module is in a low-level error mode. In such a case try to re-boot the module or try a cold-start (see "Special Settings" on page 262). Then try a firmware update (see "Replacing Module Firmware" on page 206). If this does not help, a main board replacement is required.

# Available Tests vs User Interfaces

- Depending on the user interface, the available tests and the screens/reports may vary (see chapter *"Test Functions and Calibrations"*).
- Preferred tool should be the Agilent Lab Advisor software, see "Agilent Lab Advisor Software" on page 92.
- The Agilent ChemStation may not include any maintenance/test functions.
- Screenshots used within these procedures are based on the Agilent Lab Advisor software.

# Agilent Lab Advisor Software

The Agilent Lab Advisor Software is a standalone product that can be used with or without chromatographic data system. Agilent Lab Advisor helps to manage the lab for high-quality chromatographic results by providing a detailed system overview of all connected analytical instruments with instrument status, Early Maintenance Feedback counters (EMF), instrument configuration information, and diagnostic tests. By the push of a button, a detailed diagnostic report can be generated. Upon request, the user can send this report to Agilent for a significantly improved troubleshooting and repair process.

The Agilent Lab Advisor software is available in two versions:

- Lab Advisor Basic
- Lab Advisor Advanced

Lab Advisor Basic is included with every Agilent 1200 Infinity Series and Infinity II Series pump.

The Lab Advisor Advanced features can be unlocked by purchasing a license key, and include real-time monitoring of instrument actuals, all various instrument signals, and state machines. In addition, all diagnostic test results, calibration results, and acquired signal data can be uploaded to a shared network folder. The Review Client included in Lab Advisor Advanced allows to load and examine the uploaded data no matter on which instrument it was generated. This makes Data Sharing an ideal tool for internal support groups and users who want to track the instrument history of their analytical systems.

The optional Agilent Maintenance Wizard Add-on provides an easy-to-use, step-by-step multimedia guide for performing preventive maintenance on Agilent 1200 Infinity and Infinity II Series instruments.

The tests and diagnostic features that are provided by the Agilent Lab Advisor software may differ from the descriptions in this manual. For details, refer to the Agilent Lab Advisor software help files.

# 7
# Error Information

What Are Error Messages 95
General Error Messages 96
- Timeout 96
- Shutdown 96
- Remote Timeout 97
- Lost CAN Partner 98
- Leak Sensor Short 98
- Leak Sensor Open 99
- Compensation Sensor Open 99
- Compensation Sensor Short 100
- Fan Failed 101
- Leak 101

Pump Error Messages 102
- Pressure of quaternary pump above upper limit 102
- Pressure below lower limit 102
- Target pressure not reached for quaternary pump degasser 103
- Solvent counter exceeded limit 103
- Waste counter limit exceeded 104
- Flow rate limit exceeded 104
- Quaternary pump shutdown during analysis 105
- Reading the pump encoder tag failed 105
- Writing the pump encoder tag failed 106
- Pump drive blocked or encoder failed 106
- Drive current too low 107
- Drive current too high 107
- Drive timeout 108
- Overcurrent of pump drive 108
- Deliver underrun 109

Defect connection between main board and pump drive encoder 109
Pump drive encoder defect 109
Multi Purpose Valve failed 110
Reading of multi purpose valve tag failed 110
Pump drive encoder rollover 111
Drive position limit 111
Insufficient power of drive encoder LED 111
Drive encoder error 112
Writing the multi purpose valve tag failed 112
Unknown multi purpose valve type 112
Pump drive encoder error 113
Pump drive error 113
Maximum stroke is too short 114
Pump drive stop not found 114
Timeout: Wait for Composition 115
Timeout: Wait for run volume 115
Timeout: Wait for Volume 116
Timeout: Wait for Flow 116
Timeout: Waidt for Pressure 117
Drive Encoder failed 117
Drive phases differ too much in electric resistance 118
Degasser's pressure limit violation 118
Seal wash pump was missing when tried to turn on 119
Valve hardware overcurrent (MCGV) 119

This chapter describes the meaning of error messages, and provides information on probable causes and suggested actions how to recover from error conditions.

# What Are Error Messages

Error messages are displayed in the user interface when an electronic, mechanical, or hydraulic (flow path) failure occurs which requires attention before the analysis can be continued (for example, repair, or exchange of consumables is necessary). In the event of such a failure, the red status indicator at the front of the module is switched on, and an entry is written into the module logbook.

If an error occurs outside a method run, other modules will not be informed about this error. If it occurs within a method run, all connected modules will get a notification, all LEDs get red and the run will be stopped. Depending on the module type, this stop is implemented differently. For example, for a pump the flow will be stopped for safety reasons. For a detector, the lamp will stay on in order to avoid equilibration time. Depending on the error type, the next run can only be started, if the error has been resolved, for example liquid from a leak has been dried. Errors for presumably single time events can be recovered by switching on the system in the user interface.

Special handling is done in case of a leak. As a leak is a potential safety issue and may have occurred at a different module from where it has been observed, a leak always causes a shutdown of all modules, even outside a method run.

In all cases, error propagation is done via the CAN bus or via an APG remote cable (see documentation for the APG interface).

# General Error Messages

General error messages are generic to all Agilent series HPLC modules and may show up on other modules as well.

## Timeout

**Error ID: 0062**

The timeout threshold was exceeded.

| Probable cause | Suggested actions |
|---|---|
| **1** The analysis was completed successfully, and the timeout function switched off the module as requested. | Check the logbook for the occurrence and source of a not-ready condition. Restart the analysis where required. |
| **2** A not-ready condition was present during a sequence or multiple-injection run for a period longer than the timeout threshold. | Check the logbook for the occurrence and source of a not-ready condition. Restart the analysis where required. |

## Shutdown

**Error ID: 0063**

An external instrument has generated a shutdown signal on the remote line.

The module continually monitors the remote input connectors for status signals. A LOW signal input on pin 4 of the remote connector generates the error message.

| Probable cause | Suggested actions |
|---|---|
| 1 Leak detected in another module with a CAN connection to the system. | Fix the leak in the external instrument before restarting the module. |
| 2 Leak detected in an external instrument with a remote connection to the system. | Fix the leak in the external instrument before restarting the module. |
| 3 Shut-down in an external instrument with a remote connection to the system. | Check external instruments for a shut-down condition. |

## Remote Timeout

**Error ID: 0070**

A not-ready condition is still present on the remote input. When an analysis is started, the system expects all not-ready conditions (for example, a not-ready condition during detector balance) to switch to run conditions within one minute of starting the analysis. If a not-ready condition is still present on the remote line after one minute the error message is generated.

| Probable cause | Suggested actions |
|---|---|
| 1 Not-ready condition in one of the instruments connected to the remote line. | Ensure the instrument showing the not-ready condition is installed correctly, and is set up correctly for analysis. |
| 2 Defective remote cable. | Exchange the remote cable. |
| 3 Defective components in the instrument showing the not-ready condition. | Check the instrument for defects (refer to the instrument's documentation). |

## Lost CAN Partner

**Error ID: 0071**

During an analysis, the internal synchronization or communication between one or more of the modules in the system has failed.

The system processors continually monitor the system configuration. If one or more of the modules is no longer recognized as being connected to the system, the error message is generated.

| Probable cause | Suggested actions |
|---|---|
| **1** CAN cable disconnected. | • Ensure all the CAN cables are connected correctly.<br>• Ensure all CAN cables are installed correctly. |
| **2** Defective CAN cable. | Exchange the CAN cable. |
| **3** Defective main board in another module. | Switch off the system. Restart the system, and determine which module or modules are not recognized by the system. |

## Leak Sensor Short

**Error ID: 0082**

The leak sensor in the module has failed (short circuit).

The current through the leak sensor is dependent on temperature. A leak is detected when solvent cools the leak sensor, causing the leak sensor current to change within defined limits. If the current increases above the upper limit, the error message is generated.

| Probable cause | Suggested actions |
|---|---|
| **1** Defective leak sensor. | Please contact your Agilent service representative. |
| **2** Leak sensor incorrectly routed, being pinched by a metal component. | Please contact your Agilent service representative. |

## Leak Sensor Open

**Error ID: 0083**

The leak sensor in the module has failed (open circuit).

The current through the leak sensor is dependent on temperature. A leak is detected when solvent cools the leak sensor, causing the leak-sensor current to change within defined limits. If the current falls outside the lower limit, the error message is generated.

| Probable cause | Suggested actions |
|---|---|
| **1** Leak sensor not connected to the main board. | Please contact your Agilent service representative. |
| **2** Defective leak sensor. | Please contact your Agilent service representative. |
| **3** Leak sensor incorrectly routed, being pinched by a metal component. | Please contact your Agilent service representative. |

## Compensation Sensor Open

**Error ID: 0081**

The ambient-compensation sensor (NTC) on the power switch board in the module has failed (open circuit).

The resistance across the temperature compensation sensor (NTC) on the power switch board is dependent on ambient temperature. The change in resistance is used by the leak circuit to compensate for ambient temperature changes. If the resistance across the sensor increases above the upper limit, the error message is generated.

| Probable cause | Suggested actions |
|---|---|
| **1** Loose connection between the power switch board and the main board | Please contact your Agilent service representative. |
| **2** Defective power switch board | Please contact your Agilent service representative. |
| **3** Defective main board. | Please contact your Agilent service representative. |

## Compensation Sensor Short

**Error ID: 0080**

The ambient-compensation sensor (NTC) on the power switch board in the module has failed (open circuit).

The resistance across the temperature compensation sensor (NTC) on the power switch board is dependent on ambient temperature. The change in resistance is used by the leak circuit to compensate for ambient temperature changes. If the resistance across the sensor falls below the lower limit, the error message is generated.

| Probable cause | Suggested actions |
|---|---|
| **1** Defective power switch board | Please contact your Agilent service representative. |
| **2** Loose connection between the power switch board and the main board | Please contact your Agilent service representative. |
| **3** Defective main board. | Please contact your Agilent service representative. |

## Fan Failed

**Error ID: 0068**

The cooling fan in the module has failed.

The hall sensor on the fan shaft is used by the main board to monitor the fan speed. If the fan speed falls below a certain limit for a certain length of time, the error message is generated.

Depending on the module, assemblies (e.g. the lamp in the detector) are turned off to assure that the module does not overheat inside.

| Probable cause | Suggested actions |
|---|---|
| **1** Fan cable disconnected. | Please contact your Agilent service representative. |
| **2** Defective fan. | Please contact your Agilent service representative. |
| **3** Defective main board. | Please contact your Agilent service representative. |

## Leak

**Error ID: 0064**

A leak was detected in the module.

The signals from the two temperature sensors (leak sensor and board-mounted temperature-compensation sensor) are used by the leak algorithm to determine whether a leak is present. When a leak occurs, the leak sensor is cooled by the solvent. This changes the resistance of the leak sensor which is sensed by the leak-sensor circuit on the main board.

| Probable cause | Suggested actions |
|---|---|
| **1** Loose fittings. | Ensure all fittings are tight. |
| **2** Broken capillary. | Exchange defective capillaries. |

# Pump Error Messages

These errors are pump specific.

## Pressure of quaternary pump above upper limit

**Error ID: 29163**

The pressure has exceeded the upper pressure limit.

- Parameter: Measured pressure

| Probable cause | Suggested actions |
|---|---|
| 1 Blockage in flow path after the pressure sensor. | • Check for blockages in the LC system, e.g. purge valve, Jet Weaver, degraded column, column frits, needle, needle seat, capillaries etc.<br>• Check for particles in the solvent. |
| 2 Inappropriate settings (pressure limit, flow rate). | • Decrease flow rate.<br>• Increase pressure limit. |

## Pressure below lower limit

**Error ID: 29176**

The pressure has dropped below the lower limit.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| 1 Leak | Check for leaks. |
| 2 Bottle empty | Check bottle filling. |
| 3 Wrong solvent (viscosity) | Check solvent. |
| 4 Inappropriate setting | Check flow rate and lower pressure limit. |
| 5 Column degradation | Replace column. |

## Target pressure not reached for quaternary pump degasser

**Error ID: 29221**

The target pressure of the quaternary pump degasser has not been reached within the expected time.

- Parameter: Pressure in mbar

| Probable cause | Suggested actions |
|---|---|
| **1** Condensation in degasser chamber due to temperature fluctuation. | Equilibrate and restart module. |
| **2** Degasser is defect. | Please contact your Agilent service representative. |

## Solvent counter exceeded limit

**Error ID: 29146**

The counter for the solvent volume has exceeded the limit, which has been set in the user interface.

| Probable cause | Suggested actions |
|---|---|
| **1** No solvent present. | Refill solvent bottle. |
| **2** Inappropriate setting. | Check solvent counter setting in user interface. |

## Waste counter limit exceeded

**Error ID: 29147**

The counter for the waste volume has exceeded the limit, which has been set in the user interface.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| 1 The waste container is full. | Empty waste container. |
| 2 Inappropriate setting for waste counter. | • Reset waste counter.<br>• Adjust waste counter limit. |

## Flow rate limit exceeded

**Error ID: 29164**

The flow rate of the quaternary pump has exceeded the limit, while the pump runs in pressure controlled mode, e.g. during a pressure test.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| 1 Leak | Check for leaks in the pump and flow path. |
| 2 Bottle empty. | Fill solvent bottle. |
| 3 Shutoff valve closed (if applicable). | Open shutoff valve. |
| 4 Drift of pressure sensor (unlikely for short tests taking some minutes). | Replace pressure sensor. |

## Quaternary pump shutdown during analysis

**Error ID: 29199**

The quaternary pump has been shut down by the control software or control module during an analysis.

- Parameter: 0 for off, 1 for standby.

| Probable cause | Suggested actions |
|---|---|
| **1** Pump has been shut down. | Restart pump. |

## Reading the pump encoder tag failed

**Error ID: 29201**

Reading the pump encoder tag has failed.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Defect connection between encoder and main board. | Please contact your Agilent service representative. |
| **2** Missing or defect tag Defect connection between tag and encoder. | Please contact your Agilent service representative. |

## Writing the pump encoder tag failed

**Error ID: 29200**

Writing the pump encoder tag has failed.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| 1 Defect connection between encoder and main board. | Please contact your Agilent service representative. |
| 2 Defect tag Defect connection between tag and encoder. | Please contact your Agilent service representative. |

## Pump drive blocked or encoder failed

**Error ID: 29214**

Pump drive blocked or encoder failed.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| 1 Blockage of the pump drive Drive encoder failed. | Please contact your Agilent service representative. |

## Drive current too low

**Error ID: 29205**

The current consumption of the pump drive is too low.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Drive motor defect. | Please contact your Agilent service representative. |
| **2** Wrong/missing connection of pump drive to main board. | Please contact your Agilent service representative. |

## Drive current too high

**Error ID: 29236**

The current consumption of the pump drive is too high.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Blockage of system before pressure sensor. | Check for blockage of e.g. outlet valve filter frit, Multi Purpose Valve, heat exchanger. |
| **2** Drive motor defect. | Please contact your Agilent service representative. |

## Drive timeout

**Error ID: 29204**

Movement of drive during initialization is blocked mechanically.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Blockage in flow path | Remove capillary connection to system, check outlet filter, check valves, check pump head. |
| **2** Blockage of pump drive Drive motor defect. | Please contact your Agilent service representative. |

## Overcurrent of pump drive

**Error ID: 29202**

The current consumption of the pump drive is too high.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Blockage of system before pressure sensor. | Check for blockage of e.g. outlet valve filter frit, Multi Purpose Valve, heat exchanger. |
| **2** Drive motor defect. | Please contact your Agilent service representative. |

## Deliver underrun

**Error ID: 29233**

Internal error.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| **1** Internal error. | Please contact your Agilent service representative. |
| **2** Firmware issue | Use a minimum firmware revision of B.06.55 |

## Defect connection between main board and pump drive encoder

**Error ID: 29208**

Defect connection between main board and pump drive encoder.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Defect connection between main board and pump drive encoder. | Please contact your Agilent service representative. |
| **2** Defect encoder. | Please contact your Agilent service representative. |

## Pump drive encoder defect

**Error ID: 29209**

Defect pump drive encoder.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Defect encoder. | Please contact your Agilent service representative. |

## Multi Purpose Valve failed

**Error ID: 29231**

Lost steps of the purge valve encoder.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| **1** Multi purpose valve drive mechanically blocked or defect. | • Check installation of multi purpose valve head.<br>• Replace multi purpose valve head. |

## Reading of multi purpose valve tag failed

**Error ID: 29240**

Reading the multi purpose valve tag failed.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| **1** Reading of multi purpose valve tag failed. | Check cable connection. |
| **2** Multi purpose valve head tag defect or empty. | Replace multi purpose valve head. |
| **3** Multi purpose valve tag reader is defect. | Please contact your Agilent service representative. |

## Pump drive encoder rollover

**Error ID: 29232**

Invalid pump drive encoder signals have been detected.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Pump drive encoder is defect. | Please contact your Agilent service representative. |

## Drive position limit

**Error ID: 29234**

Internal error.

- Parameter: 1 – 4 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Internal error. | Please contact your Agilent service representative. |

## Insufficient power of drive encoder LED

**Error ID: 29235**

Insufficient power of drive encoder LED.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Pump drive encoder is defect. | Please contact your Agilent service representative. |

## Drive encoder error

**Error ID: 29237, 29238, 29239, 29215**

An error has occurred for the pump drive encoder.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| 1 Pump drive encoder is defect. | Please contact your Agilent service representative. |

## Writing the multi purpose valve tag failed

**Error ID: 29241**

Writing the multi purpose valve tag failed.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| 1 Multi purpose valve head tag defect. | Replace multi purpose valve head. |
| 2 Multi purpose valve tag head reader is defect. | Please contact your Agilent service representative. |

## Unknown multi purpose valve type

**Error ID: 29242**

The type information of the multi purpose valve is invalid.

- Parameter: None

| Probable cause | Suggested actions |
|---|---|
| 1 Wrong valve head installed. | Check or replace multi purpose valve head. |
| 2 Valve head has invalid RFID tag content. | Check or replace multi purpose valve head. |

## Pump drive encoder error

**Error ID: 29211**

The pump drive encoder has generated no signal.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Pump drive encoder is defect. | Please contact your Agilent service representative. |

## Pump drive error

**Error ID: 29212, 29213**

The pump drive failed during calibration.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Pump drive encoder is defect. | Please contact your Agilent service representative. |

## Maximum stroke is too short

**Error ID: 29203**

The maximum stroke is too short.

During initialization the pump defines the operation position of the pump drives and therefore the pistons. First the pump drive moves backwards to find a mechanical stop within the ball screw. Afterwards, pistons move forwards for finding the maximum available stroke volume. These values are expected within a pre-defined range. "Maximum stroke too short" means that the outer drive position is too close. This can be caused by a drive initialization without pump head or if the pump head has not been installed properly (screws are loose).

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Wiper shifted | Please contact your Agilent service representative. |
| **2** Pump head blocks piston movement | Replace, clean or repair pump head. |
| **3** Pump drive motor is mechanically blocked. | Please contact your Agilent service representative. |

## Pump drive stop not found

**Error ID: 29207**

The pump drive stop has not been found.

- Parameter: 1 – 2 referring to pump drive

| Probable cause | Suggested actions |
|---|---|
| **1** Pump drive spindle is defect. | Please contact your Agilent service representative. |

## Timeout: Wait for Composition

**Error ID: 29180**

A target condition (composition) has been sent to the instrument which should have been reached within an expected time frame but didn't. Either the limit, time frame or the current value of the variable has been modified later directly or indirectly.

| Probable cause | Suggested actions |
|---|---|
| 1 Incorrect parameters have been sent to the instrument by the control software or manual changes. | Verify control software, macros, manual commands. |

## Timeout: Wait for run volume

**Error ID: 29181**

A target condition (run volume, which is the volume delivered since the method run start) has been sent to the instrument which should have been reached within an expected time frame but didn't. Either the limit, time frame or the current value of the variable has been modified later directly or indirectly (for example the flow rate).

| Probable cause | Suggested actions |
|---|---|
| 1 Incorrect parameters have been sent to the instrument by the control software or manual changes. | Verify control software, macros, manual commands. |

## Timeout: Wait for Volume

**Error ID: 29182**

A target condition (volume, which is the delivered flow since the limit has been set) has been sent to the instrument which should have been reached within an expected time frame but didn't. Either the limit, time frame or the current value of the variable has been modified later directly or indirectly (for example the flow rate).

| Probable cause | Suggested actions |
|---|---|
| 1 Incorrect parameters have been sent to the instrument by the control software or manual changes. | Verify control software, macros, manual commands. |

## Timeout: Wait for Flow

**Error ID: 29183**

A target condition (flow rate) has been sent to the instrument which should have been reached within an expected time frame but didn't. Either the limit, time frame or the current value of the variable has been modified later directly or indirectly.

| Probable cause | Suggested actions |
|---|---|
| 1 Incorrect parameters have been sent to the instrument by the control software or manual changes. | Verify control software, macros, manual commands. |

## Timeout: Waidt for Pressure

**Error ID: 29185**

A target condition (pressure) has been sent to the instrument which should have been reached within an expected time frame but didn't. Either the limit, time frame or the current value of the variable has been modified later directly or indirectly.

| Probable cause | Suggested actions |
|---|---|
| **1** Incorrect parameters have been sent to the instrument by the control software or manual changes. | Verify control software, macros, manual commands. |
| **2** Leak | Run system pressure test for identifying and localizing the leak. Tighten leak. |

## Drive Encoder failed

**Error ID: 29210**

Drive encoder failed during pump drive calibration.

| Probable cause | Suggested actions |
|---|---|
| **1** Internal error. | Contact Agilent support. |

## Drive phases differ too much in electric resistance

**Error ID: 29216**

Pump drive calibration has failed due to a strong difference electric resistances for different motor phases.

| Probable cause | Suggested actions |
|---|---|
| 1 Pump drive cable defect. | Please contact your Agilent service representative. |
| 2 Pump drive defect. | Please contact your Agilent service representative. |
| 3 Defective main board. | Please contact your Agilent service representative. |

## Degasser's pressure limit violation

**Error ID: 29220**

Pressure too far above the limit.

| Probable cause | Suggested actions |
|---|---|
| 1 Leak in degasser chamber or degasser tubing. | Please contact your Agilent service representative. |
| 2 Defect vacuum pump. | Please contact your Agilent service representative. |
| 3 Degasser chamber empty or connected to air. | Block unused degasser channels. |

## Seal wash pump was missing when tried to turn on

**Error ID: 29223**

The seal wash pump has not been detected (while being configured or detected before)

| Probable cause | Suggested actions |
|---|---|
| **1** Defect cable connection to seal wash pump. | Check cable connection. |
| **2** Defect seal wash pump motor. | Please contact your Agilent service representative. |
| **3** Defective main board. | Please contact your Agilent service representative. |

## Valve hardware overcurrent (MCGV)

**Error ID: 29227**

Power consumption too high for one of the MCGV valves.

| Probable cause | Suggested actions |
|---|---|
| **1** Cable defect. | Replace MCGV. |
| **2** Valve defect | Replace MCGV. |
| **3** Defective main board. | Please contact your Agilent service representative. |



# 8
# Test Functions and Calibrations

This chapter will describe the tests for the module.

Currently, no diagnostic procedures are available.

# 9 Maintenance

Introduction to Maintenance 125
Warnings and Cautions 126
Overview of Maintenance 128
Cleaning the Module 129
Installing Fittings and Capillaries 130
Replacing the Pressure Sensor 131
Replacing the Inlet Weaver 134
Replacing the Inlet Valve 136
Replacing the Outlet Valve 138
Removing the Jet Weaver 141
Installing the Jet Weaver 144
Replacing the Seal Wash Pump 146
Replacing the Multi-Channel Gradient Valve (MCGV) 148
Releasing a Stuck Inlet Valve 152
Replacing the Pump Head 155
Disassembling the Pump Head 163
Disassembling the Primary Pump Head 165
Disassembling the Secondary Pump Head 170
Replacing the Heat Exchanger 174
Replacing Wash Seal and Gasket 177
Assembling the Pump Head 179
Replacing the Multi Purpose Valve 191
Replacing Parts of the Multi Purpose Valve 194
Replacing the Outlet Filter 196
Installing the Inline Filter 198

Removing the Inline Filter 200
Replacing Parts of the Inline Filter 202
Installing the Valve Rail Kit 205
Replacing Module Firmware 206
Preparing the Pump Module for Transport 207

This chapter describes the maintenance of the Agilent 1290 Infinity Quaternary Pump.

# Introduction to Maintenance

Figure 17 on page 125 shows the main user-accessible assemblies of the Agilent 1290 Infinity Quaternary Pump. These parts can be accessed from the front (simple repairs) and don't require to remove the pump from the system stack.




**Figure 17** Overview of Maintenance Parts

# Warnings and Cautions

**WARNING**

**Toxic, flammable and hazardous solvents, samples and reagents**

**The handling of solvents, samples and reagents can hold health and safety risks.**

- ➔ When working with these substances observe appropriate safety procedures (for example by wearing goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet supplied by the vendor, and follow good laboratory practice.
- ➔ The volume of substances should be reduced to the minimum required for the analysis.
- ➔ Do not operate the instrument in an explosive atmosphere.

---

**WARNING**

**Electrical shock**

**Repair work at the module can lead to personal injuries, e.g. shock hazard, when the cover is opened.**

- ➔ Do not remove the cover of the module.
- ➔ Only certified persons are authorized to carry out repairs inside the module.

---

**WARNING**

**Personal injury or damage to the product**

**Agilent is not responsible for any damages caused, in whole or in part, by improper use of the products, unauthorized alterations, adjustments or modifications to the products, failure to comply with procedures in Agilent product user guides, or use of the products in violation of applicable laws, rules or regulations.**

- ➔ Use your Agilent products only in the manner described in the Agilent product user guides.

---

**CAUTION**

Safety standards for external equipment

➔ If you connect external equipment to the instrument, make sure that you only use accessory units tested and approved according to the safety standards appropriate for the type of external equipment.

# Overview of Maintenance

The following pages describe maintenance (simple repairs) of the module that can be carried out without opening the main cover.

# Cleaning the Module

To keep the module case clean, use a soft cloth slightly dampened with water, or a solution of water and mild detergent.

**WARNING** **Liquid dripping into the electronic compartment of your module can cause shock hazard and damage the module**

➔ Do not use an excessively damp cloth during cleaning.

➔ Drain all solvent lines before opening any connections in the flow path.

# Installing Fittings and Capillaries

**WARNING**

**Solvent can spray under high pressure.**

➔ Observe appropriate safety procedures (for example, goggles, safety gloves and protective clothing), when opening flow path.

**CAUTION**

Deformation of fittings and seals

Liquid drops under high pressure up to 1200 bar act like solid parts. Tightening connections under high pressure can deform or destroy fittings and seals.

➔ Never tighten flow connections under pressure.

**NOTE**

The lifetime of a fitting depends on how firmly it has been tightened; firm tightening reduces the lifetime.

If fitting has been overtightened, replace it.

1 Install fittings and capillaries.

2 Tighten fittings and capillaries.

# Replacing the Pressure Sensor

**When** No or invalid pressure signal

**Tools required**

| p/n | Description |
|---|---|
| 8710-2412 | Hex key 2.5 mm, 15 cm long, straight handle |
| 8710-0510 | Wrench open 1/4 — 5/16 inch |
| | Screwdriver |

**Parts required**

| # | p/n | Description |
|---|---|---|
| 1 | G4220-60001 | Pressure sensor 1200 bar |

**Preparations** Turn off pump flow, switch off pump

**NOTE** This procedure describes how to replace the pressure sensor.

In case the cable to the sensor shall be replaced as well, please contact your Agilent service representative.

**NOTE** Working on connections to the pressure sensor may slightly change the displayed pressure. In case of a pressure offset at ambient pressure, a pressure offset calibration may be run.

**1** Remove capillary connections between the pressure sensor and the Multi Purpose Valve, and between the pressure sensor and the outlet filter of the secondary pump head, respectively.



**2** Remove the screws that fix the pressure sensor to the chassis.




**3** Carefully pull out the pressure sensor for about 2 cm. Then unscrew the cable from the pressure sensor.



**4** Connect the new pressure sensor to the pressure sensor connector.



**5** Fix the pressure sensor to the instrument chassis.




**6** Connect the capillary from the pump head outlet to the pressure sensor inlet. Two arrow signs on the pressure sensor indicate the flow direction.



**7** If applicable, connect the outlet of the pressure sensor to the central port of the Multi Purpose Valve.



# Replacing the Inlet Weaver

| Parts required | p/n | Description |
|---|---|---|
| | G4204-81090 | 1290 Infinity Quaternary Pump Inlet Weaver Assembly |

**Preparations**

- Switch off pump at the main power switch
- Remove the front cover
- Use an optional solvent shutoff valve or lift up solvent filters inside solvent bottles for avoiding leakages
- For easy access to the inlet weaver assembly, remove tubing connections between MCGV and degasser

| | |
|---|---|
| **1** Open the screw at the bottom of the inlet valve. |  |

**2** Open the fitting at the center of the multi-channel gradient valve (MCGV). Do not open the screw marked with the red cross. Remove the inlet weaver from the MCGV.




**3** Pull the inlet weaver out of the inlet valve.



**4** Insert the new inlet weaver to the inlet valve. Fix the weaver with the plastic screw.



**5** Fix the fitting of the new inlet weaver to the MCGV.




**6** Reconnect tubings between MCGV and degasser.

# Replacing the Inlet Valve

| | | |
|---|---|---|
| **When** | If Inlet valve is defective. | |
| **Tools required** | **p/n** | **Description** |
| | | Wrench, 14 mm |
| | G4220-20012 | Torque wrench 2 – 25 Nm |
| **Parts required** | **p/n** | **Description** |
| | G4204-60022 | Inlet Valve 1290 Infinity Quaternary Pump |

**Preparations**

- Switch off pump at the main power switch
- Remove the front cover
- Use an optional solvent shutoff valve or lift up solvent filters inside solvent bottles for avoiding leakages
- Remove the inlet weaver, see "Replacing the Inlet Weaver" on page 134

| | |
|---|---|
| **1** With a 14 mm wrench, unscrew the inlet valve and remove it. |  |

**2** Install the new inlet valve and tighten it using a torque wrench with a 14 mm bit set to 10 Nm.




**Next Steps:**

**3** Insert the inlet weaver, see "Replacing the Inlet Weaver" on page 134.

**4** Purge the system to remove air.

# Replacing the Outlet Valve

**When**

If Outlet valve is defective.

**Tools required**

| p/n | Description |
|---|---|
| 8710-0510 | Wrench open 1/4 — 5/16 inch |
| 8710-2603 | Spanner-double open ended 12X14 mm Chrome |
| G4220-20012 | Torque wrench 2 – 25 Nm |
| G4220-20041 | Bit Torx 10x25 mm |

**Parts required**

| p/n | Description |
|---|---|
| G4220-60028 | Outlet valve<br>(primary pump head) |
| G4220-20020 | Internal gold seal for Outlet Valve |

**Preparations**

- Switch off pump at the main power switch
- Remove the front cover
- Use an optional solvent shutoff valve or lift up solvent filters inside solvent bottles for avoiding leakages

**1** Open the 2.5 mm hex screw at the top of the primary pump head, which fixes the connection capillary of the heat exchanger. Then lift up the capillary and remove it from the primary pump head.



**2** A gold seal between outlet valve and heat exchanger capillary is used for a tight connection. The seal can be replaced separately as needed.




**3** Unscrew the outlet valve with a 14 mm wrench.



**4** Insert the new outlet valve and tighten it using a torque wrench with a 14 mm bit set to 10 Nm.




**5** Insert the heat exchanger capillary into the outlet of the outlet valve. Using a torque wrench with a 2.5 mm hex bit, set 3 Nm and close the hex screw at the top of the outlet.




**6** Purge the system to remove air.

# Removing the Jet Weaver

| Tools required | p/n | Description |
|---|---|---|
| | 8710-0510 | Wrench open 1/4 — 5/16 inch |
| | 8710-0899 | Pozidriv screwdriver |

| Parts required | # | p/n | Description |
|---|---|---|---|
| | 2 | 0100-1259 | Plastic fittings |

**Preparations**

- Select **Do not use mixer** in ChemStation.
- Switch off the pump at the main power switch.

| | |
|---|---|
| **1** Remove capillary connections from the Jet Weaver to the Multi Purpose Valve. |  |

| | |
|---|---|
| **2** Open the screw, which fixes the Jet Weaver to the front panel.<br>  | **3** Lift up the Jet Weaver (1) and pull it out of the front panel (2).<br>  |

**4** If no other Jet Weaver shall be installed, use plastic fittings for closing unused ports of the valve and install the metal lid.




OR

Otherwise continue at "Installing the Jet Weaver" on page 144.

# Installing the Jet Weaver

**When**

The optional Jet Weaver 380 µL for 1290 Infinity Quaternary Pump (G4204-68000) is available for applications which require highest mixing performance, see chapter *Optimizing Performance*.

**Tools required**

| Description |
|---|
| Screwdriver Pozidriv #1 |

**Parts required**

| # | p/n | Description |
|---|---|---|
| 1 | G4204-68000 | Jet Weaver 380 µL for 1290 Infinity Quaternary Pump containing |
| 2 | 5067-5416 | Capillary ST 0.17 x 120 mm, SLV/SV Jet Weaver to Multi Purpose Valve |

**Preparations**

Switch off the pump at the main power switch

| | |
|---|---|
| **1** Open the screw of the Jet Weaver metal lid. |  |

**2** Remove the metal lid by lifting it up (1) and pulling it out of the front panel (2).




**3** Insert the Jet Weaver to the opening in the front panel (1) and push it down (2).




**4** Mount both capillary connections to the Jet Weaver observing the correct orientation.



**5** Connect the inlet capillary of the Jet Weaver to port 2 of the Multi Purpose Valve. Connect the outlet capillary to port 1.




**6** Configure the Jet Weaver as mixer in the user interface, see "Instrument Configuration" on page 53.

# Replacing the Seal Wash Pump

| | | |
|---|---|---|
| **When** | In case of wear of the seal wash pump | |
| **Parts required** | **p/n** | **Description** |
| | 5065-4445 | Peristaltic pump with Pharmed tubing |
| | 5065-9978 | Tubing, 1 mm i.d., 3 mm o.d., silicone, 5 m |
| **Preparations** | Remove the flow connections from and to the seal wash pump. | |

| | |
|---|---|
| **1** For removing the seal wash pump, press the clips (1) and pull the pump to the front (2). |  |

**2** Insert the pump clips to the holes in the module housing.



**3** Fix tubings of the peristaltic pump to the primary pump head outlet and secondary pump head inlet.



# Replacing the Multi-Channel Gradient Valve (MCGV)

| Tools required | p/n | Description |
|---|---|---|
| | 0100-1710 | Mounting Tool for Tubing Connections |
| | 8710-0899 | Pozidriv screwdriver |

| Parts required | p/n | Description |
|---|---|---|
| | G1311-67701 | Multi channel gradient valve (MCGV) |

**Preparations**

- Switch off pump at the main power switch
- Remove the front cover
- Use an optional solvent shutoff valve or lift up solvent filters inside solvent bottles for avoiding leakages

**NOTE** For best performance and life time, use lower channels A and D for aqueous solvents in buffer applications, see "Operational Hints for the Multi Channel Gradient Valve (MCGV)" on page 79 for details.

**1** Use the mounting tool for removing tubing connections between the degassing unit and the MCGV.



**2** Remove the inlet weaver, see "Replacing the Inlet Weaver" on page 134.




**3** Remove the waste funnel.



**4** Remove the cover from the MCGV.



**5** Disconnect the MCGV cable (1), unscrew the two screws (2, 3) and remove the valve.



**6** Place the new MCGV into position.




**NOTE**

Make sure that channel A of the MCGV is put at the bottom-right position.

**7** Tighten the two screws (1, 2) and connect the cable to its connector (3).



**8** Install the MCGV cover.



**9** Reconnect the waste funnel with the waste tube holder in the top cover. Insert waste tube in the holder in the leak pan and clip tube to the MCGV cover.




**Next Steps:**

**10** Install the inlet weaver, see "Replacing the Inlet Weaver" on page 134.

**11** Reconnect solvent tubes for channels A-D from the MCGV to the degasser outlets.

# Releasing a Stuck Inlet Valve

| Tools required | p/n | Description |
|---|---|---|
| | 9301-0411 | Syringe, Plastic |
| | 0100-1681 | Syringe adapter luer/barb |
| | 0100-1710 | Mounting Tool for Tubing Connections |
| | | Beaker |

**CAUTION**

Pressure damages the multi-channel gradient valve (MCGV) and/or degasser

➔ Never apply pressure to the MCGV or degasser.

➔ Directly connect the syringe to the inlet weaver.

| | |
|---|---|
| **1** Remove tubing connections channels A, B, C and D to the MCGV such that you can access the inlet weaver. |  |

**2** Open the fitting at the center of the multi-channel gradient valve (MCGV). Do not open the screw marked with the red cross. Remove the inlet weaver from the MCGV.




**3** Slightly open the black plastic screw at the bottom of the inlet valve, and rotate the inlet weaver to the front. Then retighten the screw.




**4** Disconnect the capillary from the pressure sensor inlet and route the capillary to a small beaker.




**5** Fill the syringe with a suitable wash solvent.

**NOTE**

For salt deposits, warm water is a good choice. For organic deposits, use ethanol or acetone.

**6** Connect the syringe and adapter to the inlet weaver.



**7** Push the syringe for flushing the inlet valve and pump head.



**8** Restore original connections. Flush the system for several minutes.

# Replacing the Pump Head

**When** For preventive maintenance or in case of problems with the pump performance

**Tools required**

| p/n | Description |
|---|---|
| G4220-20012 | Torque wrench 2 – 25 Nm |
| G4220-20013 | 4 mm hex bit |
| G4220-20015 | Adapter ¼ in square to hex |

**Parts required**

| p/n | Description |
|---|---|
| G4204-60200 | 1290 Infinity Quaternary Pump Head Assembly with Seal Wash |
| G4204-60400 | 1290 Infinity Quaternary Pump Head Assembly without Seal Wash |

**Preparations**

- Switch off pump at the main power switch
- Remove the front cover
- Use an optional solvent shutoff valve or lift up solvent filters inside solvent bottles for avoiding leakages

**CAUTION**

Limitation of life time

The pump head assembly is an exchange part which cannot be reassembled with standard tools. Disassembling the pump head will strongly limit its life time.

➔ Do not disassemble the pump head assembly.

**CAUTION**

Damage of connections

Disassembling the flow connection between the two pump heads of the pump head assembly (solvent channel) can damage the connection and cause leaks.

➔ Do not disconnect the flow connection between the pump heads.

**CAUTION**

Damage of internal parts

➔ Do not apply a strong force to the screws of the pump head.

➔ Use a torque hex key for that purpose.

**CAUTION**

Damage of the pump piston

Removing pump heads in a position other than the maintenance position can damage the pump piston.

➔ Before switching off the pump, bring it to the maintenance position.

**CAUTION**

Damage of pump drives

The pump drive can be damaged if the pump initializes after switching it on without having the pump head installed properly.

➔ Use the Lab Advisor maintenance procedure for replacing pump heads.

➔ Install the pump head correctly before switching on the pump.

**NOTE**

One pump head assembly consists of two pump heads including valves and the outlet filter, which are both removed at the same time.

| | |
|---|---|
| **1** In Lab Advisor go to **Tools > Remove/Install Pump Head** and follow instructions given on the screen.<br>**2** If applicable, remove flow connections of the seal wash option. | **3** DO NOT REMOVE the capillary connection between the pump heads marked by the red X.<br>  |

**4** Remove the capillary connection from the outlet filter on the secondary pump head to the pressure sensor.



**5** Open the black plastic screw of the inlet valve at the bottom of the left primary pump head (1) and remove the inlet weaver by first pushing it downwards (2) and then pulling it out to the left (3).



**6** Open the four screws holding the pump heads.




**NOTE**

Open all screws step by step, not screw by screw.

**7** Remove the complete pump head assembly by holding both heads and pulling it to the front.



**NOTE**

Do not further disassemble the pump head.

**8** Use a new pre-tested pump head assembly including valves and an outlet filter.



**CAUTION**

**Damage to the pump head**
**Using a wrong torque will damage the pump head.**

→ **For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.**

**9** Install the new pump head assembly by tightening the screws step by step. Apply 5 Nm using a torque hex key, which is included to the 1290 Infinity Service Kit p/n 5067-4699.



**CAUTION**

**Damage to the pump head**
**Using a wrong torque will damage the pump head.**

➔ **For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.**

**10** Insert the heat exchanger capillary into the outlet of the primary pump head. Using a torque key, which is included to the 1290 Infinity Service Kit p/n 5067-4699, set 3 Nm and close the hex screw at the top of the outlet.




**11** Insert the new inlet weaver to the inlet valve (1, 2). Fix the weaver with the plastic screw to the inlet valve (3).



**12** Connect the capillary from the pressure sensor to the pump head outlet filter.



# Disassembling the Pump Head

**When** If parts inside the pump head need to be replaced

**Tools required**

| p/n | Description |
|---|---|
| G4220-20012 | Torque wrench 2 – 25 Nm |
| G4220-20013 | 4 mm hex bit |
| G4220-20014 | 2.5 mm Hex Bit |
| G4220-20015 | Adapter ¼ in square to hex |
| 8710-0510 | Wrench open 1/4 — 5/16 inch |

**Preparations** Remove the pump head assembly as described in "Replacing the Pump Head" on page 155

**CAUTION**

Damaged pump head

Disassembling or reassembling the pump head with tools other than the ones recommended can damage pump heads and significantly reduce their life time.

➔ Follow all instructions step by step.

➔ Use recommended tools like the pump head alignment tool and a torque wrench.

**NOTE**

This section describes the maintenance of the 1290 Infinity Quaternary Pump head. Agilent recommends using refurbished pump heads for maintenance and repair instead of disassembling pump heads. Such pump heads are available through Agilent service.

The 1290 Infinity pump service kit (5067-4699) includes all tools required for the procedures described in subsequent sections.

**NOTE**

The primary pump head does not have a heat exchanger. Seal wash parts are optional for both pump heads.

**1** Open the 2.5 mm hex or Tx 10 screw at the top of the primary pump head, which fixes the connection capillary of the heat exchanger. Then lift up the capillary and remove it from the primary pump head.




**2** Remove both pump heads from the link plate by pushing the clips at the rear of the plate.



# Disassembling the Primary Pump Head

**CAUTION**

Damage of pump piston

The pump piston is made of $ZrO_2$-based ceramic, which is a very hard and resistant material, but it is sensitive to shearing forces from the side.

➔ Do not try to remove the pump piston from the rear.

➔ Do not use the piston for removing pump seals.

| | |
|---|---|
| **1** Remove the outlet valve at the top of the pump head and the inlet valve at the bottom of the primary pump head. |  |

**2** For disassembling the pump head, remove the 4 hex screws at the rear of the pump head.



**3** Remove the front part of the pump head including pump chamber housing with pump seal and support ring.




**4** Remove the front part of the pump head including pump chamber housing with pump seal and support ring. If the seal wash option is installed, also remove the seal holder including backup ring, and the gasket.




**5** Remove the piston from the piston housing by pushing it to the rear, then pull it out from the rear.



**6** Check the pump pistons for scratches, grooves, and dents when changing the piston seals.

**NOTE**

Damaged pistons cause micro leaks and will decrease the lifetime of the seals.




**7** Use the steel side of the insert tool for removing the pump seal from the pump chamber housing.




**NOTE**

Do not use the pump piston for that purpose, as this can break it!

**8** To flush out particles from the pump head, use pressurized air ((1), (2) and (3)).




**NOTE**

Pump head parts in contact with the piston seal need to be cleaned properly in order to get a smooth surface and a tight connection.

**9** To remove wear, use isopropanol.

**10** To clean the piston and the pump head surface, use the ultra fine abrasive mesh (p/n 8660-0852).

**11** To dry the pump head and to remove residues, use pressurized air ((1), (2) and (3)).




**12** If the wash seal shall be replaced by a new one, use the steel side of the insert tool for removing it.



# Disassembling the Secondary Pump Head

**CAUTION**

Damage of pump piston

The pump piston is made of $ZrO_2$-based ceramic, which is a very hard and resistant material, but it is sensitive to shearing forces from the side.

➔ Do not try to remove the pump piston from the rear.

➔ Do not use the piston for removing pump seals.

| | |
|---|---|
| **1** Remove the outlet filter at the top of the pump head. |  |

**2** For disassembling the pump head, remove the 4 hex screws at the rear of the pump head.



**3** Remove the front part of the pump head including pump chamber housing with pump seal and seal holder. If the seal wash option is installed, also remove the support ring with wash seal and gasket.




**4** Remove the piston from the piston housing by pushing it to the rear, then pull it out from the rear.



**5** Check the pump pistons for scratches, grooves and dents when changing the piston seals.

**NOTE**

Damaged pistons cause micro leaks and will decrease the lifetime of the seals.

**6** Use the steel side of the insert tool for removing the pump seal from the pump chamber housing.




**NOTE**

Do not use the pump piston for that purpose, as this can break it!

**7** To flush out particles from the pump head, use pressurized air ((1), (2) and (3)).




**NOTE**

Pump head parts in contact with the piston seal need to be cleaned properly in order to get a smooth surface and a tight connection.

**8** To remove wear, use isopropanol.

**9** To clean the piston and the pump head surface, use the ultra fine abrasive mesh (p/n 8660-0852).

**10** To dry the pump head and to remove residues, use pressurized air ((1), (2) and (3)).




**11** If the wash seal shall be replaced by a new one, use the steel side of the insert tool for removing it.



# Replacing the Heat Exchanger

| Tools required | p/n | Description |
|---|---|---|
| | | Wrench, 19 mm |
| | 5023-2501 | Screwdriver Torx-T10 |

| Parts required | p/n | Description |
|---|---|---|
| | G4220-81013 | Heat Exchanger (secondary pump head only) |
| | G4220-20028 | Headless screw for 1290 Infinity pump heads |
| | G4220-20001 | Spacer Fitting |

**Preparations**

- Remove the pump head assembly from the pump
- Remove the secondary pump head from the link plate

**CAUTION**

Loss of small spacer fitting

Inside the secondary pump head is a small spacer fitting, which can be dropped easily when removing the heat exchanger.

➔ The heat exchanger does not need to be removed for pump head maintenance.

**1** Remove the 19 mm screw at the front of the secondary pump head.



**2** Remove the front plate.



**3** Remove the screw at the bottom of the pump head. Do not drop the golden spacer fitting.



**4** Lift out the heat exchanger.



**5** If removed, first insert the spacer fitting. Then insert the new heat exchanger to the opening in the pump head and lift it over the pins.




**6** Use the 19 mm screw for fixing the front plate.




**7** Insert and fix the screw.



# Replacing Wash Seal and Gasket

| Tools required | p/n | Description |
|---|---|---|
| | 01018-23702 | Insert tool |

| Parts required | p/n | Description |
|---|---|---|
| | 0905-1718 | Wash Seal PE |
| | 5062-2484 | Gasket, seal wash (pack of 6) |

| | |
|---|---|
| **1** Use the steel side of the insert tool for removing the wash seal. |  |

**2** Use the soft plastic side of the insert tool for inserting the wash seal into the support ring. Wet the wash seal with isopropanol or another suitable solvent before insertion.




**3** Put the gasket into the support ring and insert the seal holder.



# Assembling the Pump Head

**When** Before installing the pump head.

**Tools required**

| p/n | Description |
|---|---|
| | Pump head alignment tool |
| G4220-20012 | Torque wrench 2 – 25 Nm |
| G4220-20013 | 4 mm hex bit |
| G4220-20041 | Bit Torx 10x25 mm |
| G4220-20015 | Adapter ¼ in square to hex |
| 01018-23702 | Insert tool |

**Parts required**

| p/n | Description |
|---|---|
| 0905-1719 | PE Seal |

See chapter *"Parts"* for details.

**CAUTION**

Damage of the pump piston

The pump piston is very sensitive to shearing forces from the side.

➔ Use the alignment piston of the pump head alignment tool for the alignment procedure described below.

**CAUTION**

Wrong orientation of pins on support ring

Assembling the pump head without paying attention to the correct orientation of the pins on the support ring can lead to leaks or damage of the piston and pump head.

➔ Observe pins on the support ring, which help assembling the parts of the pump head in the correct orientation.

**CAUTION**

Damage of the pump head assembly

When installing the pump head assembly, the pump drives need to be in maintenance position, where they are retracted. Using the pump drive in default position will damage the pump head assembly.

➔ Bring the pump drive to the maintenance position.

**CAUTION**

Damage to the pump head

Using a wrong torque will damage the pump head.

➔ For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.

**NOTE**

This procedure describes how to assemble the secondary pump head using the pump head alignment tool. Assembling the primary pump head can be done accordingly. The secondary pump head has the heat exchanger capillary, which must fit into the openings of the alignment tool, whereas the primary pump head does not have a heat exchanger.

| | |
|---|---|
| **1** Use the soft plastic side of the insert tool for inserting the piston seal to the pump chamber housing. Wet the piston seal with isopropanol or another suitable solvent before insertion. |  |

**2** Insert the support ring and pump head ferrules into the piston housing. Observe the pins on the support ring, which help you assembling the pump head correctly.




**3** Assemble the pump head by putting the pump chamber housing on top of the support ring. Observe correct orientation of the pin.




**NOTE**

Do NOT install the inlet and outlet valves and the outlet filter at this stage.

**4** Loosely close the 4 screws at the rear of the pump head. The screws will be fixed tightly later.



**5** Insert the alignment piston of the pump head alignment tool. Lubricate the alignment piston with isopropanol or another suitable solvent before insertion.



**CAUTION**

**Damage to the pump head.**

➔ **Using the alignment tool is mandatory.**

➔ **Not using will break the pump head.**

**6** Insert the pump head to the pump head alignment tool, which is included to the 1290 Infinity Service Kit p/n 5067-4699. There are openings for the seal wash support ring and heat exchanger of the secondary pump head. Observe the correct orientation of all parts.




**7** Close the tool by closing the three screws at the connection ring.



**CAUTION**

**Damage to the pump head**
**Using a wrong torque will damage the pump head.**

➔ **For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.**

**8** Using a torque key, which is included to the 1290 Infinity Service Kit p/n 5067-4699, set 5 Nm and fix the central alignment screw.




**CAUTION**

**Damage to the pump head**
**Using a wrong torque will damage the pump head.**

➔ **For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.**

**9** Using a torque key, which is included to the 1290 Infinity Service Kit p/n 5067-4699, set 5 Nm and fix the 4 screws at the rear of the alignment tool. Tighten screws crosswise.




**NOTE**

This procedure will align pump head parts to their correct positions and close the pump head tightly.

**10** Open the 3 screws which have closed the pump head alignment tool and take out the aligned pump head. In case the pump head sticks inside the alignment tool, you can use the handle and insert it to the rear of the tool for pushing out the pump head.




**CAUTION**

**Damage to the pump head**
**Using a wrong torque will damage the pump head.**

→ **For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.**

**11** For the primary pump head, install the inlet valve and outlet valve using the torque wrench, which is included to the 1290 Infinity Service Kit p/n 5067-4699. Set 10 Nm for the inlet valve and 10 Nm for the outlet valve.



**12** Push in a thin pin for removing the old filter frit from the filter assembly. Insert a new filter frit (p/n 5067-5716) in the correct orientation (observe different thickness of holder on either side).




**CAUTION**

**Damage to the pump head**
**Using a wrong torque will damage the pump head.**

➔ **For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.**

**13** For the secondary pump head, install the outlet filter using the torque wrench (14 mm hex bit), which is included to the 1290 Infinity Service Kit p/n 5067-4699, set to 16 Nm.



**14** Remove the alignment piston.



**15** Lubricate the pump piston with isopropanol or another suitable solvent, and insert it.



**16** Insert both pump heads to the link plate and make sure that the clips snap in that fix the pump heads.




**NOTE**

Observe the correct orientation of the primary and the secondary pump head. This is important for correct fixation of the heat exchanger and the capillaries, as described in the following steps.

**NOTE**

Mind the orientation of the ID tag.

**17** Bring the pump drive to the maintenance position using the Lab Advisor user interface, see "Replacing the Pump Head" on page 155. Both pump drives must be retracted.



**CAUTION**

**Damage to the pump head**
**Using a wrong torque will damage the pump head.**

➔ **For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.**

**18** Install the new pump head assembly by tightening the screws step by step. Apply 5 Nm using a torque hex key, which is included to the 1290 Infinity Service Kit p/n 5067-4699.




**CAUTION**

**Damage to the pump head**
**Using a wrong torque will damage the pump head.**

➔ **For handling the torque wrench, setting and applying the right torque, consult the manual of your torque wrench.**

**19** Insert the heat exchanger capillary into the outlet of the primary pump head. Using a torque key, which is included to the 1290 Infinity Service Kit p/n 5067-4699, set 3 Nm and close the hex screw at the top of the outlet.



**20** Insert the new inlet weaver to the inlet valve (1, 2). Fix the weaver with the plastic screw to the inlet valve (3).




**21** Connect the capillary from the pressure sensor to the pump head outlet filter.



# Replacing the Multi Purpose Valve

| Tools required | p/n | Description |
|---|---|---|
| | 5023-0240 | Hex driver, ¼", slitted |

| Parts required | p/n | Description |
|---|---|---|
| | 0100-1259 | Blank nut (plastic) |
| | 01080-83202 | Blank nut (stainless steel) |
| | 5067-4174 | Multi Purpose Valve Head |

**Preparations** Remove all capillary connections from the Multi Purpose Valve.

| | |
|---|---|
| **1** Remove the clamp with the inline filter. |  |

**2** Unscrew the black union nut and remove the head of the purge valve by pulling it to the front.



**3** Put the new valve head onto the valve drive such that the lobe fits to the groove. Screw the valve head onto the valve drive using the union nut.






The central (C) port is connected to the outlet of the pressure sensor.

- Port 1 is connected to the outlet of the optional Jet Weaver
- Port 2 is connected to the inlet of the optional Jet Weaver
- Port 3 is blocked by a blank nut (plastic)
- Port 4 is connected to the system (typically autosampler)
- Port 5 is connected to the outlet of the optional inline filter
- Port 6 is blocked by a blank nut (SST)
- Port 7 is connected to the waste capillary
- Port 8 is connected to the inlet of the optional inline filter

Block unused ports with blank nuts.

If the optional inline filter is not installed, connect ports 5 and 8 with a capillary ( Capillary ST 0.17 x 120 mm, SLV/SV (5067-5416)).

# Replacing Parts of the Multi Purpose Valve

| | | |
|---|---|---|
| **Tools required** | **p/n** | **Description** |
| | 8710-2394 | 9/64 inch hex key |

| | | |
|---|---|---|
| **Parts required** | **p/n** | **Description** |
| | 1534-4045 | Bearing ring |
| | 5068-0123 | Rotor seal, Multi Purpose Valve 1290 Infinity Quaternary Pump, 1200 bar |
| | 5068-0120 | Stator ring |
| | 5068-0001 | Stator head |
| | 1535-4857 | Stator screws, 10/Pk |

**Preparations** Remove all capillary connections from the Multi Purpose Valve.

1 Use the 9/64 inch hex key for opening the valve head.

2 Replace parts as required.

**3** Reassemble the valve head and mount it to the valve drive.



# Replacing the Outlet Filter

**When**

For removing blockages and leaks in the high pressure filter assembly. The outlet filter should be replaced as required depending on the system usage. Other parts are covered by the Agilent Preventive Maintenance (PM) Service.

**Tools required**

| p/n | Description |
|---|---|
| 8710-0510 | Wrench open 1/4 — 5/16 inch |
| 8710-1924 | Wrench open 14 mm |
| | Torque wrench |
| | Torque wrench head, 14 mm for torque wrench |

**Parts required**

| p/n | Description |
|---|---|
| G4204-60004 | Outlet filter 1290 Infinity Quaternary Pump |

| | |
|---|---|
| **1** Remove the capillary from the outlet filter to the pressure sensor. |  |

**2** Remove the outlet filter using a 14 mm wrench.



**3** Mount the new outlet filter. Using a torque wrench set to 16 Nm is recommended.




**4** Mount the capillary connection to the pressure sensor.

# Installing the Inline Filter

For certain applications, Agilent recommends using an optional inline filter, which can be installed to the Multi Purpose Valve.

| **Tools required** | **p/n** | **Description** |
|---|---|---|
| | 8710-0510 | Wrench open 1/4 — 5/16 inch |

| **Parts required** | **p/n** | **Description** |
|---|---|---|
| | G7104-68000 | Inline Filter Upgrade Kit<br>The kit includes: |
| | 5067-5407 | Inline Filter Assembly |
| | 5067-4748 | Capillary ST, 0.17 mm x 90 mm |
| | G4204-40000 | Clamp for In-Line Filter |

**Preparations** Turn the pump off.

**1** Remove the capillary between ports 5 and 8 from the Multi Purpose Valve.

**2** Clip the inline filter clamp to the Multi Purpose Valve.



**3** Connect the 90 mm capillary (part of the upgrade kit) to the filter outlet.

**4** Fix the inline filter to the clamp.



**5** Install the integrated capillary of the inline filter to port 5 of the Multi Purpose Valve.




**6** Install the removable capillary of the inline filter to port 8 of the Multi Purpose Valve.



# Removing the Inline Filter

| Tools required | p/n | Description |
|---|---|---|
| | 8710-0510 | Wrench open 1/4 — 5/16 inch |

| Parts required | p/n | Description |
|---|---|---|
| | 5067-5416 | Capillary ST 0.17 x 120 mm, SLV/SV |

| | |
|---|---|
| **1** Remove the capillaries from the Multi Purpose Valve to the inline filter. |  |

**2** Remove the clamp with the inline filter.




**3** Install the capillary between ports 5 and 8 of the Multi Purpose Valve.



# Replacing Parts of the Inline Filter

| Tools required | p/n | Description |
|---|---|---|
| | 8710-0510 | Wrench open 1/4 — 5/16 inch |

| Parts required | p/n | Description |
|---|---|---|
| | 5067-5407 | Inline Filter Assembly |
| | 5067-4748 | Capillary ST, 0.17 mm x 90 mm<br>Multi Purpose Valve to inline filter |
| | 5023-0271 | Frit 0.3 µm for inline filter, 5/pk |
| | 5067-5416 | Capillary ST 0.17 x 120 mm, SLV/SV |

**CAUTION**

Stuck Capillary in Multi Purpose Valve

Shortcutting the inline filter by directly connecting its right capillary to valve port 5 can damage the Multi Purpose Valve.

The size/position of this capillary in its fitting is incompatible, so it may get stuck irreversibly to the valve.

➔ Do not shortcut the filter by directly connecting its right capillary to valve port 5 in case the inline filter cannot or shall not be used.

➔ Use Capillary ST 0.17 x 120 mm, SLV/SV (5067-5416) instead.

**NOTE**

The inline filter can be cleaned using the back-flush function in the user interface of your Agilent instrument control software.

**1** Remove the capillaries from the Multi Purpose Valve to the inline filter.



**2** Remove the inline filter from the clamp attached to the Multi Purpose Valve.




**3** Use two 5/16 wrenches for opening the inline filter.



**4** Replace the filter frit and reassemble the inline filter.




**5** Put the inline filter to the clamp and install its capillaries. The integrated capillary is connected to port 5 of the Multi Purpose Valve. The removable capillary is connected to port 8.



# Installing the Valve Rail Kit

**When** This rail is needed for the installation of external valves.

**Tools required**

| Description |
|---|
| Pozidrive screwdriver #1 |

**Parts required**

| # | p/n | Description |
|---|---|---|
| 1 | 5067-4634 | Valve Rail Kit |

1 The valve rail is fixed to the pump cover by 4 screws. The position of the lower screws is marked on the module cover. First tighten these screws, and then tighten the upper screws.



# Replacing Module Firmware

**When**

The installation of newer firmware might be necessary
- if a newer version solves problems of older versions or
- to keep all systems on the same (validated) revision.

The installation of older firmware might be necessary
- to keep all systems on the same (validated) revision or
- if a new module with newer firmware is added to a system or
- if third party control software requires a special version.

**Tools required**

| | Description |
|---|---|
| | LAN/RS-232 Firmware Update Tool |
| OR | Agilent Lab Advisor software |
| OR | Instant Pilot G4208A<br>(only if supported by module) |

**Parts required**

| # | Description |
|---|---|
| 1 | Firmware, tools and documentation from Agilent web site |

**Preparations**

Read update documentation provided with the Firmware Update Tool.

**To upgrade/downgrade the module's firmware carry out the following steps:**

1 Download the required module firmware, the latest LAN/RS-232 FW Update Tool and the documentation from the Agilent web.
   - http://www.chem.agilent.com/_layouts/agilent/downloadFirmware.aspx?whid=69761

2 For loading the firmware into the module follow the instructions in the documentation.

*Module Specific Information*

There is no specific information for this module.

# Preparing the Pump Module for Transport

**When** If the module shall be transported or shipped.

**Parts required**

| p/n | Description |
|---|---|
| 9301-0411 | Syringe; Plastic |
| 9301-1337 | Syringe adapter |
| G4204-44000 | Transport protection foam |

**CAUTION** Mechanical damage

➔ For shipping the module, insert the Protective Foam to protected the module from mechanical damage.

➔ Be careful not to damage tubing or capillary connections while inserting the module in the Protective Foam.

1 Flush all solvent channels with isopropanol.
2 Remove solvent inlet tubes from solvent reservoirs and tubing clips at other modules.
3 Remove tubings between the seal wash function and solvent bottle/waste.
4 Remove cable and capillary connections to other modules.
5 Remove the module from the stack.
6 Remove the waste tube.

7 Disconnect the degasser outlet tubings at the MCGV one after another. Use a syringe for removing liquid from the degasser and the solvent tubings.




8 Reconnect the degasser outlet tubings to the MCGV. Remove the degasser inlet tubings.



**9** You may keep internal tubing and capillary connections.



**10** Carefully insert the protective foam to the front part of the instrument. Do not damage any tubing or capillary connections.



**11** Close the front cover.



**12** For transport or shipment, put the module and accessory kit to the original shipment box.

# 10
# Parts and Materials

Overview of Main Assemblies 212
Flow Connections 214
Seal Wash Function 215
Pump Head Assembly Parts 216
Primary Pump Head Parts 218
Secondary Pump Head Parts 222
Multi Purpose Valve 225
Solvent Cabinet 226
Cover Parts 228
Leak Parts 229
Accessory Kit 230
Others 231
HPLC System Tool Kit 231
1290 Infinity Pump Service Kit 232

This chapter provides information on parts for maintenance.

# Overview of Main Assemblies




**Figure 18** Overview of maintenance parts

| Item | | p/n | Description |
|---|---|---|---|
| | 1 | 5067-4174 | Multi Purpose Valve Head |
| | 2 | 5067-5407 | Inline Filter Assembly (OPTIONAL) |
| | | 5023-0271 | Frit 0.3 µm for inline filter, 5/pk (OPTIONAL) |
| | | G4204-40000 | Clamp for In-Line Filter (OPTIONAL) |
| | | 5067-5416 | Capillary ST 0.17 x 120 mm, SLV/SV |
| | 3 | G4204-60200 | 1290 Infinity Quaternary Pump Head Assembly with Seal Wash |
| OR | | G4204-60400 | 1290 Infinity Quaternary Pump Head Assembly without Seal Wash |
| OR | 4 | G4204-81090 | 1290 Infinity Quaternary Pump Inlet Weaver Assembly |
| OR | | 5067-5443 | Inlet tubing |
| | 5 | G1311-67701 | Multi channel gradient valve (MCGV) |
| | | 5041-8365 | Blank plug<br>for MCGV |
| | 6 | G1311-60070 | Degasser 4 Channels for Quaternary Pump |
| | 7 | 5065-4445 | Peristaltic pump with Pharmed tubing |
| | 8 | G4220-60001 | Pressure sensor 1200 bar |
| | 9 | G4204-68000 | Jet Weaver 380 µL for 1290 Infinity Quaternary Pump (OPTIONAL) |

# Flow Connections




**Figure 19** Flow connections of the pump

| Item | p/n | Description |
|---|---|---|
| 1 | G4220-60007 | Bottle Head Assembly |
| 2 | G4220-60035 | Tubing kit 140 mm, 2/pk<br>degasser to MCGV |
| 3 | 5067-4657 | Capillary ST, 0.17 mm x 300 mm<br>pump to autosampler |
| 4 | 5067-5416 | Capillary ST 0.17 x 120 mm, SLV/SV<br>for Jet Weaver |
| 5 | 5067-4748 | Capillary ST, 0.17 mm x 90 mm<br>Multi Purpose Valve to inline filter |
| 6 | 5067-4656 | Capillary ST, 0.25 mm x 80 mm<br>pressure sensor to outlet filter and Multi Purpose Valve |
| 7 | 5067-4755 | Flexible Waste Tube, 5 m |
| | G4220-68070 | Ultra Clean Tubing Kit<br>(includes bottle head assemblies and tubing connections within the pump) |
| | G4220-60070 | Tubing Kit 140 mm - Ultra Clean Tubing<br>(tubes from SSV to shutoff valve or degassing unit to MCGV) |
| | G4220-60017 | Bottle Head Assembly Ultra Clean Tubing<br>(bottle heads and tubing to shutoff panel / degasser) |

# Seal Wash Function



**Figure 20** Seal Wash Pump

| p/n | Description |
|---|---|
| 5065-4445 | Peristaltic pump with Pharmed tubing |
| 5065-9978 | Tubing, 1 mm i.d., 3 mm o.d., silicone, 5 m |

# Pump Head Assembly Parts




**Figure 21** Pump head assembly parts

| Item | # | p/n | Description |
|---|---|---|---|
| | 1 | G4204-60200 | 1290 Infinity Quaternary Pump Head Assembly with Seal Wash |
| | 1 | G4204-60400 | 1290 Infinity Quaternary Pump Head Assembly without Seal Wash |
| 1 | 1 | G4220-81013 | Heat Exchanger (secondary pump head only) |
| 2 | 1 | G4220-40001 | Link Plate |
| 3 | 1 | G4204-60022 | Inlet Valve 1290 Infinity Quaternary Pump |
| 4 | 1 | G4220-60028 | Outlet valve<br>(primary pump head) |
| 5 | 1 | G4220-20020 | Internal gold seal for Outlet Valve |
| 6 | 1 | G4204-60004 | Outlet filter 1290 Infinity Quaternary Pump |
| 7 | 4 | G4220-23704 | Stay bolt |

# Primary Pump Head Parts

## Primary Pump Head with Seal Wash (Quaternary Pump)




**Figure 22** Primary pump head (Quaternary Pump) with seal wash

| Item | # | p/n | Description |
|---|---|---|---|
| 1, 11 | 6 | 0515-1218 | Screw M5, 40 mm long |
| 2 | 1 | | Pump Chamber Housing (order pump head) |
| 3 | 1 | 0905-1719 | PE Seal |
| 4 | 1 | G4220-60016 | Seal holder including backup ring |
| 5 | 1 | 5062-2484 | Gasket, seal wash (pack of 6) |
| 6 | 1 | 0905-1718 | Wash Seal PE |
| 7 | 1 | G4220-63010 | Support Ring (Seal Wash) |
| 8 | 2 | | Pump Head Ferrules (order pump head) |
| 9 | 1 | | Piston Housing (order pump head) |
| 10 | 1 | 5067-5678 | Piston 1290 Infinity Pumps, ceramic |
| 12 | 1 | G4204-60022 | Inlet Valve 1290 Infinity Quaternary Pump |
| | 1 | G4204-40006 | Fixation screw inlet weaver to PIV (not shown) |
| 13 | 1 | G4220-60028 | Outlet valve (primary pump head) |
| 14 | 1 | G4220-20028 | Headless screw for 1290 Infinity pump heads |
| 15 | 1 | G4220-20020 | Internal gold seal for Outlet Valve |

# Primary Pump Head Without Seal Wash (Quaternary Pump)




**Figure 23** Primary pump head (Quaternary Pump) without seal wash

| Item | # | p/n | Description |
|---|---|---|---|
| 1, 8 | 6 | 0515-1218 | Screw M5, 40 mm long |
| 2 | 1 | | Pump Chamber Housing (order pump head) |
| 3 | 1 | 0905-1719 | PE Seal |
| 4 | 1 | G4220-60015 | Support ring including backup ring |
| 5 | 2 | | Pump Head Ferrules (order pump head) |
| 6 | 1 | | Piston Housing (order pump head) |
| 7 | 1 | 5067-5678 | Piston 1290 Infinity Pumps, ceramic |
| 9 | 1 | G4204-60022 | Inlet Valve 1290 Infinity Quaternary Pump |
| | 1 | G4204-40006 | Fixation screw inlet weaver to PIV (not shown) |
| 10 | 1 | G4220-60028 | Outlet valve (primary pump head) |
| 11 | 1 | G4220-20028 | Headless screw for 1290 Infinity pump heads |
| 12 | 1 | G4220-20020 | Internal gold seal for Outlet Valve |

# Secondary Pump Head Parts

## Secondary Pump Head With Seal Wash (Quaternary Pump)




**Figure 24** Secondary pump head (Quaternary Pump) with seal wash

| Item | # | p/n | Description |
|---|---|---|---|
| 1, 16 | 6 | 0515-1218 | Screw M5, 40 mm long |
| 2 | 1 | G4220-20003 | Pump Head Screw |
| 3 | 1 | | Pump Head Front Plate (order pump head) |
| 4 | 1 | G4220-81013 | Heat Exchanger (secondary pump head only) |
| 5 | 1 | G4220-20028 | Headless screw for 1290 Infinity pump heads |
| 6 | 1 | G4220-20001 | Spacer Fitting |
| 7 | 1 | | Pump Chamber Housing (order pump head) |
| 8 | 1 | 0905-1719 | PE Seal |
| 9 | 1 | G4220-60016 | Seal holder including backup ring |
| 10 | 1 | 5062-2484 | Gasket, seal wash (pack of 6) |
| 11 | 1 | 0905-1718 | Wash Seal PE |
| 12 | 1 | G4220-63010 | Support Ring (Seal Wash) |
| 13 | 2 | | Pump Head Ferrules (order pump head) |
| 14 | 1 | | Piston Housing (order pump head) |
| 15 | 1 | 5067-5678 | Piston 1290 Infinity Pumps, ceramic |
| 17 | 1 | G4204-60004 | Outlet filter 1290 Infinity Quaternary Pump |

# Secondary Pump Head Without Seal Wash (Quaternary Pump)




**Figure 25** Secondary pump head (Quaternary Pump) without seal wash

| Item | # | p/n | Description |
|---|---|---|---|
| 1, 13 | 6 | 0515-1218 | Screw M5, 40 mm long |
| 2 | 1 | G4220-20003 | Pump Head Screw |
| 3 | 1 | | Pump Head Front Plate (order pump head) |
| 4 | 1 | G4220-81013 | Heat Exchanger (secondary pump head only) |
| 5 | 1 | G4220-20028 | Headless screw for 1290 Infinity pump heads |
| 6 | 1 | G4220-20001 | Spacer Fitting |
| 7 | 1 | | Pump Chamber Housing (order pump head) |
| 8 | 1 | 0905-1719 | PE Seal |
| 9 | 1 | G4220-60015 | Support ring including backup ring |
| 10 | 2 | | Pump Head Ferrules (order pump head) |
| 11 | 1 | | Piston Housing (order pump head) |
| 12 | 1 | 5067-5678 | Piston 1290 Infinity Pumps, ceramic |
| 14 | 1 | G4204-60004 | Outlet filter 1290 Infinity Quaternary Pump |

# Multi Purpose Valve




**Figure 26** Multi-purpose valve parts

| Item | # | p/n | Description |
|---|---|---|---|
| | 1 | 5067-4174 | Multi Purpose Valve Head |
| 1 | 3 | 1535-4857 | Stator screws, 10/Pk |
| 2 | 1 | 5068-0001 | Stator head |
| 3 | 1 | 5068-0120 | Stator ring |
| 4 | 1 | 5068-0123 | Rotor seal, Multi Purpose Valve 1290 Infinity Quaternary Pump, 1200 bar |
| 5 | 1 | 1535-4045 | Bearing ring |
| 6 | 1 | 5068-0106 | Spanner nut |

# Solvent Cabinet




**Figure 27** Solvent Cabinet Parts (1)




**Figure 28** Solvent Cabinet Parts (2)

| Item | p/n | Description |
|---|---|---|
| 1 | 5065-9981 | Solvent cabinet 1200 Infinity, including all plastic parts |
| 2 | 5043-0207 | Name plate 1260 |
| 3 | 5065-9954 | Front panel, solvent cabinet |
| 4 | 5042-8907 | Leak panel |
| 5 | 9301-1450 | Solvent bottle, amber |
| 6 | 9301-1420 | Solvent bottle, transparent |
| 7 | G4220-60007 | Bottle Head Assembly |

# Cover Parts




**Figure 29** Cover parts

| Item | p/n | Description |
|---|---|---|
| 1 | 5067-5396 | 1290 Infinity Quaternary Pump Cover Kit (base, top, left, right) |
| 2 | 5042-9964 | Name plate for Agilent 1290 series |
| 3 | 5067-4683 | Front Panel |
| 4 | 5042-8914 | Serial number plate |

# Leak Parts




**Figure 30** Leak parts

| Item | p/n | Description |
|---|---|---|
| 1 | 5041-8389 | Leak funnel holder |
| 2 | 5041-8388 | Leak funnel |
| 3 | 5062-2463 | Corrugated tubing, PP, 6.5 mm id, 5 m |
| 4 | G1361-47100 | Sealing lip |
| 5 | 5042-9922 | Leak panel |
| 6 | G4280-40016 | Power Switch Coupler ZL |
| 7 | 5041-8381 | Power switch button |

# Accessory Kit

The Accessory Kit 1290 Infinity Quaternary Pump (G4204-68705) contains:

| Item | # | p/n | Description |
|---|---|---|---|
| 1 | 1 | 0100-1816 | Fitting Waste Tube to Purge Valve |
| 2 | 1 | 5067-4755 | Flexible Waste Tube, 5 m |
| 3 | 1 | 5063-6527 | Leak tubing assembly, 1 m |
| 4 | 1 | 5181-1519 | CAN cable, Agilent module to module, 1 m |
| 5 | 1 | 5042-9967 | Tubing clip (set of 5 clips) |
| 6 | 1 | 5067-4657 | Capillary ST, 0.17 mm x 300 mm<br>Pump to Autosampler |
| 7 | 1 | 5067-4670 | SST cap. 0.17 mm ID 600 mm pre-swaged |
| 8 | 1 | 5067-5443 | Inlet tubing |
| 9 | 1 | 9301-6476 | Syringe with luerlock 5 mL Polypropylene |
| 10 | 4 | 5042-9972 | Tubing grommet |
| 11 | 1 | 5023-0271 | Frit 0.3 µm for inline filter, 5/pk |

# Others

## HPLC System Tool Kit

The HPLC System Tool Kit (G4203-68708) contains some accessories and tools needed for installation and repair of the module.

| # | p/n | Description |
|---|---|---|
| 1 | 0100-1681 | Adapter syringe/seal wash tube |
| 1 | 0100-1710 | Mounting Tool for Tubing Connections |
| 1 | 01018-23702 | Insert tool |
| 1 | 5023-0240 | Hex driver, ¼", slitted |
| 1 | 8710-0060 | Hex-key wrench, 9/64 inch |
| 2 | 8710-0510 | Wrench open 1/4 — 5/16 inch |
| 1 | 8710-0641 | Hex key set 1 – 5 mm |
| 1 | 8710-0899 | Pozidriv screwdriver |
| 1 | 8710-1534 | Wrench, 4 mm both ends, open end |
| 1 | 8710-1924 | Wrench open 14 mm |
| 1 | 8710-2392 | Hex key 4 mm15 cm long T-handle |
| 1 | 8710-2393 | Hex key 1.5 mm, straight handle 10 cm |
| 1 | 8710-2394 | Hex key 9/64 inch 15 cm long T-handle |
| 1 | 8710-2409 | Wrench open end, 5/16 – 3/8 inch |
| 1 | 8710-2411 | Hex key 3 mm12 cm long |
| 1 | 8710-2412 | Hex key 2.5 mm, 15 cm long, straight handle |
| 1 | 8710-2438 | Hex key 2.0 mm |
| 1 | 8710-2509 | Screwdriver Torx TX8 |
| 1 | 8710-2594 | Double open end wrench 4 mm x 5 mm |
| 1 | 9301-0411 | Syringe, Plastic |
| 1 | 9301-1337 | Adapter syringe/solvent tube with fitting |

# 1290 Infinity Pump Service Kit

1290 Infinity pump service kit (5067-4699), contains:

| p/n | Description |
|---|---|
| G4220-20012 | Torque wrench 2 – 25 Nm |
| G4220-20013 | 4 mm hex bit |
| G4220-20014 | 2.5 mm Hex Bit |
| G4220-20015 | Adapter ¼ in square to hex |
| G4204-44000 | Transport protection foam |
| 5023-0285 | Replacement kit for 1290 Infinity pump head alignment tool (piston/handle) |

# 11
# Identifying Cables

Cable Overview 234
Analog cables 236
Remote Cables 238
BCD Cables 241
CAN/LAN Cable 243
RS-232 Cable Kit 244
Agilent 1200 Module to Printer 245

This chapter provides information on cables used with the Agilent 1200 Infinity Series modules.

# Cable Overview

**NOTE** Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.

### Analog cables

| p/n | Description |
|---|---|
| 35900-60750 | Agilent module to 3394/6 integrators |
| 35900-60750 | Agilent 35900A A/D converter |
| 01046-60105 | Analog cable (BNC to general purpose, spade lugs) |

### Remote cables

| p/n | Description |
|---|---|
| 03394-60600 | Agilent module to 3396A Series I integrators |
| | 3396 Series II / 3395A integrator, see details in section "Remote Cables" on page 238 |
| 03396-61010 | Agilent module to 3396 Series III / 3395B integrators |
| 5061-3378 | Remote Cable |
| 01046-60201 | Agilent module to general purpose |

### BCD cables

| p/n | Description |
|---|---|
| 03396-60560 | Agilent module to 3396 integrators |
| G1351-81600 | Agilent module to general purpose |

### CAN cables

| p/n | Description |
|---|---|
| 5181-1516 | CAN cable, Agilent module to module, 0.5 m |
| 5181-1519 | CAN cable, Agilent module to module, 1 m |

### LAN cables

| p/n | Description |
|---|---|
| 5023-0203 | Cross-over network cable, shielded, 3 m (for point to point connection) |
| 5023-0202 | Twisted pair network cable, shielded, 7 m (for point to point connection) |

### RS-232 cables

| p/n | Description |
|---|---|
| G1530-60600 | RS-232 cable, 2 m |
| RS232-61601 | RS-232 cable, 2.5 m<br>Instrument to PC, 9-to-9 pin (female). This cable has special pin-out, and is not compatible with connecting printers and plotters. It's also called "Null Modem Cable" with full handshaking where the wiring is made between pins 1-1, 2-3, 3-2, 4-6, 5-5, 6-4, 7-8, 8-7, 9-9. |
| 5181-1561 | RS-232 cable, 8 m |

# Analog cables



One end of these cables provides a BNC connector to be connected to Agilent modules. The other end depends on the instrument to which connection is being made.

## Agilent Module to 3394/6 Integrators

| p/n 35900-60750 | Pin 3394/6 | Pin Agilent module | Signal Name |
|---|---|---|---|
|  | 1 | | Not connected |
| | 2 | Shield | Analog - |
| | 3 | Center | Analog + |
| | | | |
| | | | |
| | | | |
| | | | |

## Agilent Module to BNC Connector

| p/n 8120-1840 | Pin BNC | Pin Agilent module | Signal Name |
|---|---|---|---|
|  | Shield | Shield | Analog - |
| | Center | Center | Analog + |

## Agilent Module to General Purpose

| p/n 01046-60105 | Pin | Pin Agilent module | Signal Name |
|---|---|---|---|
|  | 1 | | Not connected |
| | 2 | Black | Analog - |
| | 3 | Red | Analog + |

# Remote Cables




One end of these cables provides a Agilent Technologies APG (Analytical Products Group) remote connector to be connected to Agilent modules. The other end depends on the instrument to be connected to.

## Agilent Module to 3396A Integrators

| p/n 03394-60600 | Pin 3396A | Pin Agilent module | Signal Name | Active (TTL) |
|---|---|---|---|---|
|  | 9 | 1 - White | Digital ground | |
| | NC | 2 - Brown | Prepare run | Low |
| | 3 | 3 - Gray | Start | Low |
| | NC | 4 - Blue | Shut down | Low |
| | NC | 5 - Pink | Not connected | |
| | NC | 6 - Yellow | Power on | High |
| | 5,14 | 7 - Red | Ready | High |
| | 1 | 8 - Green | Stop | Low |
| | NC | 9 - Black | Start request | Low |
| | 13, 15 | | Not connected | |



## Agilent Module to 3396 Series II / 3395A Integrators

Use the cable Agilent module to 3396A Series I integrators (03394-60600) and cut pin #5 on the integrator side. Otherwise the integrator prints START; not ready.

## Agilent Module to 3396 Series III / 3395B Integrators

| p/n 03396-61010 | Pin 33XX | Pin Agilent module | Signal Name | Active (TTL) |
|---|---|---|---|---|
|  | 9 | 1 - White | Digital ground | |
| | NC | 2 - Brown | Prepare run | Low |
| | 3 | 3 - Gray | Start | Low |
| | NC | 4 - Blue | Shut down | Low |
| | NC | 5 - Pink | Not connected | |
| | NC | 6 - Yellow | Power on | High |
| | 14 | 7 - Red | Ready | High |
| | 4 | 8 - Green | Stop | Low |
| | NC | 9 - Black | Start request | Low |
| | 13, 15 | | Not connected | |

## Agilent Module to Agilent 35900 A/D Converters

| p/n 5061-3378 | Pin 35900 A/D | Pin Agilent module | Signal Name | Active (TTL) |
|---|---|---|---|---|
|  | 1 - White | 1 - White | Digital ground | |
| | 2 - Brown | 2 - Brown | Prepare run | Low |
| | 3 - Gray | 3 - Gray | Start | Low |
| | 4 - Blue | 4 - Blue | Shut down | Low |
| | 5 - Pink | 5 - Pink | Not connected | |
| | 6 - Yellow | 6 - Yellow | Power on | High |
| | 7 - Red | 7 - Red | Ready | High |
| | 8 - Green | 8 - Green | Stop | Low |
| | 9 - Black | 9 - Black | Start request | Low |

## Agilent Module to General Purpose

| p/n 01046-60201 | Wire Color | Pin Agilent module | Signal Name | Active (TTL) |
|---|---|---|---|---|
|  | White | 1 | Digital ground | |
| | Brown | 2 | Prepare run | Low |
| | Gray | 3 | Start | Low |
| | Blue | 4 | Shut down | Low |
| | Pink | 5 | Not connected | |
| | Yellow | 6 | Power on | High |
| | Red | 7 | Ready | High |
| | Green | 8 | Stop | Low |
| | Black | 9 | Start request | Low |

# BCD Cables

One end of these cables provides a 15-pin BCD connector to be connected to the Agilent modules. The other end depends on the instrument to be connected to

## Agilent Module to General Purpose

| p/n G1351-81600 | Wire Color | Pin Agilent module | Signal Name | BCD Digit |
|---|---|---|---|---|
|  | Green | 1 | BCD 5 | 20 |
| | Violet | 2 | BCD 7 | 80 |
| | Blue | 3 | BCD 6 | 40 |
| | Yellow | 4 | BCD 4 | 10 |
| | Black | 5 | BCD 0 | 1 |
| | Orange | 6 | BCD 3 | 8 |
| | Red | 7 | BCD 2 | 4 |
| | Brown | 8 | BCD 1 | 2 |
| | Gray | 9 | Digital ground | Gray |
| | Gray/pink | 10 | BCD 11 | 800 |
| | Red/blue | 11 | BCD 10 | 400 |
| | White/green | 12 | BCD 9 | 200 |
| | Brown/green | 13 | BCD 8 | 100 |
| | not connected | 14 | | |
| | not connected | 15 | + 5 V | Low |

## Agilent Module to 3396 Integrators

| p/n 03396-60560 | Pin 3396 | Pin Agilent module | Signal Name | BCD Digit |
|---|---|---|---|---|
|  | 1 | 1 | BCD 5 | 20 |
| | 2 | 2 | BCD 7 | 80 |
| | 3 | 3 | BCD 6 | 40 |
| | 4 | 4 | BCD 4 | 10 |
| | 5 | 5 | BCD0 | 1 |
| | 6 | 6 | BCD 3 | 8 |
| | 7 | 7 | BCD 2 | 4 |
| | 8 | 8 | BCD 1 | 2 |
| | 9 | 9 | Digital ground | |
| | NC | 15 | + 5 V | Low |

# CAN/LAN Cable



Both ends of this cable provide a modular plug to be connected to Agilent modules CAN or LAN connectors.

### CAN Cables

| p/n | Description |
|---|---|
| 5181-1516 | CAN cable, Agilent module to module, 0.5 m |
| 5181-1519 | CAN cable, Agilent module to module, 1 m |

### LAN Cables

| p/n | Description |
|---|---|
| 5023-0203 | Cross-over network cable, shielded, 3 m (for point to point connection) |
| 5023-0202 | Twisted pair network cable, shielded, 7 m (for point to point connection) |

# RS-232 Cable Kit

| p/n | Description |
|---|---|
| G1530-60600 | RS-232 cable, 2 m |
| RS232-61601 | RS-232 cable, 2.5 m<br>Instrument to PC, 9-to-9 pin (female). This cable has special pin-out, and is not compatible with connecting printers and plotters. It's also called "Null Modem Cable" with full handshaking where the wiring is made between pins 1-1, 2-3, 3-2, 4-6, 5-5, 6-4, 7-8, 8-7, 9-9. |
| 5181-1561 | RS-232 cable, 8 m |

# Agilent 1200 Module to Printer

| p/n | Description |
|---|---|
| 5181-1529 | Cable Printer Serial & Parallel, is a SUB-D 9 pin female vs. Centronics connector on the other end (NOT FOR FW UPDATE). For use with G1323 Control Module. |

# 12
# Hardware Information

Firmware Description 248
Electrical Connections 251
Rear View of the Module 252
Interfaces 253
Overview Interfaces 256
Setting the 8-bit Configuration Switch 260
Special Settings 262
Early Maintenance Feedback 263
Instrument Layout 264

This chapter describes the pump in more detail on hardware and electronics.

# Firmware Description

The firmware of the instrument consists of two independent sections:

- a non-instrument specific section, called *resident system*
- an instrument specific section, called *main system*

## Resident System

This resident section of the firmware is identical for all Agilent 1100/1200/1220/1260/1290 series modules. Its properties are:

- the complete communication capabilities (CAN, LAN and RS-232C)
- memory management
- ability to update the firmware of the 'main system'

## Main System

Its properties are:

- the complete communication capabilities (CAN, LAN and RS-232C)
- memory management
- ability to update the firmware of the 'resident system'

In addition the main system comprises the instrument functions that are divided into common functions like

- run synchronization through APG remote,
- error handling,
- diagnostic functions,
- or module specific functions like
  - internal events such as lamp control, filter movements,
  - raw data collection and conversion to absorbance.

## Firmware Updates

Firmware updates can be done using the following tools (latest version should be used):

- Agilent Lab Advisor software with files on the hard disk (*)
- Firmware Update Tool with local files on the hard disk (*)
- Instant Pilot (G4208A) with files on a USB Flash Disk

(*) Required tools, firmware and documentation are available from the Agilent web:
http://www.chem.agilent.com/_layouts/agilent/downloadFirmware.aspx?whid=69761

The file naming conventions are:

PPPP_RVVV_XXX.dlb, where

PPPP is the product number, for example, 1315B for the G1315B DAD,

R the firmware revision, for example, A for G1315B or B for the G1315C DAD,

VVV is the revision number, for example 650 is revision 6.50,

XXX is the build number of the firmware.

For instructions on firmware updates refer to section *Replacing Firmware* in chapter *"Maintenance"* or use the documentation provided with the *Firmware Update Tools*.

**NOTE**

Update of main system can be done in the resident system only. Update of the resident system can be done in the main system only.

Main and resident firmware must be from the same set.




**Figure 31** Firmware Update Mechanism

**NOTE**

Some modules are limited in downgrading due to their main board version or their initial firmware revision. For example, a G1315C DAD SL cannot be downgraded below firmware revision B.01.02 or to a A.xx.xx.

Some modules can be re-branded (e.g. G1314C to G1314B) to allow operation in specific control software environments. In this case the feature set of the target type are use and the feature set of the original are lost. After re-branding (e.g. from G1314B to G1314C), the original feature set is available again.

All these specific informations are described in the documentation provided with the firmware update tools.

The firmware update tools, firmware and documentation are available from the Agilent web.

- http://www.chem.agilent.com/_layouts/agilent/downloadFirmware.aspx?whid=69761

# Electrical Connections

- The CAN bus is a serial bus with high speed data transfer. The two connectors for the CAN bus are used for internal module data transfer and synchronization.
- The REMOTE connector may be used in combination with other analytical instruments from Agilent Technologies if you want to use features such as start, stop, common shut down, prepare, and so on.
- With the appropriate software, the RS-232C connector may be used to control the module from a computer through a RS-232C connection. This connector is activated and can be configured with the configuration switch.
- The power input socket accepts a line voltage of 100 – 240 VAC ± 10 % with a line frequency of 50 or 60 Hz. Maximum power consumption varies by module. There is no voltage selector on your module because the power supply has wide-ranging capability. There are no externally accessible fuses, because automatic electronic fuses are implemented in the power supply.

**NOTE** Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.

# Rear View of the Module




**Figure 32** Rear of quaternary pump

# Interfaces

The Agilent 1200 Infinity Series modules provide the following interfaces:

**Table 6** Agilent 1200 Infinity Series Interfaces

| Module | CAN | LAN/BCD (optional) | LAN (on-board) | RS-232 | Analog | APG Remote | Special |
|---|---|---|---|---|---|---|---|
| **Pumps** | | | | | | | |
| G1310B Iso Pump<br>G1311B Quat Pump<br>G1311C Quat Pump VL<br>G1312B Bin Pump<br>K1312B Bin Pump Clinical Ed.<br>G1312C Bin Pump VL<br>1376A Cap Pump<br>G2226A Nano Pump<br>G5611A Bio-inert Quat Pump | 2 | Yes | No | Yes | 1 | Yes | |
| G4220A/B Bin Pump<br>G4204A Quat Pump | 2 | No | Yes | Yes | No | Yes | CAN-DC- OUT for CAN slaves |
| G1361A Prep Pump | 2 | Yes | No | Yes | No | Yes | CAN-DC- OUT for CAN slaves |
| **Samplers** | | | | | | | |
| G1329B ALS<br>G2260A Prep ALS | 2 | Yes | No | Yes | No | Yes | THERMOSTAT for G1330B/K1330B |
| G1364B FC-PS<br>G1364C FC-AS<br>G1364D FC-µS<br>G1367E HiP ALS<br>K1367E HiP ALS Clinical Ed.<br>G1377A HiP micro ALS<br>G2258A DL ALS<br>G5664A Bio-inert FC-AS<br>G5667A Bio-inert Autosampler | 2 | Yes | No | Yes | No | Yes | THERMOSTAT for G1330B/K1330B<br>CAN-DC- OUT for CAN slaves |
| G4226A ALS | 2 | Yes | No | Yes | No | Yes | |

**Table 6** Agilent 1200 Infinity Series Interfaces

| Module | CAN | LAN/BCD (optional) | LAN (on-board) | RS-232 | Analog | APG Remote | Special |
|---|---|---|---|---|---|---|---|
| **Detectors** | | | | | | | |
| G1314B VWD VL<br>G1314C VWD VL+ | 2 | Yes | No | Yes | 1 | Yes | |
| G1314E/F VWD<br>K1314F Clinical Ed. | 2 | No | Yes | Yes | 1 | Yes | |
| G4212A/B DAD<br>K4212B DAD Clinical Ed. | 2 | No | Yes | Yes | 1 | Yes | |
| G1315C DAD VL+<br>G1365C MWD<br>G1315D DAD VL<br>G1365D MWD VL | 2 | No | Yes | Yes | 2 | Yes | |
| G1321B FLD<br>K1321B FLD Clinical Ed.<br>G1321C FLD | 2 | Yes | No | Yes | 2 | Yes | |
| G1362A RID | 2 | Yes | No | Yes | 1 | Yes | |
| G4280A ELSD | No | No | No | Yes | Yes | Yes | EXT Contact<br>AUTOZERO |
| **Others** | | | | | | | |
| G1170A Valve Drive | 2 | No | No | No | No | No | 1 |
| G1316A/C TCC<br>K1316C TCC Clinical Ed. | 2 | No | No | Yes | No | Yes | |
| G1322A DEG<br>K1322A DEG Clinical Ed. | No | No | No | No | No | Yes | AUX |
| G1379B DEG | No | No | No | Yes | No | Yes | |
| G4225A DEG<br>K4225A DEG Clinical Ed. | No | No | No | Yes | No | Yes | |

**Table 6** Agilent 1200 Infinity Series Interfaces

| Module | CAN | LAN/BCD (optional) | LAN (on-board) | RS-232 | Analog | APG Remote | Special |
|---|---|---|---|---|---|---|---|
| G4227A Flex Cube | 2 | No | No | No | No | No | CAN-DC- OUT for CAN slaves [2] |
| G4240A CHIP CUBE | 2 | Yes | No | Yes | No | Yes | CAN-DC- OUT for CAN slaves THERMOSTAT for G1330A/B (NOT USED), K1330B |

[1] Requires a HOST module with on-board LAN (e.g. G4212A or G4220A with minimum firmware B.06.40 or C.06.40) or with additional G1369C LAN Card

**NOTE** The detector (DAD/MWD/FLD/VWD/RID) is the preferred access point for control via LAN. The inter-module communication is done via CAN.

- CAN connectors as interface to other modules
- LAN connector as interface to the control software
- RS-232C as interface to a computer
- REMOTE connector as interface to other Agilent products
- Analog output connector(s) for signal output

# Overview Interfaces

## CAN

The CAN is inter-module communication interface. It is a 2-wire serial bus system supporting high speed data communication and real-time requirement.

## LAN

The modules have either an interface slot for an LAN card (e.g. Agilent G1369B/C LAN Interface) or they have an on-board LAN interface (e.g. detectors G1315C/D DAD and G1365C/D MWD). This interface allows the control of the module/system via a PC with the appropriate control software. Some modules have neither on-board LAN nor an interface slot for a LAN card (e.g. G1170A Valve Drive or G4227A Flex Cube). These are hosted modules and require a Host module with firmware B.06.40 or later or with additional G1369C LAN Card.

**NOTE** If an Agilent detector (DAD/MWD/FLD/VWD/RID) is in the system, the LAN should be connected to the DAD/MWD/FLD/VWD/RID (due to higher data load). If no Agilent detector is part of the system, the LAN interface should be installed in the pump or autosampler.

## RS-232C (Serial)

The RS-232C connector is used to control the module from a computer through RS-232C connection, using the appropriate software. This connector can be configured with the configuration switch module at the rear of the module. Refer to *Communication Settings for RS-232C*.

**NOTE** There is no configuration possible on main boards with on-board LAN. These are pre-configured for

- 19200 baud,
- 8 data bit with no parity and
- one start bit and one stop bit are always used (not selectable).

The RS-232C is designed as DCE (data communication equipment) with a 9-pin male SUB-D type connector. The pins are defined as:

**Table 7** RS-232C Connection Table

| Pin | Direction | Function |
|---|---|---|
| 1 | In | DCD |
| 2 | In | RxD |
| 3 | Out | TxD |
| 4 | Out | DTR |
| 5 | | Ground |
| 6 | In | DSR |
| 7 | Out | RTS |
| 8 | In | CTS |
| 9 | In | RI |




**Figure 33** RS-232 Cable

## Analog Signal Output

The analog signal output can be distributed to a recording device. For details refer to the description of the module's main board.

## APG Remote

The APG Remote connector may be used in combination with other analytical instruments from Agilent Technologies if you want to use features as common shut down, prepare, and so on.

Remote control allows easy connection between single instruments or systems to ensure coordinated analysis with simple coupling requirements.

The subminiature D connector is used. The module provides one remote connector which is inputs/outputs (wired- or technique).

To provide maximum safety within a distributed analysis system, one line is dedicated to **SHUT DOWN** the system's critical parts in case any module detects a serious problem. To detect whether all participating modules are switched on or properly powered, one line is defined to summarize the **POWER ON** state of all connected modules. Control of analysis is maintained by signal readiness **READY** for next analysis, followed by **START** of run and optional **STOP** of run triggered on the respective lines. In addition **PREPARE** and **START REQUEST** may be issued. The signal levels are defined as:

- standard TTL levels (0 V is logic true, + 5.0 V is false),
- fan-out is 10,
- input load is 2.2 kOhm against + 5.0 V, and
- output are open collector type, inputs/outputs (wired- or technique).

**NOTE** All common TTL circuits operate with a 5 V power supply. A TTL signal is defined as "low" or L when between 0 V and 0.8 V and "high" or H when between 2.0 V and 5.0 V (with respect to the ground terminal).

**Table 8** Remote Signal Distribution

| Pin | Signal | Description |
|---|---|---|
| 1 | DGND | Digital ground |
| 2 | PREPARE | (L) Request to prepare for analysis (for example, calibration, detector lamp on). Receiver is any module performing pre-analysis activities. |
| 3 | START | (L) Request to start run / timetable. Receiver is any module performing run-time controlled activities. |
| 4 | SHUT DOWN | (L) System has serious problem (for example, leak: stops pump). Receiver is any module capable to reduce safety risk. |
| 5 | | Not used |
| 6 | POWER ON | (H) All modules connected to system are switched on. Receiver is any module relying on operation of others. |
| 7 | READY | (H) System is ready for next analysis. Receiver is any sequence controller. |
| 8 | STOP | (L) Request to reach system ready state as soon as possible (for example, stop run, abort or finish and stop injection). Receiver is any module performing run-time controlled activities. |
| 9 | START REQUEST | (L) Request to start injection cycle (for example, by start key on any module). Receiver is the autosampler. |

## Special Interfaces

The module includes a DC-Out (24 VDC) power line that is intended to be used with certain modules that operate as CAN slaves, for example external valves. The line has a limited output of 1.7 A and is self resetting.

# Setting the 8-bit Configuration Switch

The 8-bit configuration switch is located at the rear of the module. Switch settings provide configuration parameters for LAN, serial communication protocol and instrument specific initialization procedures.

All modules with on-board LAN:

- Default is ALL switches DOWN (best settings).
  - Bootp mode for LAN and
  - 19200 baud, 8 data bit / 1 stop bit with no parity for RS-232
- For specific LAN modes switches 3-8 must be set as required.
- For boot/test modes switches 1+2 must be UP plus required mode.

**NOTE** For normal operation use the default (best) settings.




**Figure 34** Location of Configuration Switch (example shows a G4212A DAD)

**NOTE** To perform any LAN configuration, SW1 and SW2 must be set to OFF. For details on the LAN settings/configuration refer to chapter LAN Configuration.

**Table 9** 8-bit Configuration Switch (with on-board LAN)

| | Mode | | Function | | | | | |
|---|---|---|---|---|---|---|---|---|
| | SW 1 | SW 2 | SW 3 | SW 4 | SW 5 | SW 6 | SW 7 | SW 8 |
| LAN | *0* | *0* | Link Configuration | | | Init Mode Selection | | |
| Auto-negotiation | | | *0* | x | x | x | x | x |
| 10 MBit, half-duplex | | | 1 | *0* | *0* | x | x | x |
| 10 MBit, full-duplex | | | 1 | *0* | 1 | x | x | x |
| 100 MBit, half-duplex | | | 1 | 1 | *0* | x | x | x |
| 100 MBit, full-duplex | | | 1 | 1 | 1 | x | x | x |
| Bootp | | | x | x | x | *0* | *0* | *0* |
| Bootp & Store | | | x | x | x | *0* | *0* | 1 |
| Using Stored | | | x | x | x | *0* | 1 | *0* |
| DHCP | | | x | x | x | 1 | *0* | *0* |
| Using Default | | | x | x | x | *0* | 1 | 1 |
| TEST | 1 | 1 | System | | | | | NVRAM |
| Boot Resident System | | | 1 | | | | | x |
| Revert to Default Data (Coldstart) | | | x | x | x | | | 1 |

**Legend:**

0 (switch down), 1 (switch up), x (any position)

**NOTE** When selecting the mode TEST, the LAN settings are: Auto-Negotiation & Using Stored.

**NOTE** For explanation of "Boot Resident System" and "Revert to Default Data (Coldstart)" refer to "Special Settings" on page 262.

## Special Settings

The special settings are required for specific actions (normally in a service case).

**NOTE** The tables include both settings for modules – with on-board LAN and without on-board LAN. They are identified as LAN and no LAN.

### Boot-Resident

Firmware update procedures may require this mode in case of firmware loading errors (main firmware part).

If you use the following switch settings and power the instrument up again, the instrument firmware stays in the resident mode. It is not operable as a module. It only uses basic functions of the operating system for example, for communication. In this mode the main firmware can be loaded (using update utilities).

**Table 10** Boot Resident Settings (On-board LAN)

| Mode Select | SW1 | SW2 | SW3 | SW4 | SW5 | SW6 | SW7 | SW8 |
|---|---|---|---|---|---|---|---|---|
| TEST/BOOT | 1 | 1 | 1 | 0 | 0 | 0 | 0 | 0 |

### Forced Cold Start

A forced cold start can be used to bring the module into a defined mode with default parameter settings.

**CAUTION** Loss of data

Forced cold start erases all methods and data stored in the non-volatile memory. Exceptions are calibration settings, diagnosis and repair log books which will not be erased.

→ Save your methods and data before executing a forced cold start.

If you use the following switch settings and power the instrument up again, a forced cold start has been completed.

**Table 11** Forced Cold Start Settings (On-board LAN)

| Mode Select | SW1 | SW2 | SW3 | SW4 | SW5 | SW6 | SW7 | SW8 |
|---|---|---|---|---|---|---|---|---|
| TEST/BOOT | 1 | 1 | 0 | 0 | 0 | 0 | 0 | 1 |

# Early Maintenance Feedback

Maintenance requires the exchange of components which are subject to wear or stress. Ideally, the frequency at which components are exchanged should be based on the intensity of usage of the module and the analytical conditions, and not on a predefined time interval. The early maintenance feedback (**EMF**) feature monitors the usage of specific components in the instrument, and provides feedback when the user-selectable limits have been exceeded. The visual feedback in the user interface provides an indication that maintenance procedures should be scheduled.

## EMF Counters

**EMF counters** increment with use and can be assigned a maximum limit which provides visual feedback in the user interface when the limit is exceeded. Some counters can be reset to zero after the required maintenance procedure.

## Using the EMF Counters

The user-settable **EMF** limits for the **EMF Counters** enable the early maintenance feedback to be adapted to specific user requirements. The useful maintenance cycle is dependent on the requirements for use. Therefore, the definition of the maximum limits need to be determined based on the specific operating conditions of the instrument.

## Setting the EMF Limits

The setting of the **EMF** limits must be optimized over one or two maintenance cycles. Initially the default **EMF** limits should be set. When instrument performance indicates maintenance is necessary, take note of the values displayed by the **EMF counters**. Enter these values (or values slightly less than the displayed values) as **EMF** limits, and then reset the **EMF counters** to zero. The next time the **EMF counters** exceed the new **EMF** limits, the **EMF** flag will be displayed, providing a reminder that maintenance needs to be scheduled.

# Instrument Layout

The industrial design of the module incorporates several innovative features. It uses Agilent's E-PAC concept for the packaging of electronics and mechanical assemblies. This concept is based upon the use of expanded polypropylene (EPP) layers of foam plastic spacers in which the mechanical and electronic boards components of the module are placed. This pack is then housed in a metal inner cabinet which is enclosed by a plastic external cabinet. The advantages of this packaging technology are:

- virtual elimination of fixing screws, bolts or ties, reducing the number of components and increasing the speed of assembly/disassembly,
- the plastic layers have air channels molded into them so that cooling air can be guided exactly to the required locations,
- the plastic layers help cushion the electronic and mechanical parts from physical shock, and
- the metal inner cabinet shields the internal electronics from electromagnetic interference and also helps to reduce or eliminate radio frequency emissions from the instrument itself.

# 13
# LAN Configuration

What You Have To Do First 266
TCP/IP Parameter Configuration 267
Configuration Switch 268
Initialization Mode Selection 269
Dynamic Host Configuration Protocol (DHCP) 273
- General Information (DHCP) 273
- Setup (DHCP) 274

Link Configuration Selection 276
Automatic configuration with Bootp 277
- About Agilent BootP Service 277
- How BootP Service Works 278
- Situation: Cannot Establish LAN Communication 278
- Installation of BootP Service 279
- Two Methods to Determine the MAC Address 281
- Assigning IP Addresses Using the Agilent BootP Service 282
- Changing the IP Address of an Instrument Using the Agilent BootP Service 285

Manual Configuration 287
- With Telnet 288
- With the Instant Pilot (G4208A) 292

PC and User Interface Software Setup Setup 293
- PC Setup for Local Configuration 293
- User Interface Software Setup 294

This chapter provides information on connecting the module to the controller software.

# What You Have To Do First

The module has an on-board LAN communication interface.

1 Note the MAC (Media Access Control) address for further reference. The MAC or hardware address of the LAN interfaces is a world wide unique identifier. No other network device will have the same hardware address. The MAC address can be found on a label at the rear of the module underneath the configuration switch (see Figure 36 on page 266).

**Figure 35** MAC label

2 Connect the instrument's LAN interface (see Figure 36 on page 266) to

- the PC network card using a crossover network cable (point-to-point) or
- a hub or switch using a standard LAN cable.




**Figure 36** Location of LAN interfaces and MAC label

# TCP/IP Parameter Configuration

To operate properly in a network environment, the LAN interface must be configured with valid TCP/IP network parameters. These parameters are:

- IP address
- Subnet Mask
- Default Gateway

The TCP/IP parameters can be configured by the following methods:

- by automatically requesting the parameters from a network-based BOOTP Server (using the so-called Bootstrap Protocol)
- by automatically requesting the parameters from a network-based DHCP Server (using the so-called Dynamic Host Configuration Protocol). This mode requires a LAN-onboard Module or a G1369C LAN Interface card, see "Setup (DHCP)" on page 274
- by manually setting the parameters using Telnet
- by manually setting the parameters using the Instant Pilot (G4208A)

The LAN interface differentiates between several initialization modes. The initialization mode (short form 'init mode') defines how to determine the active TCP/IP parameters after power-on. The parameters may be derived from a Bootp cycle, non-volatile memory or initialized with known default values. The initialization mode is selected by the configuration switch, see Table 13 on page 269.

# Configuration Switch

The configuration switch can be accessed at the rear of the module.




**Figure 37** Location of Configuration Switch

The module is shipped with all switches set to OFF, as shown above.

**NOTE** To perform any LAN configuration, SW1 and SW2 must be set to OFF.

**Table 12** Factory Default Settings

| | |
|---|---|
| Initialization ('Init') Mode | Bootp, all switches down. For details see "Initialization Mode Selection" on page 269 |
| Link Configuration | speed and duplex mode determined by auto-negotiation, for details see "Link Configuration Selection" on page 276 |

# Initialization Mode Selection

The following initialization (init) modes are selectable:

**Table 13** Initialization Mode Switches

| | SW 6 | SW 7 | SW 8 | Init Mode |
|---|---|---|---|---|
| | OFF | OFF | OFF | Bootp |
| | OFF | OFF | ON | Bootp & Store |
| | OFF | ON | OFF | Using Stored |
| | OFF | ON | ON | Using Default |
| | ON | OFF | OFF | DHCP [1] |

[1] Requires firmware B.06.40 or above. Modules without LAN on board, see G1369C LAN Interface Card

## Bootp

When the initialization mode **Bootp** is selected, the module tries to download the parameters from a **Bootp** Server. The parameters obtained become the active parameters immediately. They are not stored to the non-volatile memory of the module. Therefore, the parameters are lost with the next power cycle of the module.




**Figure 38** **Bootp** (Principle)

## Bootp & Store

When **Bootp & Store** is selected, the parameters obtained from a **Bootp** Server become the active parameters immediately. In addition, they are stored to the non-volatile memory of the module. Thus, after a power cycle they are still available. This enables a kind of bootp once configuration of the module.

*Example:* The user may not want to have a **Bootp** Server be active in his network all the time. But on the other side, he may not have any other configuration method than **Bootp**. In this case he starts the **Bootp** Server temporarily, powers on the module using the initialization mode **Bootp & Store**, waits for the **Bootp** cycle to be completed, closes the **Bootp** Server and powers off the module. Then he selects the initialization mode Using Stored and powers on the module again. From now on, he is able to establish the TCP/IP connection to the module with the parameters obtained in that single **Bootp** cycle.




**Figure 39** Bootp & Store (Principle)

**NOTE** Use the initialization mode **Bootp & Store** carefully, because writing to the non-volatile memory takes time. Therefore, when the module shall obtain its parameters from a **Bootp** Server every time it is powered on, the recommended initialization mode is **Bootp**!

## Using Stored

When initialization mode **Using Stored** is selected, the parameters are taken from the non-volatile memory of the module. The TCP/IP connection will be established using these parameters. The parameters were configured previously by one of the described methods.




**Figure 40** Using Stored (Principle)

## Using Default

When **Using Default** is selected, the factory default parameters are taken instead. These parameters enable a TCP/IP connection to the LAN interface without further configuration, see Table 14 on page 271.




**Figure 41** Using Default (Principle)

**NOTE** Using the default address in your local area network may result in network problems. Take care and change it to a valid address immediately.

**Table 14** Using Default Parameters

| | |
|---|---|
| IP address: | 192.168.254.11 |
| Subnet Mask: | 255.255.255.0 |
| Default Gateway | not specified |

Since the default IP address is a so-called local address, it will not be routed by any network device. Thus, the PC and the module must reside in the same subnet.

The user may open a Telnet session using the default IP address and change the parameters stored in the non-volatile memory of the module. He may then close the session, select the initialization mode Using Stored, power-on again and establish the TCP/IP connection using the new parameters.

When the module is wired to the PC directly (e.g. using a cross-over cable or a local hub), separated from the local area network, the user may simply keep the default parameters to establish the TCP/IP connection.

**NOTE**

In the **Using Default** mode, the parameters stored in the memory of the module are not cleared automatically. If not changed by the user, they are still available, when switching back to the mode Using Stored.

# Dynamic Host Configuration Protocol (DHCP)

## General Information (DHCP)

The Dynamic Host Configuration Protocol (DHCP) is an auto configuration protocol used on IP networks. The DHCP functionality is available on all Agilent HPLC modules with on-board LAN Interface or LAN Interface Card, and "B"-firmware (B.06.40 or above).

When the initialization mode "DHCP" is selected, the card tries to download the parameters from a DHCP Server. The parameters obtained become the active parameters immediately. They are not stored to the non-volatile memory of the card.

Besides requesting the network parameters, the card also submits its hostname to the DHCP Server. The hostname equals the MAC address of the card, e.g. *0030d3177321*. It is the DHCP server's responsibility to forward the hostname/address information to the Domain Name Server. The card does not offer any services for hostname resolution (e.g. NetBIOS).




**Figure 42** DHCP (Principle)

**NOTE**

1. It may take some time until the DHCP server has updated the DNS server with the hostname information.
2. It may be necessary to fully qualify the hostname with the DNS suffix, e.g. *0030d3177321.country.company.com*.
3. The DHCP server may reject the hostname proposed by the card and assign a name following local naming conventions.

# Setup (DHCP)

**Software required** The modules in the stack must have at least firmware from set A.06.34 and the above mentioned modules B.06.40 or above (must from the same firmware set).

1 Note the MAC address of the LAN interface (provided with G1369C LAN Interface Card or Main Board). This MAC address is on a label on the card or at the rear of the main board, e.g. *0030d3177321*.

On the Instant Pilot the MAC address can be found under **Details** in the LAN section.




**Figure 43** LAN Setting on Instant Pilot

2 Set the Configuration Switch to DHCP either on the G1369C LAN Interface Card or the main board of above mentioned modules.

**Table 15** G1369C LAN Interface Card (configuration switch on the card)

| SW 4 | SW 5 | SW 6 | SW 7 | SW 8 | Initialization Mode |
|---|---|---|---|---|---|
| ON | OFF | OFF | OFF | OFF | DHCP |

**Table 16** LC Modules inclusive 1120/1220 (configuration switch at rear of the instrument)

| SW 6 | SW 7 | SW 8 | Initialization Mode |
|---|---|---|---|
| ON | OFF | OFF | DHCP |

3 Turn on the module that hosts the LAN interface.

4 Configure your Control Software (e.g. Agilent ChemStation, Lab Advisor, Firmware Update Tool) and use MAC address as host name, e.g. *0030d3177321*.

The LC system should become visible in the control software (see Note in section "General Information (DHCP)" on page 273).

# Link Configuration Selection

The LAN interface supports 10 or 100 Mbps operation in full- or half-duplex modes. In most cases, full-duplex is supported when the connecting network device - such as a network switch or hub - supports IEEE 802.3u auto-negotiation specifications.

When connecting to network devices that do not support auto-negotiation, the LAN interface will configure itself for 10- or 100-Mbps half-duplex operation.

For example, when connected to a non-negotiating 10-Mbps hub, the LAN interface will be automatically set to operate at 10-Mbps half-duplex.

If the module is not able to connect to the network through auto-negotiation, you can manually set the link operating mode using link configuration switches on the module.

**Table 17** Link Configuration Switches

| SW 3 | SW 4 | SW 5 | Link Configuration |
|---|---|---|---|
| OFF | - | - | speed and duplex mode determined by auto-negotiation |
| ON | OFF | OFF | manually set to 10 Mbps, half-duplex |
| ON | OFF | ON | manually set to 10 Mbps, full-duplex |
| ON | ON | OFF | manually set to 100 Mbps, half-duplex |
| ON | ON | ON | manually set to 100 Mbps, full-duplex |

# Automatic configuration with Bootp

**NOTE** All examples shown in this chapter will not work in your environment. You need your own IP-, Subnet-Mask- and Gateway addresses.

**NOTE** Assure that the detector configuration switch is set properly. The setting should be either **BootP** or **BootP & Store**, see Table 13 on page 269.

**NOTE** Assure that the detector connected to the network is powered off.

**NOTE** If the Agilent BootP Service program is not already installed on your PC, then install it from your Agilent ChemStation DVD, located in folder **BootP**.

## About Agilent BootP Service

The Agilent BootP Service is used to assign the LAN Interface with an IP address.

The Agilent BootP Service is provided on the ChemStation DVD. The Agilent BootP Service is installed on a server or PC on the LAN to provide central administration of IP addresses for Agilent instruments on a LAN. The BootP service must be running TCP/IP network protocol and cannot run a DHCP server.

## How BootP Service Works

When an instrument is powered on, an LAN Interface in the instrument broadcasts a request for an IP address or host name and provides its hardware MAC address as an identifier. The Agilent BootP Service answers this request and passes a previously defined IP address and host name associated with the hardware MAC address to the requesting instrument.

The instrument receives its IP address and host name and maintains the IP address as long as it is powered on. Powering down the instrument causes it to lose its IP address, so the Agilent BootP Service must be running every time the instrument powers up. If the Agilent BootP Service runs in the background, the instrument will receive its IP address on power-up.

The Agilent LAN Interface can be set to store the IP address and will not lose the IP address if power cycled.

## Situation: Cannot Establish LAN Communication

If a LAN communication with BootP service cannot be established, check the following on the PC:

- Is the BootP service started? During installation of BootP, the service is not started automatically.
- Does the Firewall block the BootP service? Add the BootP service as an exception.
- Is the LAN Interface using the BootP-mode instead of "Using Stored" or "Using Default" modes?

## Installation of BootP Service

Before installing and configuring the Agilent BootP Service, be sure to have the IP addresses of the computer and instruments on hand.

1 Log on as Administrator or other user with Administrator privileges.

2 Close all Windows programs.

3 Insert the Agilent ChemStation software DVD into the drive. If the setup program starts automatically, click **Cancel** to stop it.

4 Open Windows Explorer.

5 Go to the BootP directory on the Agilent ChemStation DVD and double-click **BootPPackage.msi**.

6 If necessary, click the **Agilent BootP Service...** icon in the task bar.

7 The **Welcome** screen of the **Agilent BootP Service Setup Wizard** appears. Click **Next**.

8 The **End-User License Agreement** screen appears. Read the terms, indicate acceptance, then click **Next**.

9 The **Destination Folder** selection screen appears. Install BootP to the default folder or click **Browse** to choose another location. Click **Next**.

The default location for installation is:

C:\Program Files\Agilent\BootPService\

10 Click **Install** to begin installation.

11 Files load; when finished, the **BootP Settings** screen appears.




**Figure 44** BootP Settings screen

12 In the **Default Settings** part of the screen, if known, you can enter the subnet mask and gateway.

Defaults can be used:

- The default subnet mask is 255.255.255.0
- The default gateway is 192.168.254.11

13 On the **BootP Settings** screen, click **OK**. The **Agilent BootP Service Setup** screen indicates completion.

14 Click **Finish** to exit the **Agilent BootP Service Setup** screen.

15 Remove the DVD from the drive.

This completes installation.

16 Start BootP Service in the Windows® services: On the Windows® desktop click right on **Computer** icon, select **Manage > Services and Applications > Services**. Select the **Agilent BootP Service** and click **Start**.

# Two Methods to Determine the MAC Address

## Enabling logging to discover the MAC address using BootP

If you want to see the MAC address, select the **Do you want to log BootP requests?** check box.

1 Open BootP Settings from **Start > All Programs > Agilent BootP Service > EditBootPSettings**.

2 In **BootP Settings...** check **Do you want to log BootP requests?** to enable logging.


Logging
Do you want to log bootP requests?
BootP Log File:
C:\Documents and Settings\All Users\Application Data\Agilent\BootP\LogFile

**Figure 45** Enable BootP logging

The log file is located in

C:\Documents and Settings\All Users\Application Data\Agilent\BootP\LogFile

It contains a MAC address entry for each device that requests configuration information from BootP.

3 Click **OK** to save the values or **Cancel** to discard them. The editing ends.

4 After each modification of the BootP settings (i.e. **EditBootPSettings**) a stop or start of the BootP service is required for the BootP service to accept changes. See "Stopping the Agilent BootP Service" on page 285 or "Restarting the Agilent BootP Service" on page 286.

5 Uncheck the **Do you want to log BootP requests?** box after configuring instruments; otherwise, the log file will quickly fill up disk space.

## Determining the MAC address directly from the LAN Interface card label

1 Turn off the instrument.

2 Read the MAC address from the label and record it.

The MAC address is printed on a label on the rear of the module.

See Figure 35 on page 266 and Figure 36 on page 266.

3 Turn on the instrument.

# Assigning IP Addresses Using the Agilent BootP Service

The Agilent BootP Service assigns the Hardware MAC address of the instrument to an IP address.

## Determining the MAC address of the instrument using BootP Service

1 Power cycle the Instrument.

2 After the instrument completes self-test, open the log file of the BootP Service using Notepad.

- The default location for the logfile is C:\Documents and Settings\All Users\Application Data\Agilent\BootP\LogFile.
- The logfile will not be updated if it is open.

The contents will be similar to the following:

```
02/25/10 15:30:49 PM

Status: BootP Request received at outermost layer

Status: BootP Request received from hardware address: 0010835675AC

Error: Hardware address not found in BootPTAB: 0010835675AC

Status: BootP Request finished processing at outermost layer
```

3 Record the hardware (MAC) address (for example, 0010835675AC).

4 The Error means the MAC address has not been assigned an IP address and the Tab File does not have this entry. The MAC address is saved to the Tab File when an IP address is assigned.

5 Close the log file before turning on another instrument.

6 Uncheck the **Do you want to log BootP requests?** box after configuring instruments to avoid having the logfile use up excessive disk space.

## Adding each instrument to the network using BootP

1 Follow **Start > All Programs > Agilent BootP Service** and select **Edit BootP Settings**. The BootP Settings screen appears.

2 Uncheck the **Do you want to log BootP requests?** once all instruments have been added.

The **Do you want to log BootP requests?** box must be unchecked when you have finished configuring instruments; otherwise, the log file will quickly fill up disk space.

3 Click **Edit BootP Addresses...** The **Edit BootP Addresses** screen appears.

4 Click **Add...** The **Add BootP Entry** screen appears.




**Figure 46** Enable BootP logging

5 Make these entries for the instrument:

- MAC address
- Host name, Enter a Hostname of your choice.

  The Host Name must begin with "alpha" characters (i.e. LC1260)
- IP address
- Comment (optional)
- Subnet mask
- Gateway address (optional)

The configuration information entered is saved in the Tab File.

6 Click **OK**.

7 Leave **Edit BootP Addresses** by pressing **Close**.

8 Exit **BootP Settings** by pressing **OK**.

9 After each modification of the BootP settings (i.e. EditBootPSettings) a stop or start of the BootP service is required for the BootP service to accept changes. See "Stopping the Agilent BootP Service" on page 285 or "Restarting the Agilent BootP Service" on page 286.

10 Power cycle the Instrument.

OR

If you changed the IP address, power cycle the instrument for the changes to take effect.

11 Use the PING utility to verify connectivity by opening a command window and typing:

Ping 192.168.254.11 for example.

The Tab File is located at

C:\Documents and Settings\All Users\Application Data\Agilent\BootP\TabFile

# Changing the IP Address of an Instrument Using the Agilent BootP Service

Agilent BootP Service starts automatically when your PC reboots. To change Agilent BootP Service settings, you must stop the service, make the changes, and then restart the service.

## Stopping the Agilent BootP Service

1 From the Windows control panel, select **Administrative Tools > Services**. The **Services** screen appears.




**Figure 47** Windows Services screen

2 Right-click **Agilent BootP Service**.

3 Select **Stop**.

4 Close the **Services and Administrative Tools** screen.

## Editing the IP address and other parameters in EditBootPSettings

1 Select **Start > All Programs > Agilent BootP Service** and select **Edit BootP Settings**. The **BootP Settings** screen appears.

2 When the **BootP Settings** screen is first opened, it shows the default settings from installation.

3 Press **Edit BootP Addresses...** to edit the Tab File.




**Figure 48** Edit BootP Adresses screen

4 In the **Edit BootP Addresses...** screen press **Add...** to create a new entry or select an existing line from the table and press **Modify...** or **Delete** to change the IP address, comment, subnet mask, for example, in the Tab File.

If you change the IP address, it will be necessary to power cycle the instrument for the changes to take effect.

5 Leave **Edit BootP Addresses...** by pressing **Close**.

6 Exit BootP Settings by pressing OK.

## Restarting the Agilent BootP Service

1 In the Windows control panel, select **Administrative Tools > Services**. The **Services** screen appears, see Figure 47 on page 285.

2 Right-click **Agilent BootP Service** and select **Start**.

3 Close the **Services and Administrative Tools** screens.

# Manual Configuration

Manual configuration only alters the set of parameters stored in the non-volatile memory of the module. It never affects the currently active parameters. Therefore, manual configuration can be done at any time. A power cycle is mandatory to make the stored parameters become the active parameters, given that the initialization mode selection switches are allowing it.




**Figure 49** Manual Configuration (Principle)

## With Telnet

Whenever a TCP/IP connection to the module is possible (TCP/IP parameters set by any method), the parameters may be altered by opening a Telnet session.

1 Open the system (DOS) prompt window by clicking on Windows **START** button and select **"Run..."**. Type "cmd" and press OK.

2 Type the following at the system (DOS) prompt:

- `c:\>telnet <IP address>` or
- `c:\>telnet <host name>`




**Figure 50** Telnet - Starting a session

where <IP address> may be the assigned address from a Bootp cycle, a configuration session with the Handheld Controller, or the default IP address (see "Configuration Switch" on page 268).

When the connection was established successfully, the module responds with the following:




**Figure 51** A connection to the module is made

3 Type
`?` and press enter to see the available commands.




**Figure 52** Telnet Commands

**Table 18** Telnet Commands

| Value | Description |
|---|---|
| ? | displays syntax and descriptions of commands |
| / | displays current LAN settings |
| ip <x.x.x.x> | sets new ip address |
| sm <x.x.x.x> | sets new subnet mask |
| gw <x.x.x.x> | sets new default gateway |
| exit | exits shell and saves all changes |

4 To change a parameter follows the style:

- parameter value, for example:
  `ip 134.40.27.230`

Then press [Enter], where parameter refers to the configuration parameter you are defining, and value refers to the definitions you are assigning to that parameter. Each parameter entry is followed by a carriage return.

5 Use the “/” and press Enter to list the current settings.




information about the LAN interface
MAC address, initialization mode
Initialization mode is Using Stored
active TCP/IP settings
TCP/IP status - here ready
connected to PC with controller software (e.g. Agilent ChemStation), here not connected

**Figure 53** Telnet - Current settings in "Using Stored" mode

6 Change the IP address (in this example 134.40.27.99) and type “/” to list current settings.




change of IP setting to
Initialization mode is Using Stored
active TCP/IP settings
stored TCP/IP settings in non-volatile memory

connected to PC with controller software (e.g. Agilent ChemStation), here not connected

**Figure 54** Telnet - Change IP settings

7 When you have finished typing the configuration parameters, type **exit** and press **Enter** to exit with storing parameters.




**Figure 55** Closing the Telnet Session

**NOTE** If the Initialization Mode Switch is changed now to “Using Stored” mode, the instrument will take the stored settings when the module is re-booted. In the example above it would be 134.40.27.99.

## With the Instant Pilot (G4208A)

To configure the TCP/IP parameters before connecting the module to the network, the Instant Pilot (G4208A) can be used.

1 From the Welcome screen press the **More** button.

2 Select **Configure.**

3 Press the **DAD** button.

4 Scroll down to the LAN settings.




**Figure 56** Instant Pilot - LAN Configuration

5 Press the **Edit** button (only visible if not in Edit mode), perform the required changes and press the **Done** button.

6 Leave the screen by clicking **Exit**.

# PC and User Interface Software Setup Setup

## PC Setup for Local Configuration

This procedure describes the change of the TCP/IP settings on your PC to match the module's default parameters in a local configuration (see also "Initialization Mode Selection" on page 269).


FIXED IP address
Local Area Connection Properties
General
Connect using:
Xircom CreditCard Ethernet 10/100 + Modem 56
Configure
Components checked are used by this connection:
File and Printer Sharing for Microsoft Networks
PPP over Ethernet Protocol
Internet Protocol (TCP/IP)
Install...
Uninstall
Properties
Description
Transmission Control Protocol/Internet Protocol. The default wide area network protocol that provides communication across diverse interconnected networks.
Show icon in taskbar when connected
OK
Cancel
Internet Protocol (TCP/IP) Properties
General
You can get IP settings assigned automatically if your network supports this capability. Otherwise, you need to ask your network administrator for the appropriate IP settings.
Use the following IP address:
IP address:
192 . 168 . 254 . 1
Subnet mask:
255 . 255 . 248 . 0
Default gateway:
Internet Protocol (TCP/IP) Properties
General
You can get IP settings assigned automatically if your network supports this capability. Otherwise, you need to ask your network administrator for the appropriate IP settings.
Obtain an IP address automatically
IP address:
Subnet mask:
Default gateway:
Obtain DNS server address automatically
Use the following DNS server addresses:
Preferred DNS server:
Alternate DNS server:
Advanced...
OK
Cancel
automatic IP address via DHCP
(requires special IT setup of the network)

**Figure 57** Changing the TCP/IP settings of the PC

## User Interface Software Setup

Install you user interface software according the provided *User Interface Software Setup Guide*.

# 14
# Appendix

General Safety Information 296

The Waste Electrical and Electronic Equipment (WEEE) Directive (2002-96-EC) 299

Radio Interference 300

Sound Emission 301

Agilent Technologies on Internet 302

This chapter provides additional information on safety, legal and web.

# General Safety Information

## Safety Symbols

**Table 19** Safety Symbols

| Symbol | Description |
|---|---|
| | The apparatus is marked with this symbol when the user should refer to the instruction manual in order to protect risk of harm to the operator and to protect the apparatus against damage. |
| | Indicates dangerous voltages. |
| | Indicates a protected ground terminal. |
| | Indicates eye damage may result from directly viewing the light produced by the deuterium lamp used in this product. |
| | The apparatus is marked with this symbol when hot surfaces are available and the user should not touch it when heated up. |

**WARNING**

**A WARNING**

**alerts you to situations that could cause physical injury or death.**

➔ Do not proceed beyond a warning until you have fully understood and met the indicated conditions.

**CAUTION**

A CAUTION

alerts you to situations that could cause loss of data, or damage of equipment.

➔ Do not proceed beyond a caution until you have fully understood and met the indicated conditions.

## General Safety Information

The following general safety precautions must be observed during all phases of operation, service, and repair of this instrument. Failure to comply with these precautions or with specific warnings elsewhere in this manual violates safety standards of design, manufacture, and intended use of the instrument. Agilent Technologies assumes no liability for the customer's failure to comply with these requirements.

**WARNING**

**Ensure the proper usage of the equipment.**

**The protection provided by the equipment may be impaired.**

➔ The operator of this instrument is advised to use the equipment in a manner as specified in this manual.

## Safety Standards

This is a Safety Class I instrument (provided with terminal for protective earthing) and has been manufactured and tested according to international safety standards.

## Operation

Before applying power, comply with the installation section. Additionally the following must be observed.

Do not remove instrument covers when operating. Before the instrument is switched on, all protective earth terminals, extension cords, auto-transformers, and devices connected to it must be connected to a protective earth via a ground socket. Any interruption of the protective earth grounding will cause a potential shock hazard that could result in serious personal injury. Whenever it is likely that the protection has been impaired, the instrument must be made inoperative and be secured against any intended operation.

Make sure that only fuses with the required rated current and of the specified type (normal blow, time delay, and so on) are used for replacement. The use of repaired fuses and the short-circuiting of fuse holders must be avoided.

Some adjustments described in the manual, are made with power supplied to the instrument, and protective covers removed. Energy available at many points may, if contacted, result in personal injury.

Any adjustment, maintenance, and repair of the opened instrument under voltage should be avoided whenever possible. When inevitable, this has to be carried out by a skilled person who is aware of the hazard involved. Do not attempt internal service or adjustment unless another person, capable of rendering first aid and resuscitation, is present. Do not replace components with power cable connected.

Do not operate the instrument in the presence of flammable gases or fumes. Operation of any electrical instrument in such an environment constitutes a definite safety hazard.

Do not install substitute parts or make any unauthorized modification to the instrument.

Capacitors inside the instrument may still be charged, even though the instrument has been disconnected from its source of supply. Dangerous voltages, capable of causing serious personal injury, are present in this instrument. Use extreme caution when handling, testing and adjusting.

When working with solvents, observe appropriate safety procedures (for example, goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet by the solvent vendor, especially when toxic or hazardous solvents are used.

# The Waste Electrical and Electronic Equipment (WEEE) Directive (2002-96-EC)

## Abstract

The Waste Electrical and Electronic Equipment (WEEE) Directive (2002/96/EC), adopted by EU Commission on 13 February 2003, is introducing producer responsibility on all Electric and Electronic appliances from 13 August 2005.

**NOTE**

This product complies with the WEEE Directive (2002/96/EC) marking requirements. The affixed label indicates that you must not discard this electrical/electronic product in domestic household waste.

Product Category: With reference to the equipment types in the WEEE Directive Annex I, this product is classed as a “Monitoring and Control instrumentation” product.

*Do not dispose off in domestic household waste*

To return unwanted products, contact your local Agilent office, or see www.agilent.com for more information.

# Radio Interference

Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.

## Test and Measurement

If test and measurement equipment is operated with equipment unscreened cables and/or used for measurements on open set-ups, the user has to assure that under operating conditions the radio interference limits are still met within the premises.

# Sound Emission

## Manufacturer's Declaration

This statement is provided to comply with the requirements of the German Sound Emission Directive of 18 January 1991.

This product has a sound pressure emission (at the operator position) < 70 dB.

- Sound Pressure Lp < 70 dB (A)
- At Operator Position
- Normal Operation
- According to ISO 7779:1988/EN 27779/1991 (Type Test)

# Agilent Technologies on Internet

For the latest information on products and services visit our worldwide web site on the Internet at:

http://www.agilent.com

# Index


## 8

8-bit configuration switch
  on-board LAN 260

## A

Agilent Lab Advisor software 92
Agilent Lab Advisor 92
Agilent
  on internet 302
  User Interface Software Setup 293
ambient non-operating temperature 25
ambient operating temperature 25
analog signal 257
analog
  cable 236
apg remote 258
assembling
  pump head 179
automatic configuration with Bootp 277

## B

BCD
  cable 241
bench space 24
BootP service
  installation 279
  restart 286
  settings 286
  stop 285
Bootp
  & Store 270
  automatic configuration 277
  initialization modes 269
  using default 271
  using stored 271

## C

cable
  analog 236
  BCD 241
  CAN 243
  LAN 243
  remote 238
  RS-232 244
cables
  analog 234
  BCD 234
  CAN 235
  LAN 235
  overview 234
  remote 234
  RS-232 235
CAN
  cable 243
capillaries 130
cleaning 129
compensation sensor open 99
compensation sensor short 100
condensation 24
configuration switch 268
configuration
  one stack 31
  two stack front 34
  two stack rear 35
  two stack 34
cover parts 228

## D

defect connection between main board and pump drive encoder 109
defect on arrival 30
degasser, when to use 78
degasser's pressure limit violation 118
delay volume
  description 80
deliver underrun 109
delivery checklist 30
DHCP
  general information 273
  setup 274
dimensions 25
disassembling
  primary pump head 165
  pump head 163
  secondary pump head 170
drive current too high 107
drive current too low 107
drive encoder error 112
drive encoder failed 117
drive phases differ too much in electric resistance 118
drive position limit 111
drive timeout 108

## E

electrical connections
  descriptions of 251
EMF
  early maintenance feedback 263
error messages
  compensation sensor open 99

compensation sensor short 100
defect connection between main board and pump drive encoder 109
degasser's pressure limit violation 118
deliver underrun 109
drive current too high 107
drive current too low 107
drive encoder error 112
drive encoder failed 117
drive phases differ too much in electric resistance 118
drive position limit 111
drive timeout 108
fan failed 101
flow rate limit exceeded 104
insufficient power of drive encoder LED 111
leak sensor open 99
leak sensor short 98
leak 101
lost CAN partner 98
maximum stroke too short 114
multi purpose valve failed 110
overcurrent of pump drive 108
pressure below lower limit 102
pressure exceeded upper pressure limit 102
pump drive blocked or encoder failed 106
pump drive encoder defect 109
pump drive encoder error 113
pump drive encoder rollover 111
pump drive error 113
pump drive stop not found 114
pump error 102
quaternary pump shutdown during analysis 105
reading of multi purpose valve tag failed 110
reading the pump encoder tag failed 105
remote timeout 97
seal wash pump was missing when tried to turn on 119
shutdown 96
solvent counter exceeded limit 103
target pressure not reached for quaternary pump degasser 103
timeout: wait for composition 115
timeout: wait for flow 116
timeout: wait for pressure 117
timeout: wait for run volume 115
timeout: wait for volume 116
timeout 96
unknown multi purpose valve type 112
valve hardware overcurrent (MCGV) 119
waste counter limit exceeded 104
writing the multi purpose valve tag failed 112
writing the pump encoder tag failed 106
extra-column volume 80

## F

fan failed 101
filter
  flushing 66
firmware
  description 248
  main system 248
  resident system 248
  update tool 249
  updates 249, 206
  upgrade/downgrade 206
fittings 130
flow connections 42, 214
flow rate limit exceeded 104
frequency range 25

## G

general error messages 96

## H

handling acetonitrile 74
handling acids 74
handling buffers 73
heat exchanger
  replace 174
humidity 25

## I

initialization mode selection 269
inlet valve stuck
  release 152
inlet valve
  replace 136
inlet weaver
  replace 134
inline filter
  install 198
  remove 200
  replacing parts 202
installation
  bench space 24
  seal wash 45
  site requirements 21
install
  inline filter 198
  valve rail kit 205
instrument layout 264
insufficient power of drive encoder LED 111
interfaces 253
internet 302

## J

jet weaver

install 144
remove 141

## L

LAN
automatic configuration with Bootp 277
Bootp & Store 270
Bootp 269
cable 243
configuration switch 268
configuration 265
first steps 266
initialization mode selection 269
link configuration selection 276
manual configuration with telnet 288
manual configuration 287
PC and User Interface Software Setup 293
TCP/IP parameter configuration 267
using default 271
using stored 271
leak sensor open 99
leak sensor short 98
leak 101
line frequency 25
line voltage 25
link configuration selection 276
lost CAN partner 98

## M

MAC address
determine 282
MAC
address 266
maintenance
feedback 263
introduction 125
replacing firmware 206
manual configuration
of LAN 287
maximum stroke too short 114
MCGV
replace 148
message
remote timeout 97
multi purpose valve failed 110
multi purpose valve 225
positions 14

## N

non-operating altitude 25
non-operating temperature 25

## O

operating Altitude 25
operating principle 12
operating temperature 25
operational hints, MCGV 79
optimization
achieving higher resolution 83
stack configuration 31
outlet filter
replace 196
outlet valve
replace 138
overcurrent of pump drive 108

## P

packaging
damaged 30
parts
accessory kit 230
cover 228
damaged 30
missing 30
multi purpose valve 225
overview 212
pump head assembly 216
pump service kit 232
seal wash 215
system tool kit 231
PC and User Interface Software Setup 293
performance specifications 26
performance
Optimization 77
physical specifications 25
power consideration 22
power consumption 25
power cords 23
power supply indicator 89
power switch 40
pressure below lower limit 102
pressure exceeded upper pressure limit 102
pressure sensor
replace 131, 131
primary pump head
disassemble 165
pump drive blocked or encoder failed 106
pump drive encoder defect 109
pump drive encoder error 113
pump drive encoder rollover 111
pump drive error 113
pump drive stop not found 114
pump error messages 102
pump head assembly
parts 216
pump head
assemble 179
disassemble 163
replace 155
pump
features 10
overview 11

## Q

quaternary pump shutdown during analysis 105

## R

reading of multi purpose valve tag failed 110
reading the pump encoder tag failed 105
remote
  cable 238
remove
  inline filter 200
repairs
  replacing firmware 206
replace
  inline filter 202
  outlet filter 196
replacing
  Multi Purpose Valve parts 194
  Multi Purpose Valve 191
  pump head 155
resolution
  Optimization 83
RS-232C
  cable 244

## S

safety class I 297
safety
  general information 297
  standards 25
  symbols 296
seal wash pump was missing when tried to turn on 119
seal wash
  installation 45
  parts 215
secondary pump head
  disassemble 170
shutdown 96
site requirements 21
  power cords 23
solvent counter exceeded limit 103
solvent handling 73
special interfaces 259
special settings
  boot-resident 262
  forced cold start 262
specification
  physical 25
specifications 21
  performance 26
status indicator 90
system setup and installation
  optimizing stack configuration 31
system tool kit 231

## T

target pressure not reached for quaternary pump degasser 103
TCP/IP parameter configuration 267
telnet
  configuration 288
temperature sensor 101
test functions 88
timeout: wait for composition 115
timeout: wait for flow 116
timeout: wait for pressure 117
timeout: wait for run volume 115
timeout: wait for volume 116
timeout 96
transport foam
  remove 39
transport
  prepare 207
troubleshooting
  error messages 88, 95
  status indicators 88, 89

## U

unknown multi purpose valve type 112
unpacking 30
user interfaces 91

## V

valve hardware overcurrent (MCGV) 119
valve position
  filter flush mode 17
  normal operating mode without mixer 14
  purge mode 15
  with Jet Weaver and inline filter 16
valve rail kit
  install 205
valve
  multi purpose 225
voltage range 25

## W

waste counter limit exceeded 104
weight 25
writing the multi purpose valve tag failed 112
writing the pump encoder tag failed 106

www.agilent.com

# In This Book

This manual contains technical reference information about the Agilent 1290 Infinity Quaternary Pump G4204A.

The manual describes the following:

- Introduction,
- Site requirements and specifications,
- installation,
- configuration,
- using and optimizing,
- troubleshooting and diagnostic,
- error information,
- test functions,
- maintenance,
- parts identification,
- hardware information,
- safety and related information.

© Agilent Technologies 2012-2014, 2015

Printed in Germany
04/2015

G4204-90001

Agilent Technologies